



Standard Test Methods for Chromium in Water¹

This standard is issued under the fixed designation D1687; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

This standard has been approved for use by agencies of the Department of Defense.

1. Scope*

1.1 These test methods cover the determination of hexavalent and total chromium in water. Three test methods are included as follows:

Test Method	Concentration Range	Sections
A—Photometric Diphenyl-carbohydrazide	0.01 to 0.5 mg/L	7-15
B—Atomic Absorption, Direct	0.1 to 10 mg/L	16-24
C—Atomic Absorption, Graphite Furnace	5 to 100 µg/L	25-33

1.2 Test Method A is a photometric method that measures dissolved hexavalent chromium only. Test Methods B and C are atomic absorption methods that are generally applicable to the determination of dissolved or total recoverable chromium in water without regard to valence state.

1.3 Test Method A has been used successfully with reagent grade water Types I, II, and III, tap water, 10 % NaCl solution, treated water from a synthetic organic industrial plant that meets National Pollution Discharge Elimination System (NPDES) permit requirements, and EPA-extraction procedure leachate water, process water, lake water, effluent treatment, that is, lime neutralization and precipitation of spent pickle liquor and associated rinse water from stainless steel pickling. Test Method C has been used successfully with reagent water, stock scrubber water, lake water, filtered tap water, river water, well water, production plant water, and a condensate from a medium BTU coal gasification process. Matrices used, except for reagent water, are not available for Test Method B. It is the user's responsibility to ensure the validity of these test methods for waters of untested matrices.

1.4 The values stated in either SI units or inch-pound units are to be regarded separately as standard. The values stated in each system are mathematical conversions and may not be exact equivalents; therefore, each system shall be used independently of the other.

¹ These test methods are under the jurisdiction of ASTM Committee D19 on Water and are the direct responsibility of Subcommittee D19.05 on Inorganic Constituents in Water.

Current edition approved Sept. 1, 2012. Published August 2007. Originally approved in 1959. Last previous edition approved in 2007 as D1687 – 02(2007)E01. DOI: 10.1520/D1687-12.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* For specific hazard statements, see 4.2 and Note 6 and Note 7.

2. Referenced Documents

2.1 ASTM Standards:²

D858 Test Methods for Manganese in Water
D1066 Practice for Sampling Steam
D1068 Test Methods for Iron in Water
D1129 Terminology Relating to Water
D1193 Specification for Reagent Water
D1688 Test Methods for Copper in Water
D1691 Test Methods for Zinc in Water
D1886 Test Methods for Nickel in Water
D2777 Practice for Determination of Precision and Bias of Applicable Test Methods of Committee D19 on Water
D3370 Practices for Sampling Water from Closed Conduits
D3557 Test Methods for Cadmium in Water
D3558 Test Methods for Cobalt in Water
D3559 Test Methods for Lead in Water
D3919 Practice for Measuring Trace Elements in Water by Graphite Furnace Atomic Absorption Spectrophotometry
D4691 Practice for Measuring Elements in Water by Flame Atomic Absorption Spectrophotometry
D4841 Practice for Estimation of Holding Time for Water Samples Containing Organic and Inorganic Constituents
D5810 Guide for Spiking into Aqueous Samples
D5847 Practice for Writing Quality Control Specifications for Standard Test Methods for Water Analysis
E60 Practice for Analysis of Metals, Ores, and Related Materials by Spectrophotometry
E275 Practice for Describing and Measuring Performance of Ultraviolet and Visible Spectrophotometers

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

*A Summary of Changes section appears at the end of this standard

3. Terminology

3.1 *Definitions*—For definitions of terms used in these test methods, refer to Terminology **D1129**.

3.2 *Definitions of Terms Specific to This Standard:*

3.2.1 *Laboratory Control Sample, n*—a solution with the certified concentration(s) of the analytes.

4. Significance and Use

4.1 Hexavalent chromium salts are used extensively in metal finishing and plating applications, in anodizing aluminum, and in the manufacture of paints, dyes, explosives, and ceramics. Trivalent chromium salts are used as mordants in textile dyeing, in the ceramic and glass industry, in the leather industry as a tanning agent, and in photography. Chromium may be present in wastewater from these industries and may also be discharged from chromate-treated cooling waters.

4.2 The hexavalent state of chromium is toxic to humans, animals, and aquatic life. It can produce lung tumors when inhaled and readily induces skin sensitization.

5. Purity of Reagents

5.1 Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society³ where such specifications are available. Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

5.2 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean reagent water conforming to Specification **D1193**, Type I, II, or III water. Type I is preferred and more commonly used. Type II water was specified at the time of round robin testing of these test methods.

NOTE 1—The user must ensure the type of reagent water chosen is sufficiently free of interferences. The water should be analyzed using the test method.

6. Sampling

6.1 Collect the sample in accordance with the applicable ASTM standard as follows: Practice **D1066**, or Practice **D3370**. The holding time for the samples may be calculated in accordance with Practice **D4841**.

6.2 Samples to be analyzed by Test Method A should be stabilized upon collection by addition of sodium hydroxide solution to a pH greater than or equal to 8, or analyzed immediately. Minor delays in stabilization or analyses of samples containing sulfur reduction compounds can produce significant loss in hexavalent chromium. Acidic samples con-

taining hypobromite, persulfate, or chlorine could oxidize trivalent chromium, if present, to hexavalent form upon preservation, resulting in a positive interference. When the presence of these oxidizing compounds is suspected, samples should not be preserved but analyzed immediately. It will be evident that in this case, the simultaneous presence of reducing compounds could not be anticipated.

6.3 Samples to be analyzed by Test Methods B and C shall be preserved by addition of HNO_3 (sp gr 1.42) to pH of 2 or less immediately at the time of collection, normally about 2 mL HNO_3 /L. If only dissolved chromium is to be determined, the sample shall be filtered through a 0.45- μm membrane filter before acidification.

NOTE 2—Alternatively, the pH may be adjusted in the laboratory if the sample is returned within 14 days. This could reduce hazards of working with acids in the field when appropriate.

TEST METHOD A—PHOTOMETRIC DIPHENYLCARBOHYDRAZIDE

7. Scope

7.1 This test method covers the determination of dissolved hexavalent chromium in water.

7.2 The test method is applicable in the range from 0.01 to 0.5 mg/L chromium. The range may be extended by appropriate sample dilution.

7.3 This test method has been used successfully with reagent grade water Types I, II, and III, tap water, 10 % NaCl solution, treated water from a synthetic organic industrial plant that meets NPDES permit requirements, EPA-extraction procedure leachate water, process water, lake water, effluent from treatment that is, lime neutralization and precipitation of spent pickle liquor and associated rinse water from stainless steel pickling. It is the responsibility of the user to ensure the validity of the test method to waters of untested matrices.

8. Summary of Test Method

8.1 Hexavalent chromium reacts with 1.5-diphenylcarbohydrazide (s-diphenylcarbazine) in an acid medium to produce a reddish-purple color. The intensity of the color formed is proportional to the hexavalent chromium concentration.

9. Interferences

9.1 Vanadium, titanium, or iron present at concentrations of 5 mg/L yield a 10 to 30 % reduction in recovery of chromium. Copper at 100 mg/L yields a 20 to 30 % reduction in recovery in the presence of sulfate. Mercury gives a blue-purple color, but the reaction is not very sensitive at the pH employed for the test.

9.2 Nitrite concentrations in excess of 10 mg/L as NO_2 yield low test results. Sulfamic acid may be added (~ 10.1 g) prior to the addition of diphenylcarbazine solution to minimize nitrite interference. Add sulfamic acid only when the presence of nitrite has been positively established. Excess sulfamic acid itself creates a slightly positive interference.

³ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

9.3 Sulfide and sulfite reduce chromate in an acid medium to give low results.

9.4 Several sample matrices have been identified which produce a yellow-orange complex that interferes with this quantification. When this occurs, it may be remedied by inverting the indicator-buffer sequence. In these cases the analyst should evaluate the matrix effect with the additions of spikes. (Guide [D5810](#))

9.5 Although each interferent has been reported, most of the common interferences are eliminated by the preservation procedure at the time of collection. The potentially interfering metals are precipitated and the reducing effect of sulfur compounds has been overcome.

10. Apparatus

10.1 *Photometer*—Spectrophotometer or filter photometer suitable for use at 540 nm and equipped with a cell having a minimum path length of 10 mm. The photometers and photometric practice prescribed in this test method shall conform to Practice [E60](#). Spectrophotometers and spectrophotometric practice shall conform to Practice [E275](#).

11. Reagents

11.1 *Chromium Solution, Stock* (1 mL = 0.10 mg Cr)—Dissolve 0.2828 g of potassium dichromate ($K_2Cr_2O_7$) that has been oven dried at 105°C for 1 h in water. Dilute to 1 L with water. Alternatively, certified stock solutions are commercially available through chemical supply vendors and may be used.

11.2 *Chromium Solution, Standard* (1 mL = 0.001 mg Cr)—Dilute 10.0 mL of chromium stock solution (see [11.1](#)) to 1 L with water.

11.3 *Diphenylcarbazide Indicator Solution*—Dissolve 0.25 g of powdered 1,5-diphenylcarbohydrazide in 100 mL of acetone. Store in an amber glass-stoppered flask at 4°C when not in use. This solution is stable for about one week when kept refrigerated. Prepare fresh reagent when the solution becomes discolored.

NOTE 3—Allow the indicator solution to warm to room temperature before use.

11.4 *Phosphoric Acid* (1 + 1)—Dilute 500 mL of concentrated phosphoric acid (sp gr 1.69) to 1 L with water.

11.5 *Phosphoric Acid* (1 + 19)—Dilute 50 mL of concentrated phosphoric acid (sp gr 1.69) to 1 L with water.

11.6 *Sodium Hydroxide Solution* (40 mg/L)—Dissolve 40 mg of sodium hydroxide (NaOH) in water. Cool and dilute to 1 L. This solution is used for sample preservation.

11.7 *Sulfamic Acid* (NH_2SO_3H)—Crystals.

12. Calibration

12.1 Prepare a series of at least four standard solutions containing from 0 to 0.50 mg/L of chromium by diluting measured volumes of the standard chromium solution (see [11.2](#)) to 100 mL with water in separate volumetric flasks.

12.2 Transfer 50 mL of each prepared standard solution to separate 125-mL Erlenmeyer flasks and proceed with [13.3-13.6](#).

12.3 Prepare a calibration curve by plotting milligrams per liter of chromium versus absorbance on linear graph paper.

12.4 A calibration curve must be prepared for each photometer. A recalibration must be made if any alterations of the instrument are made or if new reagents are prepared. At the least, a blank and three chromium standard solutions must be analyzed to verify the original test calibration each time the test is performed.

13. Procedure

13.1 Filter a portion of the sample through a 0.45-μm membrane filter and adjust the pH into the 8 to 8.5 range if it is greater than 8.5 with a few drops of the phosphoric acid solution (1 + 19).

13.2 Transfer 50.0 mL of the filtered sample, or a smaller aliquot of sample diluted to 50.0 mL, to a 125-mL Erlenmeyer flask.

13.3 Add 2.0 mL of the diphenylcarbazide solution to each solution and swirl to mix.

NOTE 4—If the sample is colored, prepare a separate aliquot as described in [13.1](#) and [13.2](#). Add 2.0 mL of acetone instead of diphenylcarbazide solution and proceed with [13.4](#) and [13.5](#). Use this solution as the sample blank.

13.4 Immediately add 5.0 mL of phosphoric acid solution (1 + 1) to each solution and swirl to mix.

13.5 Permit the solutions to stand 15 min for full color development. Measure the absorbance within 30 min after the addition of the diphenylcarbazide solution at 540 nm with a cell having a minimum path length of 10 mm.

13.6 Determine milligrams per liter of chromium as Cr^{+6} in the test sample by referring the absorbance to the prepared calibration curve (see [12.3](#)).

14. Calculation

14.1 Calculate the hexavalent chromium concentration as follows:

$$Cr^{+6}, \text{ mg/L} = (W_S - W_B)(50/S) \quad (1)$$

where:

W_S = chromium found in the sample, mg/L (see [13.6](#)),
 W_B = chromium found in the sample blank, mg/L (see [13.6](#)),
 and
 S = volume of sample used, mL (see [13.2](#)).

15. Precision and Bias

15.1 The collaborative test data were obtained on reagent grade water Types I, II, and III, tap water, 10 % NaCl solution, treated water from a synthetic organic industrial plant which meets NPDES permit requirements, EPA-extraction procedure leachate water, process water, lake water, effluent from treatment, that is, lime neutralization and precipitation of spent pickle liquor and associated rinse water from stainless steel pickling.

TABLE 1 Determination of Bias and Precision, Photometric Diphenylcarbohydrazide

	Amount Added, mg/L	Mean Recovery ($\bar{X}$), mg/L	$\pm$ Bias	$\pm$ % Bias	Statistically Significant at 5 % Level	S_T	S_o
Reagent water	0.010	0.0125	+ 0.0025	+ 25.0	yes	0.006	0.0031
	0.050	0.0502	+ 0.0002	+ 0.40	no	0.007	0.0053
	0.350	0.3484	–0.0016	–0.46	no	0.022	0.0130
	0.500	0.4964	–0.0036	–0.72	no	0.022	0.0139
Water of choice	0.010	0.0112	+ 0.0012	+ 12.0	no	0.005	0.0025
	0.050	0.0468	–0.0032	–6.40	yes	0.007	0.0042
	0.350	0.3378	–0.0122	–3.49	yes	0.026	0.0159
	0.500	0.4776	–0.0224	–4.48	yes	0.038	0.0204
Leachate	0.010	0.0148	+ 0.0048	+ 48.0	yes	0.008	0.0037
	0.050	0.0513	+ 0.0013	+ 2.60	no	0.009	0.0062
	0.350	0.3422	–0.0078	–2.23	yes	0.015	0.0093
	0.500	0.4887	–0.0113	–2.26	yes	0.025	0.0130

15.2 Single-operator and overall precision of this test method within its designated range and recovery data for the above waters for 16 laboratories, which include a total of 16 operators analyzing each sample on three different days, is given in Table 1.

15.3 Single-operator and overall precision of this test method within its designated range and recovery data for a prepared leachate water for 8 laboratories, which include a total of 8 operators analyzing each sample on three different days, is also given in Table 1.

15.4 It is the user's responsibility to ensure the validity of the test method for waters of untested matrices.

15.5 Precision and bias for this test method conforms to Practice D2777 – 77, which was in place at the time of collaborative testing. Under the allowances made in 1.4 of D2777 – 08, these precision and bias data do meet existing requirements for interlaboratory studies of Committee D19 test methods.

TEST METHOD B—ATOMIC ABSORPTION, DIRECT

16. Scope

16.1 This test method covers the determination of dissolved and total recoverable chromium in most waters, wastewaters, and brines.

16.2 The test method is applicable in the range from 0.1 to 10 mg/L of chromium. The range may be extended to concentrations greater than 10 mg/L by dilution of the sample.

16.3 It is the user's responsibility to ensure the validity of this test method for waters of untested matrices.

17. Summary of Test Method

17.1 Chromium is determined by atomic absorption spectrophotometry. Dissolved chromium is determined by aspirating a portion of the filtered sample directly with no pretreatment. Total recoverable chromium is determined by aspirating the sample following hydrochloric-nitric acid digestion and filtration. The same digestion procedure is used to determine total recoverable cadmium (Test Methods D3557), nickel (Test Methods D1886), cobalt (Test Methods D3558), copper (Test Methods D1688), iron (Test Methods D1068), lead (Test

Methods D3559), manganese (Test Methods D858) and zinc (Test Methods D1691).

18. Interferences

18.1 Iron, nickel, and cobalt at 100 µg/L and magnesium at 30 mg/L interfere by depressing the absorption of chromium. These interferences are eliminated in solutions containing 10,000 mg/L of 8-hydroxyquinoline. Samples adjusted to this concentration show no interference from 700 mg/L of iron and 10 mg/L each of nickel and cobalt, or from 1000 mg/L of magnesium.

18.2 Potassium above 500 mg/L enhances the chromium absorption.

18.3 Sodium, sulfate, and chloride (9000 mg/L each), calcium and magnesium (4000 mg/L each), nitrate (2000 mg/L), and cadmium, lead, copper, and zinc, (10 mg/L each) do not interfere.

19. Apparatus and Materials

19.1 *Atomic Absorption Spectrophotometer*, for use at 357.9 nm. A general guide for the use of flame atomic absorption applications is given in Practice D4691.

NOTE 5—The manufacturer's instructions should be followed for all instrumental parameters. Wavelengths other than 357.9 nm may be used if they have been determined to be equally suitable.

19.1.1 *Chromium Hollow Cathode Lamp*, multielement hollow-cathode lamps.

19.2 *Oxidant*:

19.2.1 *Air* that has been passed through a suitable filter to remove oil, water, and other foreign substances, is the usual oxidant.

19.2.2 *Nitrous Oxide*, medical grade, is satisfactory.

19.3 *Fuel*:

19.3.1 *Acetylene*—Standard, commercially available acetylene is the usual fuel. Acetone, always present in acetylene cylinders, can affect analytical results. The cylinder should be replaced at 345 kPa (50 psi).

NOTE 6—**Warning:** "Purified" grade acetylene containing a special proprietary solvent rather than acetone should not be used with poly(vinyl chloride) tubing as weakening of the tubing walls can cause a hazardous situation.

19.4 *Pressure-Reducing Valves*—The supplies of fuel and oxidant shall be maintained at pressures somewhat higher than the controlled operating pressure of the instrument by suitable valves.

20. Reagents

20.1 *Chromium Solution, Stock* (1 mL = 1.0 mg Cr)—Dissolve 2.828 g of primary standard potassium dichromate ($K_2Cr_2O_7$) in 200 mL of water and dilute to 1 L. Alternatively, certified stock solutions are commercially available through chemical supply vendors and may be used.

20.2 *Chromium Solution, Standard* (1 mL = 0.1 mg Cr)—Dilute 100.0 mL of the chromium stock solution and 1 mL of HNO_3 (sp gr 1.42) to 1 L with water.

20.3 *Hydrochloric Acid* (sp gr 1.19)—Concentrated hydrochloric acid (HCl).

NOTE 7—If a high reagent blank is obtained, distill the HCl or use a spectrograde acid. **Caution**—When HCl is distilled an azeotropic mixture is obtained (approximately 6 N HCl). Therefore, whenever concentrated HCl is specified for the preparation of a reagent or in the procedure, use double the amount specified if a distilled acid is used.

20.4 *8-Hydroxyquinoline Solution* (100 g/L)—Dissolve 50 g of 8-hydroxyquinoline in 35 mL of HCl (sp gr 1.19). Warm the mixture gently on a hot plate to facilitate dissolution. Transfer to a 500-mL volumetric flask and bring to volume with the careful addition of water. Use a hood.

20.5 *Nitric Acid* (sp gr 1.42)—Concentrated nitric acid (HNO_3).

NOTE 8—If a high reagent blank is obtained, distill the HNO_3 or use a spectrograde acid.

20.6 Nitric Acid (1 + 499)—Add 1 volume of HNO_3 (sp gr 1.42) to 499 volumes of water.

21. Standardization

21.1 Prepare 100 mL each of a blank and at least four standard solutions, containing 1 mL of 8-hydroxyquinoline solution (100 g/L)/10 mL of standard, to bracket the expected chromium concentration range of the samples to be analyzed, by diluting the standard chromium solution (see 20.2) with HNO_3 (1 + 499). Prepare the standards each time the test is to be performed.

21.2 To determine the total recoverable chromium, add 0.5 mL of HNO_3 (sp gr 1.42) and proceed as directed in 22.2–22.4. To determine dissolved chromium, proceed with 21.3.

21.3 Aspirate the blank and standards and record the absorbance or concentration at 357.9 nm. Aspirate HNO_3 (1 + 499) between each standard.

21.4 Prepare an analytical curve by plotting the absorbance versus concentration for each standard on linear graph paper. Alternatively, read directly in concentration if this capability is provided with the instrument.

22. Procedure

22.1 Measure 100.0 mL of a well-mixed acidified sample into a 125-mL beaker or flask.

NOTE 9—If only dissolved chromium is to be determined, start with 22.5.

22.2 Add 5 mL of HCl (sp gr 1.19) to each sample.

22.3 Heat the samples on a steam bath or hotplate in a well-ventilated hood until the volume has been reduced to 15 to 20 mL, making certain that the samples do not boil.

NOTE 10—When analyzing brines and samples containing appreciable amounts of suspended matter or dissolved solids, the amount of reduction in the volume is left to the discretion of the analyst.

NOTE 11—Many laboratories have found block digestion systems a useful way to digest samples for trace metals analysis. Systems typically consist of either a metal or graphite block with wells to hold digestion tubes. The block temperature controller must be able to maintain uniformity of temperature across all positions of the block. For trace metals analysis, the digestion tubes should be constructed of polypropylene and have a volume accuracy of at least 0.5%. All lots of tubes should come with a certificate of analysis to demonstrate suitability for their intended purpose.

22.4 Cool and filter the samples through a suitable filter such as fine-textured, acid-washed, ashless paper, into 100-mL volumetric flasks. Wash the filter paper two to three times with water and bring to volume.

22.5 Pipette 10.0 mL of sample into a 50-mL beaker and add 1.0 mL of 8-hydroxyquinoline solution.

22.6 Aspirate each filtered and acidified sample and determine its absorbance or concentration. Aspirate HNO_3 (1 + 499) between each sample.

23. Calculation

23.1 Calculate the concentration of chromium in the sample, in milligrams per liter, using the analytical curve prepared in 21.4. Alternatively, read directly in concentration if this capability is provided with the instrument.

24. Precision and Bias⁴

24.1 The overall precision (S_T) of this test method within its designated range for six laboratories, which include a total of nine operators analyzing each sample on three different days, varies linearly with the chromium concentration, X , in milligrams per liter.

24.1.1 For reagent water:

$$S_T = 0.097X + 0.010 \quad (2)$$

24.1.2 For selected water matrices:

$$S_T = 0.079X + 0.019 \quad (3)$$

where:

S_T = overall precision, mg/L, and
 X = concentration of chromium, mg/L.

24.2 Single-operator precision did not differ significantly from overall precision.

24.3 Recoveries of known amounts of chromium from reagent water and selected water matrices are given in Table 2.

⁴ Supporting data are available from ASTM Headquarters. Request RR:D19-1036.

TABLE 2 Determination of Bias, Atomic Absorption, Direct

	Amount Added, mg/L	Amount Found, mg/L	Bias	Bias, %	Statistically Significant (95 % confidence level)
Reagent water	0.4	0.399	−0.001	−0.25	no
	3.0	2.89	−0.11	−3.7	no
	7.0	6.99	−0.01	−0.14	no
Selected water	0.4	0.425	+ 0.025	+ 6.2	yes
	3.0	3.095	+ 0.095	+ 3.2	no
matrices	7.0	7.180	+ 0.180	+ 2.6	no

24.4 The selected waters used in this study are not available. It is the user's responsibility to ensure the validity of the test method for waters of untested matrices.

24.5 Precision and bias for this test method conforms to Practice D2777 – 77, which was in place at the time of collaborative testing. Under the allowances made in 1.4 of D2777 – 08, these precision and bias data do meet existing requirements for interlaboratory studies of Committee D19 test methods.

TEST METHOD C—ATOMIC ABSORPTION, GRAPHITE FURNACE

25. Scope

25.1 This test method covers the determination of dissolved and total recoverable chromium in most waters and wastewaters.

25.2 This test method is applicable in the range from 5 to 100 µg/L of chromium (Refer to Practice D3919, Footnote in Table 1) based on a 20-µL sample size. The range can be increased or decreased by varying the volume of sample injected or the instrumental settings. High concentrations may be diluted but preferably should be analyzed by direct aspiration atomic absorption spectrophotometry.

25.3 This test method has been used successfully with reagent water, stack scrubber water, lake water, filtered tap water, river water, condensate from medium BTU coal gasification process, well water, and production plant water. It is the user's responsibility to ensure the validity of the test method for waters of untested matrices.

26. Summary of Test Method

26.1 Chromium is determined by an atomic absorption spectrophotometer used in conjunction with a graphite furnace. A sample is placed in a graphite tube, evaporated to dryness, charred (pyrolyzed or ashed), and atomized. Since the graphite furnace uses the sample much more efficiently than flame atomization, the detection of low concentrations of elements in small sample volumes is possible. Finally, the absorption signal during atomization is recorded and compared to standards. A general guide for the application of the graphite furnace is given in Practice D3919.

26.2 Dissolved chromium is determined on a filtered sample with no pretreatment.

26.3 Total recoverable chromium is determined acid digestion and filtration. Because chlorides interfere with furnace

procedures for some metals, the use of hydrochloric acid in any digestion or solubilization step shall be avoided. If suspended material is not present, this digestion and filtration may be omitted.

27. Interferences

27.1 For a complete discussion on general interferences with furnace procedures, refer to Practice D3919.

28. Apparatus and Materials

28.1 *Atomic Absorption Spectrophotometer*, for use at 357.9 nm with background correction. See Note 12 and Note 13.

NOTE 12—A wavelength other than 357.9 nm may be used if it has been determined to be suitable. Greater linearity may be obtained at high concentrations by using a less sensitive wavelength.

NOTE 13—The manufacturer's instructions should be followed for all instrumental parameters.

28.2 *Chromium Light Source*, chromium hollow-cathode lamp. A single-element lamp is preferred, but multielement lamps may be used.

28.3 *Graphite Furnace*, capable of reaching temperatures sufficient to atomize the element of interest.

28.4 *Graphite Tubes*, compatible with furnace device. In this instance and to eliminate the possible formation of carbides, pyrolytically coated graphite tubes are recommended.

28.5 *Pipettes*, microlitre with disposable tips. Sizes may range from 1 µL to 100 µL, as required.

28.6 *Argon*, standard, welders grade, commercially available. Hydrogen may also be used if recommended by the instrument manufacturer.

28.7 *Data Storage and Reduction Device*—Computer and microprocessor controlled devices, or a strip chart recorder, shall be utilized for data collection, reduction, storage, and problem recognition (drift, incomplete atomization, changes in sensitivity, etc.). Strip chart recorders shall have a full scale deflection time of 0.2 s or less to ensure accuracy.

28.8 *Automatic Sampling*, may be used if available.

29. Reagents

29.1 *Chromium Solution, Stock* (1.0 mL = 1.0 mg Cr)—See 20.1.

29.2 *Chromium Intermediate Solution*, (1.0 mL = 10 µg Cr)—Dilute 10.0 mL of chromium stock solution (20.1) and 1 mL of HNO₃ (sp gr 1.42) to 1 L with water.

29.3 *Chromium Solution, Standard* (1.0 mL = 0.10 µg Cr)—Dilute 10.0 mL of chromium intermediate solution (29.2) and 1 mL of HNO₃ (sp. gr. 1.42) to 1 L with water. This standard is used to prepare working standards at the time of the analysis.

29.4 *Nitric Acid* (sp gr 1.42)—Concentrated nitric acid (HNO₃).

30. Standardization

30.1 Initially, set the instrument in accordance with the manufacturer's specifications. Follow the general instructions in Practice D3919.

31. Procedure

31.1 Clean all glassware to be used for preparation of standard solutions or in the digestion step, or both, by soaking the glassware overnight in HNO_3 (1 + 1) and then rinsing with water.

NOTE 14—Traces of chromium may be sometimes found in laboratory distilled water. It is the responsibility of the analyst to make certain, through analysis of appropriate blanks, that water used for diluting and rinsing is free from detectable amounts of chromium.

31.2 Measure 100.0 mL of each standard and well-mixed sample into 125-mL beakers or flasks.

31.3 For total recoverable chromium, add 5 mL HNO_3 (sp gr 1.42) to each standard and sample and proceed as directed in 31.4–31.6. If only dissolved chromium is to be determined, filter the unacidified sample through a 0.45- μm membrane filter, acidify, and proceed to 31.6.

31.4 Heat the samples at 95°C on a steam bath or hotplate in a well-ventilated fume hood until the volume has been reduced to 15 to 20 mL, making certain that the samples do not boil (see Note 10).

NOTE 15—Many laboratories have found block digestion systems a useful way to digest samples for trace metals analysis. Systems typically consist of either a metal or graphite block with wells to hold digestion tubes. The block temperature controller must be able to maintain uniformity of temperature across all positions of the block. For trace metals analysis, the digestion tubes should be constructed of polypropylene and have a volume accuracy of at least 0.5%. All lots of tubes should come with a certificate of analysis to demonstrate suitability for their intended purpose.

31.5 Cool and filter the sample through a suitable filter such as fine-textured, acid washed, ashless paper, into a 100-mL volumetric flask. Wash the filter paper two or three times with water and bring to volume. See Note 16.

NOTE 16—If suspended material is not present, this filtration may be omitted. However, the sample must still be diluted to 100 mL.

31.6 Inject a measured aliquot of sample into the furnace device following the directions as provided by the particular instrument manufacturer. Refer to Practice D3919.

32. Calculation

32.1 Determine the concentration of chromium in each sample by referring to Practice D3919.

33. Precision and Bias⁵

33.1 The precision of this test method was tested by 15 laboratories in reagent water, stack scrubber water, lake water, filtered tap water, river water, tap water, condensate from a medium BTU coal gasification process, well water, and production plant water. The round-robin study upon which these precision data are based involved the determination of numerous other metals. Replicate determinations were not requested in order to simplify the study and ensure generation of data for all metals. Thus, no single-operator precision data can be calculated. Bias data and overall precision data are given in Table 3 and Table 4.

⁵ Supporting data are available from ASTM Headquarters. Request RR:D19-1103.

TABLE 3 Determination of Bias and Overall Precision in Reagent Water, Atomic Absorption, Graphite Furnace

Amount Added, $\mu\text{g/L}$	Amount Found, $\mu\text{g/L}$	s_T , $\mu\text{g/L}$	Bias, $\mu\text{g/L}$	Bias, %	Statistically Significant
8.0	8.1	1.78	+ 0.1	+ 1.25	no
10.0	9.5	2.28	−0.5	−5.0	no
28.0	27.9	3.93	−0.1	−0.36	no

TABLE 4 Determination of Bias and Overall Precision in Water of Choice, Atomic Absorption, Graphite Furnace

Amount Added, $\mu\text{g/L}$	Amount Found, $\mu\text{g/L}$	s_T , $\mu\text{g/L}$	Bias, $\mu\text{g/L}$	Bias, %	Statistically Significant
8.0	6.67	1.85	−1.33	−16.6	yes
10.0	10.83	3.42	+ 0.83	+ 8.3	no
28.0	28.2	5.0	+ 0.2	+ 0.7	no

33.2 These data may not apply to waters of other matrices, therefore, it is the responsibility of the analyst to ensure the validity of the test method in a particular matrix.

33.3 Precision and bias for this test method conforms to Practice Practice D2777 – 77, which was in place at the time of collaborative testing. Under the allowances made in 1.4 of Practice D2777 – 08, these precision and bias data do meet existing requirements for interlaboratory studies of Committee D19 test methods.

34. Quality Control (QC)

34.1 The following quality control information is recommended for the determination of chromium in water.

34.2 For each method the instrument shall be calibrated using a minimum of four calibration standards and a calibration blank. The calibration correlation coefficient shall be equal to or greater than 0.990. In addition to the initial calibration blank, a calibration blank shall be analyzed at the end of the batch run to ensure contamination was not a problem during the batch analysis.

34.3 An instrument check standard shall be analyzed at a minimum frequency of 10 % throughout the batch analysis. The value of the instrument check standard shall fall between 80 % and 120 % of the true value.

34.4 Two method blanks shall be prepared ensuring that an adequate method blank volume is present for a minimum of seven repetitive analyses. The standard deviation of the method blank is used to determine the minimum detectable concentration of each sample and control in the batch.

34.5 A Laboratory Control Sample shall be analyzed with each batch of samples at a minimum frequency of 10 %.

34.6 If the QC for the sample batch is not within the established control limits, reanalyze the samples or qualify the results with the appropriate flags, or both. (Practice D5847)

34.7 Blind control samples should be submitted by an outside agency in order to determine the laboratory performance capabilities.

35. Keywords

35.1 atomic absorption; chromium; graphite furnace;
hexavalent chromium; photometric water

SUMMARY OF CHANGES

Committee D19 has identified the location of selected changes to this standard since the last issue (D1687 – 02(2007)E01) that may impact the use of this standard.

- | | |
|--|---|
| (1) The SI statement was added to section 1. | (5) Reagent references were added to sections 8, 9, 10, 11, 12. |
| (2) D1192 was removed from section 2 and section 6.1. | (6) Sections 22 and 31 were modified to include note about the |
| (3) Section 6 was modified to allow for pH of the samples in the laboratory. | use of block digestion systems. |
| (4) Sections 11.1 and 20.1 were modified to allow for commercial standards. | |

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the ASTM website (www.astm.org/COPYRIGHT/).