



Designation: D2887 – 13

Designation: 406

Standard Test Method for Boiling Range Distribution of Petroleum Fractions by Gas Chromatography^{1,2}

This standard is issued under the fixed designation D2887; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

This standard has been approved for use by agencies of the Department of Defense.

1. Scope*

1.1 This test method covers the determination of the boiling range distribution of petroleum products. The test method is applicable to petroleum products and fractions having a final boiling point of 538°C (1000°F) or lower at atmospheric pressure as measured by this test method. This test method is limited to samples having a *boiling range* greater than 55.5°C (100°F), and having a vapor pressure sufficiently low to permit sampling at ambient temperature.

NOTE 1—Since a *boiling range* is the difference between two temperatures, only the constant of 1.8°F/°C is used in the conversion of the temperature range from one system of units to another.

1.1.1 *Procedure A (Sections 6-14)*—Allows a larger selection of columns and analysis conditions such as packed and capillary columns as well as a Thermal Conductivity Detector in addition to the Flame Ionization Detector. Analysis times range from 14 to 60 min.

1.1.2 *Procedure B (Sections 15-23)*—Is restricted to only 3 capillary columns and requires no sample dilution. In addition, Procedure B is used not only for the sample types described in Procedure A but also for the analysis of samples containing biodiesel mixtures B5, B10, and B20. The analysis time, when using Procedure B (Accelerated D2887), is reduced to about 8 min.

1.2 This test method is not to be used for the analysis of gasoline samples or gasoline components. These types of samples must be analyzed by Test Method D3710.

1.3 The values stated in SI units are to be regarded as standard. The inch-pound units given in parentheses are for information only.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:³

- D86 Test Method for Distillation of Petroleum Products at Atmospheric Pressure
- D1160 Test Method for Distillation of Petroleum Products at Reduced Pressure
- D2892 Test Method for Distillation of Crude Petroleum (15-Theoretical Plate Column)
- D3710 Test Method for Boiling Range Distribution of Gasoline and Gasoline Fractions by Gas Chromatography
- D4057 Practice for Manual Sampling of Petroleum and Petroleum Products
- D4626 Practice for Calculation of Gas Chromatographic Response Factors
- D6708 Practice for Statistical Assessment and Improvement of Expected Agreement Between Two Test Methods that Purport to Measure the Same Property of a Material
- E260 Practice for Packed Column Gas Chromatography
- E355 Practice for Gas Chromatography Terms and Relationships
- E516 Practice for Testing Thermal Conductivity Detectors Used in Gas Chromatography
- E594 Practice for Testing Flame Ionization Detectors Used in Gas or Supercritical Fluid Chromatography

¹ This test method is under the jurisdiction of ASTM Committee D02 on Petroleum Products and Lubricants and is the direct responsibility of Subcommittee D02.04.0H on Chromatographic Distribution Methods.

Current edition approved May 1, 2013. Published June 2013. Originally approved in 1973. Last previous edition approved in 2012 as D2887-12. DOI: 10.1520/D2887-13.

² This standard has been developed through the cooperative effort between ASTM International and the Energy Institute, London. The EI and ASTM International logos imply that the ASTM International and EI standards are technically equivalent, but does not imply that both standards are editorially identical.

³ For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

*A Summary of Changes section appears at the end of this standard

3. Terminology

3.1 *Definitions*—This test method makes reference to many common gas chromatographic procedures, terms, and relationships. Detailed definitions of these can be found in Practices E260, E355, and E594.

3.2 *Definitions of Terms Specific to This Standard:*

3.2.1 *area slice, n*—the area, resulting from the integration of the chromatographic detector signal, within a specified retention time interval. In area slice mode (see 6.3.2), peak detection parameters are bypassed and the detector signal integral is recorded as area slices of consecutive, fixed duration time intervals.

3.2.2 *corrected area slice, n*—an area slice corrected for baseline offset, by subtraction of the exactly corresponding area slice in a previously recorded blank (non-sample) analysis.

3.2.3 *cumulative corrected area, n*—the accumulated sum of corrected area slices from the beginning of the analysis through a given retention time, ignoring any non-sample area (for example, solvent).

3.2.4 *final boiling point (FBP), n*—the temperature (corresponding to the retention time) at which a cumulative corrected area count equal to 99.5 % of the total sample area under the chromatogram is obtained.

3.2.5 *initial boiling point (IBP), n*—the temperature (corresponding to the retention time) at which a cumulative corrected area count equal to 0.5 % of the total sample area under the chromatogram is obtained.

3.2.6 *slice rate, n*—the time interval used to integrate the continuous (analog) chromatographic detector response during an analysis. The slice rate is expressed in hertz (for example, integrations or slices per second).

3.2.7 *slice time, n*—the time associated with the end of each contiguous area slice. The slice time is equal to the slice number divided by the slice rate.

3.2.8 *total sample area, n*—the cumulative corrected area, from the initial area point to the final area point, where the chromatographic signal is considered to have returned to baseline after complete sample elution.

3.3 *Abbreviations:*

3.3.1 A common abbreviation of hydrocarbon compounds is to designate the number of carbon atoms in the compound. A prefix is used to indicate the carbon chain form, while a subscripted suffix denotes the number of carbon atoms (for example, normal decane = $n\text{-C}_{10}$; isotetradecane = $i\text{-C}_{14}$).

4. Summary of Test Method

4.1 The boiling range distribution determination by distillation is simulated by the use of gas chromatography. A nonpolar packed or open tubular (capillary) gas chromatographic column is used to elute the hydrocarbon components of the sample in order of increasing boiling point. The column temperature is raised at a reproducible linear rate and the area under the chromatogram is recorded throughout the analysis. Boiling points are assigned to the time axis from a calibration curve obtained under the same chromatographic conditions by ana-

lyzing a known mixture of hydrocarbons covering the boiling range expected in the sample. From these data, the boiling range distribution can be obtained.

4.2 Procedure A and Procedure B yield essentially the same results. See Sections 14 and 23, and the accompanying research reports.

5. Significance and Use

5.1 The boiling range distribution of petroleum fractions provides an insight into the composition of feedstocks and products related to petroleum refining processes. The gas chromatographic simulation of this determination can be used to replace conventional distillation methods for control of refining operations. This test method can be used for product specification testing with the mutual agreement of interested parties.

5.2 Boiling range distributions obtained by this test method are essentially equivalent to those obtained by true boiling point (TBP) distillation (see Test Method D2892). They are not equivalent to results from low efficiency distillations such as those obtained with Test Method D86 or D1160.

5.3 Procedure B was tested with biodiesel mixtures and reports the Boiling Point Distribution of FAME esters of vegetable and animal origin mixed with ultra low sulfur diesel.

Procedure A

6. Apparatus

6.1 *Chromatograph*—The gas chromatograph used must have the following performance characteristics:

6.1.1 *Detector*—Either a flame ionization or a thermal conductivity detector may be used. The detector must have sufficient sensitivity to detect 1.0 % dodecane with a peak height of at least 10 % of full scale on the recorder under conditions prescribed in this test method and without loss of resolution as defined in 9.3.1. When operating at this sensitivity level, detector stability must be such that a baseline drift of not more than 1 % of full scale per hour is obtained. The detector must be capable of operating continuously at a temperature equivalent to the maximum column temperature employed. Connection of the column to the detector must be such that no temperature below the column temperature exists.

NOTE 2—It is not desirable to operate a thermal conductivity detector at a temperature higher than the maximum column temperature employed. Operation at higher temperature generally contributes to higher noise levels and greater drift and can shorten the useful life of the detector.

6.1.2 *Column Temperature Programmer*—The chromatograph must be capable of linear programmed temperature operation over a range sufficient to establish a retention time of at least 1 min for the IBP and to elute compounds up to a boiling temperature of 538°C (1000°F) before reaching the upper end of the temperature program. The programming rate must be sufficiently reproducible to obtain retention time repeatability of 0.1 min (6 s) for each component in the calibration mixture described in 7.8.

6.1.3 *Cryogenic Column Cooling*—Column starting temperatures below ambient will be required if samples with IBPs of less than 93°C (200°F) are to be analyzed. This is typically

TABLE 1 Typical Operating Conditions

Packed Columns	1	2	3	4	Open Tubular Columns	5	6	7
Column length, m (ft)	1.2 (4)	1.5 (5)	0.5 (1.5)	0.6 (2)	Column length (m)	7.5	5	10
Column outside diameter, mm (in.)	6.4 (1/4)	3.2 (1/8)	3.2 (1/8)	6.4 (1/8)	Column inner diameter (mm)	0.53	0.53	0.53
Liquid phase	OV-1	SE-30	UC-W98	SE-30	Stationary phase	DB-1	HP-1	HP-1
Percent liquid phase	3	5	10	10	Stationary phase thickness (m)	1.5	0.88	2.65
Support material	S ^A	G ^B	P ^C	P ^C	Carrier gas	nitrogen	helium	helium
Support mesh size	60/80	60/80	80/100	60/80	Carrier gas flow rate, mL/min	30	12	12
Initial column temperature, °C	−20	−40	−30	−50	Initial column temperature, °C	40	35	35
Final column temperature, °C	360	350	360	390	Final column temperature, °C	340	350	350
Programming rate, °C/min	10	6.5	10	7.5	Programming rate, °C/min	10	10	20
Carrier gas	helium	helium	N ₂	helium	Detector	FID	FID	FID
Carrier gas flow, mL/min	40	30	25	60	Detector temperature, °C	350	380	370
Detector	TC	FID	FID	TC	Injector temperature, °C	340	cool on-column	cool on-column
Detector temperature, °C	360	370	360	390	Sample size, µL	0.5	1	0.1–0.2
Injection port temperature, °C	360	370	350	390	Sample concentration mass %	25	2	neat
Sample size, µ	4	0.3	1	5				

^A Diatoport S; silane treated.

^B Chromosorb G (AW-DMS).

^C Chromosorb P, acid washed.

provided by adding a source of either liquid carbon dioxide or liquid nitrogen, controlled through the oven temperature circuitry. Excessively low initial column temperature must be avoided to ensure that the stationary phase remains liquid. The initial temperature of the column should be only low enough to obtain a calibration curve meeting the specifications of the method.

6.1.4 Sample Inlet System—The sample inlet system must be capable of operating continuously at a temperature equivalent to the maximum column temperature employed, or provide for on-column injection with some means of programming the entire column, including the point of sample introduction, up to the maximum temperature required. Connection of the column to the sample inlet system must be such that no temperature below the column temperature exists.

6.1.5 Flow Controllers—The gas chromatograph must be equipped with mass flow controllers capable of maintaining carrier gas flow constant to $\pm 1\%$ over the full operating temperature range of the column. The inlet pressure of the carrier gas supplied to the gas chromatograph must be sufficiently high to compensate for the increase in column back-pressure as the column temperature is raised. An inlet pressure of 550 kPa (80 psig) has been found satisfactory with the packed columns described in Table 1. For open tubular columns, inlet pressures from 10 to 70 kPa (1.5 to 10 psig) have been found to be suitable.

6.1.6 Microsyringe—A microsyringe is needed for sample introduction.

NOTE 3—Automatic sampling devices or other sampling means, such as indium encapsulation, can be used provided: the system can be operated at a temperature sufficiently high to completely vaporize hydrocarbons with atmospheric boiling points of 538°C (1000°F), and the sampling system is connected to the chromatographic column avoiding any cold temperature zones.

6.2 Column—Any column and conditions may be used that provide separation of typical petroleum hydrocarbons in order of increasing boiling point and meet the column performance requirements of 9.3.1 and 9.3.3. Successfully used columns and conditions are given in Table 1.

6.3 Data Acquisition System:

6.3.1 Recorder—A 0 to 1 mV range recording potentiometer or equivalent, with a full-scale response time of 2 s or less may be used.

6.3.2 Integrator—Means must be provided for determining the accumulated area under the chromatogram. This can be done by means of an electronic integrator or computer-based chromatography data system. The integrator/computer system must have normal chromatographic software for measuring the retention time and areas of eluting peaks (peak detection mode). In addition, the system must be capable of converting the continuously integrated detector signal into area slices of fixed duration. These contiguous area slices, collected for the entire analysis, are stored for later processing. The electronic range of the integrator/computer (for example, 1 V, 10 V) must be within the linear range of the detector/electrometer system used. The system must be capable of subtracting the area slice of a blank run from the corresponding area slice of a sample run.

NOTE 4—Some gas chromatographs have an algorithm built into their operating software that allows a mathematical model of the baseline profile to be stored in memory. This profile is automatically subtracted from the detector signal on subsequent sample analyses to compensate for any baseline offset. Some integration systems also store and automatically subtract a blank analysis from subsequent analytical determinations.

7. Reagents and Materials

7.1 Purity of Reagents—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society where such specifications are available.⁴ Other grades may be used,

⁴ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For Suggestions on the testing of reagents not listed by the American Chemical Society, see *Annual Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

7.2 Liquid Phase for Columns—Methyl silicone gums and liquids provide the proper chromatographic hydrocarbon elution characteristics for this test method.

7.3 Solid Support for Packed Columns—Chromatographic grade diatomaceous earth solid support material within a particle size range from 60 to 100 sieve mesh size is recommended.

7.4 Carrier Gas—Helium or nitrogen of high purity. (**Warning**—Helium and nitrogen are compressed gases under high pressure.) Additional purification is recommended by the use of molecular sieves or other suitable agents to remove water, oxygen, and hydrocarbons. Available pressure must be sufficient to ensure a constant carrier gas flow rate (see 6.1.5).

7.5 Hydrogen—Hydrogen of high purity (for example, hydrocarbon free) is used as fuel for the flame ionization detector (FID). (**Warning**—Hydrogen is an extremely flammable gas under high pressure.)

7.6 Air—High purity (for example, hydrocarbon free) compressed air is used as the oxidant for the flame ionization detector (FID). (**Warning**—Compressed air is a gas under high pressure and supports combustion.)

7.7 Column Resolution Test Mixture—For packed columns, a nominal mixture of 1 mass % each of *n*-C₁₆ and *n*-C₁₈ paraffin in a suitable solvent, such as *n*-octane, for use in testing the column resolution. (**Warning**—*n*-octane is flammable and harmful if inhaled.) The calibration mixture specified in 7.8.2 may be used as a suitable alternative, provided the concentrations of the *n*-C₁₆ and *n*-C₁₈ components are nominally 1.0 mass % each. For open tubular columns, use the mixture specified in 7.8.3.

7.8 Calibration Mixture—An accurately weighed mixture of approximately equal mass quantities of *n*-hydrocarbons dissolved in carbon disulfide (CS₂). (**Warning**—Carbon disulfide is extremely volatile, flammable, and toxic.) The mixture shall cover the boiling range from *n*-C₅ to *n*-C₄₄, but does not need to include every carbon number (see Note 5).

7.8.1 At least one compound in the mixture must have a boiling point lower than the IBP of the sample and at least one compound in the mixture must have a boiling point higher than the FBP of the sample. Boiling points of *n*-paraffins are listed in Table 2.

7.8.1.1 If necessary, for the calibration mixture to have a compound with a boiling point below the IBP of the sample, propane or butane can be added to the calibration mixture, non-quantitatively, by bubbling the gaseous compound into the calibration mixture in a septum sealed vial using a gas syringe.

NOTE 5—Calibration mixtures containing normal paraffins with the carbon numbers 5, 6, 7, 8, 9, 10, 11, 12, 14, 15, 16, 17, 18, 20, 24, 28, 32, 36, 40, and 44 have been found to provide a sufficient number of points to generate a reliable calibration curve.

7.8.2 Packed Columns—The final concentration should be approximately ten parts of the *n*-paraffin mixture to one hundred parts of CS₂.

TABLE 2 Boiling Points of Normal Paraffins^{A,B}

Carbon Number	Boiling Point, °C	Boiling Point, °F	Carbon Number	Boiling Point, °C	Boiling Point, °F
1	−162	−259	23	380	716
2	−89	−127	24	391	736
3	−42	−44	25	402	755
4	0	31	26	412	774
5	36	97	27	422	791
6	69	156	28	431	808
7	98	209	29	440	825
8	126	258	30	449	840
9	151	303	31	458	856
10	174	345	32	466	870
11	196	385	33	474	885
12	216	421	34	481	898
13	235	456	35	489	912
14	254	488	36	496	925
15	271	519	37	503	937
16	287	548	38	509	948
17	302	576	39	516	961
18	316	601	40	522	972
19	330	626	41	528	982
20	344	651	42	534	993
21	356	674	43	540	1004
22	369	695	44	545	1013

^A API Project 44, October 31, 1972 is believed to have provided the original normal paraffin boiling point data that are listed in Table 2. However, over the years some of the data contained in both API Project 44 (Thermodynamics Research Center Hydrocarbon Project) and Test Method D2887 have changed, and they are no longer equivalent. Table 2 represents the current normal paraffin boiling point values accepted by Subcommittee D02.04 and found in all test methods under the jurisdiction of Section D02.04.0H.

^B Test Method D2887 has traditionally used *n*-paraffin boiling points rounded to the nearest whole degree for calibration. The boiling points listed in Table 2 are correct to the nearest whole number in both degrees Celsius and degrees Fahrenheit. However, if a conversion is made from one unit to the other and then rounded to a whole number, the result will not agree with the table value for a few carbon numbers. For example, the boiling point of *n*-heptane is 98.425°C, which is correctly rounded to 98°C in the table. However, converting 98.425°C gives 209.165°F, which rounds to 209°F, while converting 98°C gives 208.4°F, which rounds to 208°F. Carbon numbers 2, 4, 7, 8, 9, 13, 14, 15, 16, 25, 27, and 32 are affected by rounding.

7.8.3 Open Tubular Columns—The final concentration should be approximately one part of the *n*-paraffin mixture to one hundred parts of CS₂.

7.9 Reference Gas Oil No. 1 or No. 2—A reference sample that has been analyzed by laboratories participating in the test method cooperative study. Consensus values for the boiling range distribution of this sample are given in Tables 3 and 4.

8. Sampling

8.1 Samples to be analyzed by this test method must be obtained using the procedures outlined in Practice D4057.

8.2 The test specimen to be analyzed must be homogeneous and free of dust or undissolved material.

9. Preparation of Apparatus

9.1 Chromatograph—Place in service in accordance with the manufacturer's instructions. Typical operating conditions are shown in Table 1.

9.1.1 When a FID is used, regularly remove the deposits formed in the detector from combustion of the silicone liquid phase decomposition products. These deposits will change the response characteristics of the detector.

TABLE 3 Test Method D2887 Reference Gas Oil No. 1^A

% Off	Batch 1		Allowable Difference		Batch 2		Allowable Difference	
	°C	°F	°C	°F	°C	°F	°C	°F
IBP	114	238	7.5	13.6	115	240	7.6	13.7
5	143	289	3.6	6.6	151	304	3.8	6.8
10	169	336	4.0	7.3	176	348	4.1	7.4
15	196	384	4.4	8.0	201	393	4.5	8.1
20	221	429	4.8	8.7	224	435	4.9	8.7
25					243	470		
30	258	496	4.7	8.4	259	499	4.7	8.4
35					275	527		
40	287	548	4.3	7.7	289	552	4.3	7.7
45					302	576		
50	312	594	4.3	7.7	312	594	4.3	7.7
55					321	611	4.3	7.7
60	332	629	4.3	7.7	332	629	4.3	7.7
65	343	649	4.3	7.7	343	649	4.3	7.7
70	354	669	4.3	7.7	354	668	4.3	7.7
75	364	688	4.3	7.7	365	690	4.3	7.7
80	376	709	4.3	7.7	378	712	4.3	7.7
85	389	732	4.3	7.7	391	736	4.3	7.7
90	404	759	4.3	7.7	407	764	4.3	7.7
95	425	797	5.0	9.0	428	803	5.0	9.0
FBP	475	887	11.8	21.2	475	888	11.8	21.2

^A Consensus results for Batch 2 obtained from 30 laboratories in 1995 (supporting data have been filed at ASTM International Headquarters and may be obtained by requesting Research Report RR:D02-1407).

TABLE 4 Test Method D2887 Reference Gas Oil No. 2^A

% Off	Allowable Difference		Allowable Difference	
	°C	°F	°C	°F
IBP	106	223	7.0	12.6
5	173	343	4.1	7.4
10	196	384	4.4	8.0
15	216	420	4.7	8.5
20	233	452	5.0	9.0
25	251	483	...	...
30	267	512	4.8	8.6
35	283	541	...	...
40	298	568	4.3	7.7
45	310	590	...	...
50	321	610	4.3	7.7
55	331	629	4.3	7.7
60	342	647	4.3	7.7
65	350	662	4.3	7.7
70	358	677	4.3	7.7
75	368	694	4.3	7.7
80	378	712	4.3	7.7
85	390	734	4.3	7.7
90	406	763	4.3	7.7
95	431	808	5.0	9.0
FBP	496	925	11.8	21.2

^A Consensus results for Reference Gas Oil No. 2 obtained from 32 laboratories in 2009.

9.1.2 If the sample inlet system is heated above 300°C (572°F), a blank analysis must be made after a new septum is installed to ensure that no extraneous detector response is produced by septum bleed. At the sensitivity levels commonly employed in this test method, conditioning of the septum at the operating temperature of the sample inlet system for several hours will minimize this problem. A recommended practice is to change the septum at the end of a series of analyses rather than at the beginning of the series.

9.2 Column Preparation:

9.2.1 *Packed Columns*—Any satisfactory method that will produce a column meeting the requirements of 9.3.1 and 9.3.3 can be used. In general, use liquid phase loadings of 3 to 10 %. Condition the column at the maximum operating temperature

to reduce baseline shifts due to bleeding of the column substrate. The column can be conditioned very rapidly and effectively using the following procedure:

9.2.1.1 Connect the column to the inlet but leave the detector end free.

9.2.1.2 Purge the column thoroughly at ambient temperature with carrier gas.

9.2.1.3 Turn off the carrier gas and allow the column to depressurize completely.

9.2.1.4 Seal off the open end (detector) of the column with an appropriate fitting.

9.2.1.5 Raise the column temperature to the maximum operating temperature.




FIG. 1 Column Resolution Parameters

9.2.1.6 Hold the column at this temperature for at least 1 h with no flow through the column.

9.2.1.7 Cool the column to ambient temperature.

9.2.1.8 Remove the cap from the detector end of the column and turn the carrier gas back on.

9.2.1.9 Program the column temperature up to the maximum several times with normal carrier gas flow. Connect the free end of the column to the detector.

9.2.1.10 An alternative method of column conditioning that has been found effective for columns with an initial loading of 10 % liquid phase consists of purging the column with carrier gas at the normal flow rate while holding the column at the maximum operating temperature for 12 to 16 h, while detached from the detector.

9.2.2 *Open Tubular Columns*—Open tubular columns with cross-linked and bonded stationary phases are available from many manufacturers and are usually pre-conditioned. These columns have much lower column bleed than packed columns. Column conditioning is less critical with these columns but some conditioning may be necessary. The column can be conditioned very rapidly and effectively using the following procedure.

9.2.2.1 Once the open tubular column has been properly installed into the gas chromatograph and tested to be leak free, set the column and detector gas flows. Before heating the column, allow the system to purge with carrier gas at ambient temperature for at least 30 min.

9.2.2.2 Increase the oven temperature about 5 to 10°C per minute to the final operating temperature and hold for about 30 min.

9.2.2.3 Cycle the gas chromatograph several times through its temperature program until a stable baseline is obtained.

9.3 System Performance Specification:

9.3.1 *Column Resolution*—The column resolution, influenced by both the column physical parameters and operating conditions, affects the overall determination of boiling range distribution. Resolution is therefore specified to maintain equivalence between different systems (laboratories) employing this test method. Resolution is determined using Eq 1 and

the C₁₆ and C₁₈ paraffins from a column resolution test mixture analysis (see 7.7 and Section 10), and is illustrated in Fig. 1. Resolution (*R*) should be at least three, using the identical conditions employed for sample analyses:

$$R = 2(t_2 - t_1) / [1.699(w_2 + w_1)] \quad (1)$$

where:

R = resolution,

*t*₁ = time(s) for the *n*-C₁₆ peak maximum,

*t*₂ = time(s) for the *n*-C₁₈ peak maximum,

*w*₁ = peak width(s), at half height, of the *n*-C₁₆ peak, and

*w*₂ = peak width(s), at half height, of the *n*-C₁₈ peak.

9.3.2 *Detector Response Calibration*—This test method assumes that the detector response to petroleum hydrocarbons is proportional to the mass of individual components. This must be verified when the system is put in service, and whenever any changes are made to the system or operational parameters. Analyze the calibration mixture using the identical procedure to be used for the analysis of samples (see Section 10). Calculate the relative response factor for each *n*-paraffin (relative to *n*-decane) in accordance with Practice D4626 and Eq 2:

$$F_n = (M_n/A_n) / (M_{10}/A_{10}) \quad (2)$$

where:

*F*_{*n*} = relative response factor,

*M*_{*n*} = mass of the *n*-paraffin in the mixture,

*A*_{*n*} = peak area of the *n*-paraffin in the mixture,

*M*₁₀ = mass of the *n*-decane in the mixture, and

*A*₁₀ = peak area of the *n*-decane in the mixture.

The relative response factor (*F*_{*n*}) of each *n*-paraffin must not deviate from unity (1) by more than ±10 %.

9.3.3 *Column Elution Characteristics*—The column material, stationary phase, or other parameters can affect the elution order of non-paraffinic sample components, resulting in deviations from a TBP versus retention time relationship. If stationary phases other than those referenced in 7.3 are used, the retention times of a few alkylbenzenes (for example,

o-xylene, *n*-butyl-benzene, 1,3,5-triisopropylbenzene, *n*-decylbenzene, and tetradecylbenzene) across the boiling range should be analyzed to make certain that the column is separating in accordance with the boiling point order (see [Appendix X1](#)).

10. Calibration and Standardization

10.1 Analysis Sequence Protocol—Define and use a predetermined schedule of analysis events designed to achieve maximum reproducibility for these determinations. The schedule will include cooling the column oven to the initial starting temperature, equilibration time, sample injection and system start, analysis, and final upper temperature hold time.

10.1.1 After chromatographic conditions have been set to meet performance requirements, program the column temperature upward to the maximum temperature to be used and hold that temperature for the selected time. Following the analysis sequence protocol, cool the column to the initial starting temperature.

10.1.2 During the cool down and equilibration time, ready the integrator/computer system. If a retention time or detector response calibration is being performed, use the peak detection mode. For samples and baseline compensation determinations, use the area slice mode of integration. The recommended slice rate for this test method is given in [12.1.2](#). Other slice rates may be used if within the limits of 0.02 and 0.2 % of the retention time of the final calibration component (C_{44}). Larger slice rates may be used, as may be required for other reasons, if provision is made to accumulate (bunch) the slice data to within these limits prior to determination of the boiling range distribution.

10.1.3 At the exact time set by the schedule, inject either the calibration mixture or sample into the chromatograph; or make no injection (baseline blank). At the time of injection, start the chromatograph time cycle and the integrator/computer data acquisition. Follow the analysis sequence protocol for all subsequent repetitive analyses or calibrations. Since complete resolution of sample peaks is not expected, do not change the detector sensitivity setting during the analysis.

10.2 Baseline Compensation Analysis—A baseline compensation analysis, or baseline blank, is performed exactly like an analysis except no injection is made. A blank analysis must be performed at least once per day. The blank analysis is necessary due to the usual occurrence of chromatographic baseline instability and is subtracted from sample analyses to remove any nonsample slice area from the chromatographic data. The blank analysis is typically performed prior to sample analyses, but may be useful if determined between samples or at the end of a sample sequence to provide additional data regarding instrument operation or residual sample carryover from previous sample analyses. Attention must be given to all factors that influence baseline stability, such as column bleed, septum bleed, detector temperature control, constancy of carrier gas flow, leaks, instrument drift, and so forth. Periodic baseline blank analyses should be made, following the analysis sequence protocol, to give an indication of baseline stability.

NOTE 6—If automatic baseline correction (see [Note 4](#)) is provided by the gas chromatograph, further correction of area slices may not be

required. However, if an electronic offset is added to the signal after baseline compensation, additional area slice correction may be required in the form of offset subtraction. Consult the specific instrumentation instructions to determine if an offset is applied to the signal. If the algorithm used is unclear, the slice area data can be examined to determine if further correction is necessary. Determine if any offset has been added to the compensated signal by examining the corrected area slices of those time slices that precede the elution of any chromatographic unretained substance. If these corrected area slices (representing the true baseline) deviate from zero, subtract the average of these corrected area slices from each corrected area slice in the analysis.

10.3 Retention Time Versus Boiling Point Calibration—In order to analyze samples, a retention time versus boiling point calibration must be performed. Inject an appropriate aliquot (0.2 to 2.0 μL) of the calibration mixture (see [7.8](#)) into the chromatograph, using the analysis sequence protocol. Obtain a normal (peak detection) data record in order to determine the peak retention times and the peak areas for each component. Collect a time slice area record if a boiling range distribution report is desired.

10.3.1 Inspect the chromatogram of the calibration mixture for evidence of skewed (non-Gaussian shaped) peaks. Skewness is often an indication of overloading the sample capacity of the column that will result in displacement of the peak apex relative to nonoverloaded peaks. Distortion in retention time measurement and hence errors in boiling point temperature determination will be likely if column overloading occurs. The column liquid phase loading has a direct bearing on acceptable sample size. Reanalyze the calibration mixture using a smaller sample size or a more dilute solution to avoid peak distortion.

10.3.2 Prepare a calibration table based upon the results of the analysis of the calibration mixture by recording the time of each peak maximum and the boiling point temperature in degrees Celsius (or Fahrenheit) for every component in the mixture. *n*-Paraffin boiling point temperatures are listed in [Table 2](#).

10.3.3 Plot the retention time of each peak versus the corresponding normal boiling point temperature of that component in degrees Celsius (or Fahrenheit) as shown in [Fig. 2](#).

10.3.4 Ideally, the retention time versus boiling point temperature calibration plot would be linear, but it is impractical to operate the chromatograph such that curvature is eliminated completely. The greatest potential for deviation from linearity will be associated with the lower boiling point paraffins. They will elute from the column relatively fast and have the largest difference in boiling point temperature. In general, the lower the sample IBP, the lower will be the starting temperature of the analysis. Although extrapolation of the curve at the upper end is more accurate, calibration points must bracket the boiling range of the sample at both the low and high ends.

10.4 Reference Gas Oil Analysis—The Reference Gas Oil sample is used to verify both the chromatographic and calculation processes involved in this test method. Perform an analysis of the gas oil following the analysis sequence protocol. Collect the area slice data and provide a boiling point distribution report as in [Sections 12 and 13](#).

10.4.1 The results of this reference analysis must agree with the values given in [Table 3](#) within the range specified by the test method reproducibility (see [14.1.2](#)). If it does not meet the criteria in [Table 3](#), check that all hardware is operating properly




FIG. 2 Typical Calibration Curve

and all instrument settings are as recommended by the manufacturer. Rerun the retention boiling point calibration as described in 10.3.

10.4.2 Perform this reference gas oil confirmation test at least once per day or as often as required to establish confidence in consistent compliance with 10.4.1.

11. Procedure

11.1 Sample Preparation:

11.1.1 The amount of sample injected must not overload the column stationary phase nor exceed the detector linear range. A narrow boiling range sample will require a smaller amount injected than a wider boiling range sample.

11.1.1.1 To determine the detector linear range, refer to Practice E594 for flame ionization detectors or Practice E516 for thermal conductivity detectors.

11.1.1.2 The column stationary phase capacity can be estimated from the chromatogram of the calibration mixture (see 9.3.2). Different volumes of the calibration standard can be injected to find the maximum amount of a component that the stationary phase can tolerate without overloading (see 10.3.1).

Note the peak height for this amount of sample. The maximum sample signal intensity should not exceed this peak height.

11.1.2 Samples that are of low enough viscosity to be sampled with a syringe at ambient temperature may be injected neat. This type of sample may also be diluted with CS₂ to control the amount of sample injected to comply with 11.1.1.

11.1.3 Samples that are too viscous or waxy to sample with a syringe may be diluted with CS₂.

11.1.4 Typical sample injection volumes are listed below.

Packed Columns:	
Stationary Phase Loading, %	Neat Sample Volume, μ L
10	1.0
5	0.5
Open Tubular Columns:	
Film Thickness, μ	Neat Sample Volume, μ L
0.8 to 1.5	0.1 to 0.2
1.8 to 3.0	0.1 to 0.5
3.0 to 5.0	0.2 to 1.0

11.2 Sample Analysis—Using the analysis sequence protocol, inject a sample aliquot into the gas chromatograph. Collect a contiguous time slice area record of the entire analysis.

12. Calculations⁵

12.1 Acquisition Rate Requirements:

12.1.1 The number of slices required at the beginning of data acquisition depends on chromatographic variables such as the column flow, column film thickness, and initial column temperature as well as column length. In addition the detector signal level has to be as low as possible at the initial temperature of the analysis. The detector signal level for both the sample signal and the blank at the beginning of the run has to be similar for proper zeroing of the signals.

12.1.2 The sampling frequency has to be adjusted so that at least a significant number of slices are acquired prior to the start of elution of sample or solvent. For example, if the time for start of sample elution is 0.06 min (3.6 s), a sampling rate of 5 Hz would acquire 18 slices. However a rate of 1 Hz would only acquire 3.6 slices which would not be sufficient for zeroing the signals. Rather than specifying number of slices, it is important to select an initial time segment that is, one or two seconds. Insure that the smallest number of slices is 5 or greater.

12.1.3 Verify that the slice width used to acquire the sample chromatogram is the same used to acquire the blank run chromatogram.

12.2 *Chromatograms Offset for Sample and Blank*—Perform a slice offset for the sample chromatogram and blank chromatogram. This operation is necessary so that the signal is corrected from its displacement from the origin. This is achieved by determining an average slice offset from the slices accumulated in the first segment (that is, first s) and performing a standard deviation calculation for the first N slices accumulated. It is carried out for both sample signal and baseline signal.

12.2.1 Sample Offset:

12.2.1.1 Calculate the average slice offset of sample chromatogram using the first second of acquired slices. Insure that no sample has eluted during this time and that the number of slices acquired is at least 5. Throw out any of the first N slices selected that are not within one standard deviation of the average and recompute the average. This eliminates any area that is due to possible baseline upset from injection.

12.2.1.2 Subtract the average slice offset from all the slices of the sample chromatogram. Set negative slices to zero. This will zero the chromatogram.

12.2.2 Blank Offset:

NOTE 7—If you are using electronic baseline compensation proceed to 12.4. It is strongly recommended that the offset method use the slices acquired by running a blank with or without solvent according to how the sample was prepared. Use these acquired blank slices for the offset or zero calculations.

12.2.2.1 Repeat 12.2.1 using the blank run table.

12.3 *Offset the Sample Chromatogram with Blank Chromatogram*—Subtract from each slice in the sample chromatogram table with its correspondent slice in the blank run chromatogram table. Set negative slices to zero.

12.4 Determine the Start of Sample Elution Time:

12.4.1 *Calculate the Total Area*—Add all the corrected slices in the table. If the sample to be analyzed has a solvent peak, start counting area from the point at which the solvent peak has eluted completely. Otherwise, start at the first corrected slice.

12.4.2 *Calculate the Rate of Change between each Two Consecutive Area Slices*—Begin at the slice set in 12.4.1 and work forward. The rate of change is obtained by subtracting the area of a slice from the area of the immediately preceding slice and dividing by the slice width. The time where the rate of change first exceeds 0.0001 % per second of the total area (see 12.4.1) is defined as the start of the sample elution time. To reduce the possibility of noise or an electronic spike falsely indicating the start of sample elution time, a 1 s slice average can be used instead of a single slice. For noisier baselines, a slice average larger than 1 s may be required.

12.5 Determine the End of Sample Elution Time:

12.5.1 *Calculate the Rate of Change between each Two Consecutive Area Slices*—Begin at the end of run and work backward. The rate of change is obtained by subtracting the area of a slice from the area of the immediately preceding slice and dividing by the slice width. The time where the rate of change first exceeds 0.0001 % per second of the total area (see 12.4.1) is defined as the end of sample elution time. To reduce the possibility of noise or an electronic spike falsely indicating the end of sample elution a 1 s slice average can be used instead of a single slice. For noisier baselines a slice average larger than 1 s may be required.

12.6 *Calculate the Sample Total Area*—Add all the slices from the slice corresponding to the start of sample elution time to the slice corresponding to the end of sample elution time.

12.7 *Normalize to Area Percent*—Divide each slice in the sample chromatogram table by the total area (see 12.6) and multiply it by 100.

12.8 Calculate the Boiling Point Distribution Table:

12.8.1 *Initial Boiling Point*—Add slices in the sample chromatogram until the sum is equal to or greater than 0.5 %. If the sum is greater than 0.5 %, interpolate (refer to the algorithm in 12.9.1) to determine the time that will generate the exact 0.5 % of the area. Calculate the boiling point temperature corresponding to this slice time using the calibration table. Use interpolation when required (refer to the algorithm in 12.9.2).

12.8.2 *Final Boiling Point*—Add slices in the sample chromatogram until the sum is equal to or greater than 99.5 %. If the sum is greater than 99.5 %, interpolate (refer to the algorithm in 12.9.1) to determine the time that will generate the exact 99.5 % of the area. Calculate the boiling point temperature corresponding to this slice time using the calibration table. Use interpolation when required (refer to the algorithm in 12.9.2).

12.8.3 *Intermediate Boiling Point*—For each point between 1 % and 99 %, find the time where the accumulative sum is equal to or greater than the area percent being analyzed. As in 12.8.1 and 12.8.2, use interpolation when the accumulated sum exceeds the area percent to be estimated (refer to the algorithm in 12.9.1). Use the calibration table to assign the boiling point.

⁵ Supporting data have been filed at ASTM International Headquarters and may be obtained by requesting Research Report RR:D02-1477.

12.9 Calculations Algorithms:

12.9.1 Calculations to determine the exact point in time that will generate the X percent of total area, where X = 0.5, 1, 2, ..., 99.5 %.

12.9.1.1 Record the time of the slice just prior to the slice that will generate an accumulative slice area larger than the X percent of the total area. Let us call this time, T_s , and the accumulative area at this point, A_c .

12.9.1.2 Calculate the fraction of the slice required to produce the exact X percent of the total area:

$$A_x = \frac{X - A_c}{A_{c+1} - A_c} \quad (3)$$

where:

A_x = fraction of the slice that will yield the exact percent,
 A_c = cumulative percent up to the slice prior to X,
 A_{c+1} = cumulative percent up to the slice right after X, and
 X = desired cumulative percent.

12.9.1.3 Calculate the time required to generate the fraction of area A_x :

$$T_f = A_x \cdot W \quad (4)$$

where:

W = slice width,
 A_x = fraction of the slice that will yield the exact percent, and
 T_f = fraction of time that will yield A_x .

12.9.1.4 Record the exact time where the accumulative area is equal to the X percent of the total area:

$$T_i = T_s + T_f \quad (5)$$

where:

T_s = fraction of the slice that yields the cumulative percent up to the slice prior to X,
 T_f = fraction of time that will yield A_x , and
 T_i = time where the cumulative area is equal to X percent of the total area.

12.9.2 Interpolate to determine the exact boiling point given the retention time corresponding to the cumulative slice area.

12.9.2.1 Compare the given time against each retention time in the calibration table. Select the nearest standard having a retention time equal to or larger than the interpolation time. (**Warning**—The retention time table shall be sorted in ascending order.)

12.9.2.2 If the interpolation time is equal to the retention time of the standard, record the corresponding boiling point.

12.9.2.3 If the retention time is not equal to the retention time of the standards (see 9.3), interpolate the boiling point temperature as follows:

12.9.2.4 If the interpolation time is less than the first retention time in the calibration table, then extrapolate using the first two components in the table:

$$BP_x = m_1 \cdot (RT_x - RT_1) + BP_1 \quad (6)$$

where:

m_1 = $(BP_2 - BP_1) / (RT_2 - RT_1)$,
 BP_x = boiling point extrapolated,
 RT_x = retention time to be extrapolated,
 RT_1 = retention time of the first component in the calibration table,
 BP_1 = boiling point of the first component in the calibration table,
 RT_2 = retention time of the second component in the calibration table, and
 BP_2 = boiling point of the second component in the calibration table.

12.9.2.5 If the interpolation time is between two retention times in the calibration table, then interpolate using the upper and lower standard components:

$$BP_x = m_u \cdot (RT_x - RT_l) + BP_l \quad (7)$$

where:

m_u = $(BP_u - BP_l) / (RT_u - RT_l)$,
 BP_x = boiling point extrapolated,
 RT_x = retention time to be extrapolated,
 RT_l = retention time of the lower bound component in the calibration table,
 BP_l = boiling point of the lower bound component in the calibration table,
 RT_u = retention time of the upper bound component in the calibration table, and
 BP_u = boiling point of the upper bound component in the calibration table.

12.9.2.6 If the interpolation time is larger than the last retention time in the calibration table, then extrapolate using the last two standard components in the table:

$$BP_x = m_n \cdot (RT_x - RT_{n-1}) + BP_{n-1} \quad (8)$$

where:

m_n = $(BP_n - BP_{n-1}) / (RT_n - RT_{n-1})$,
 BP_x = boiling point extrapolated,
 RT_x = retention time to be extrapolated,
 RT_{n-1} = retention time of the standard component eluting prior to the last component in the calibration table,
 BP_{n-1} = boiling point of the standard component eluting prior to the last component in the calibration table,
 RT_n = retention time of the last component in the calibration table, and
 BP_n = boiling point of the standard component in the calibration table.

13. Report

13.1 Report the temperature to the nearest 0.5°C (1°F) at 1 % intervals between 1 and 99 % and at the IBP (0.5 %) and the FBP (99.5 %). Other report formats based upon users' needs may be employed.

NOTE 8—If a plot of the boiling point distribution curve is desired, use a spreadsheet with uniform subdivisions and use either retention time or temperature as the horizontal axis. The vertical axis will represent the boiling range distribution (0 to 100 %). Plot each boiling temperature against its corresponding normalized percent. Draw a smooth curve connecting the points.

TABLE 5 Repeatability

NOTE 1— x = the average of the two results in °C and y = the average of the two results in °F.

% Off	Repeatability	
	°C	°F
IBP	0.011 x	0.011 ($y - 32$)
5 %	0.0032 ($x + 100$)	0.0032 ($y + 148$)
10–20 %	0.8	1.4
30 %	0.8	1.4
40 %	0.8	1.4
50–90 %	1.0	1.8
95 %	1.2	2.2
FBP	3.2	5.8

TABLE 6 Reproducibility

NOTE 1— x = the average of the two results in °C and y = the average of the two results in °F.

% Off	Reproducibility	
	°C	°F
IBP	0.066 x	0.066 ($y - 32$)
5 %	0.015 ($x + 100$)	0.015 ($y + 148$)
10–20 %	0.015 ($x + 100$)	0.015 ($y + 148$)
30 %	0.013 ($x + 100$)	0.013 ($y + 148$)
40 %	4.3	7.7
50–90 %	4.3	7.7
95 %	5.0	9.0
FBP	11.8	21.2

14. Precision and Bias⁶

14.1 *Precision*—The precision of this test method as determined by the statistical examination of the interlaboratory test results is as follows:

14.1.1 *Repeatability*—The difference between successive test results obtained by the same operator with the same apparatus under constant operating conditions on identical test material would, in the long run, in the normal and correct operation of the test method, exceed the following values by only one case in twenty (see Table 5).

14.1.2 *Reproducibility*—The difference between two single and independent results obtained by different operators working in different laboratories on identical test material would, in the long run, exceed the following values only one case in twenty (see Table 6).

NOTE 9—This precision estimate is based on the analysis of nine samples by 19 laboratories using both packed and open tubular columns. The range of results found in the round robin are listed in Table 7.

14.2 *Bias*—The procedure in Test Method D2887 for determining the boiling range distribution of petroleum fractions by gas chromatography has no bias because the boiling range distribution can only be defined in terms of a test method.

14.2.1 A rigorous, theoretical definition of the boiling range distribution of petroleum fractions is not possible due to the complexity of the mixture as well as the unquantifiable interactions among the components (for example, azeotropic behavior). Any other means used to define the distribution would require the use of a physical process, such as a conventional distillation or gas chromatographic characterization. This would therefore result in a method-dependent definition and would not constitute a true value from which bias can be calculated.

Procedure B, Accelerated Method

15. Introduction

15.1 Procedure B was developed for carrying out Test Method D2887 in an accelerated mode. By changing variables such as carrier flow, oven heating and type of column, it is possible to significantly reduce the analysis time. The term accelerated is used here to distinguish this technique from

TABLE 7 Round Robin Range of Results

% Off	Range of Results, °C	Range of Results, °F
IBP	112–213	234–415
5 %	133–286	271–547
10 %	139–312	282–594
20 %	151–341	304–646
30 %	161–358	322–676
40 %	171–370	340–698
50 %	182–381	360–718
60 %	196–390	385–734
70 %	206–401	403–754
80 %	219–412	426–774
90 %	233–426	451–799
95 %	241–437	466–819
FBP	274–475	525–887

ultrafast chromatography, which requires direct heating of the column only. In addition, the precision study involved mixtures of ultra low sulfur diesel and B100. The need to use solvent for sample dilution is not required.

15.2 Procedure B requires the use of a Flame Ionization detector only. Sections common to both procedures are referenced in Procedure B.

16. Apparatus

16.1 *Chromatograph*—The gas chromatograph used shall have the following performance characteristics:

16.1.1 *Detector*—A flame ionization detector (FID) must be used. The detector must have a Minimum Detectable Quantity of 2.0 pg carbon/s for n-C13 or better. The detector requires a sensitivity of 0.005C/g-0.010C/g of Carbon. Operating at this sensitivity level, detector stability must be such that a baseline drift of not more than 10^{-12} to 10^{-13} A/h(Pico Amps/Hour). This drift is measured as change in detector current per unit time. The detector must be capable of operating continuously at a temperature equivalent to the maximum column temperature employed (see Table 8). Connection of the column to the detector must be such that no temperature below the column temperature exists. It is recommended that the Flame Jet have an orifice of or (0.5 ± 0.08 mm) in order to avoid premature decrease of the flame tip orifice due to accumulation of column bleed substrate.

16.1.2 *Programmable Oven*—The gas chromatograph must be capable of achieving linear programmed temperature operation at rates of 35°C/min over the entire range of the conditions in Table 8.

⁶ Supporting data have been filed at ASTM International Headquarters and may be obtained by requesting Research Report RR:D02-1406.

TABLE 8 Typical Operating Parameters for Procedure B (Accelerated D2887) Test

	Column 1	Column 2	Column 3
Column length (m)	10	5	7.5
Column ID (mm)	0.53	0.53	0.53
Stationary phase thickness (μm) ^A	0.88	2.65	1.5
Carrier gas	helium	helium	helium
Carrier gas flow rate (mL/min)	26	35	37
Initial column temperature (°C)	60	40	40 (0.5 min)
Final column temperature (°C)	360	350	360
Oven programming rate (°C/min)	35	35	35
Detector	FID	FID	FID
Detector temperature (°C)	360	360	365
Injector	PTV	PTV	Cool on column
Injector initial temperature (°C)	100	100	100 (0.5 min)
Injector programming rate (°C/min)	35	35	35
Injector final temperature (°C)	360	350	350
Sample size (μL)	0.1	0.1	0.1
Dilution concentration	neat	neat	neat
Analysis time (minutes)	8	7.8	8

^A All columns contain a polydimethylsiloxane stationary phase.

NOTE 10—Some instrument manufacturers may require different line voltages in order to rapidly heat the oven.

16.1.3 Sample Inlet System—Temperature programmable inlets or Cool on Column inlets should be used preferentially for this method. Temperature programmable inlet (PTV) is an inlet that transfers the sample directly to the column without venting a portion of the sample and usually contains a liner. Cool on column inlets contain no liners. Isothermally operated inlets are not recommended for this test method.

16.1.4 Inlet Septa—It is important that septa be chosen that provide maximum stability at the inlet highest operational temperature. The septa should be periodically replaced after 50 runs. Septa particles in the inlet are responsible for ghost peaks in the blank signal.

16.1.5 Electronic Pneumatic Control—The gas chromatograph must be equipped with electronic flow controllers capable of maintaining carrier gas flow constant to $\pm 1\%$ or better over the full operating temperature range of the column. The flow control should be carried out by flow sensors rather than a calculated pressure program to maintain constant flow. The carrier gas supply pressure must have at least a differential of 135 kPa (20 psi) between the column pressure at 350°C and the gas supply pressure.

16.1.6 Automatic Sample Injectors—The use of autosamplers equipped with a micro syringe capable of delivering 0.1 μL is required for reproducible retention time.

16.2 Column—Use one of the three columns listed in [Table 8](#). These columns contain Polydimethyl-Siloxane (PDMS) as

the liquid phase. These columns elute n-paraffins hydrocarbons according to boiling point.

16.3 Data Acquisition System:

16.3.1 Computer—A computer with data acquisition software is necessary to control the instrument, perform the injections, syringe washes, sample aspiration, sample injections, and signal digitization and acquisition. The data acquisition software is operated in the peak processing mode and or in the slice mode.

16.3.2 Simulated distillation calculations are carried out with software conforming to the algorithms specified in [Section 12](#).

17. Reagents and Materials

17.1 Calibration Mixture—An accurately weighed mixture of approximately equal mass quantities of n-hydrocarbons dissolved in CS₂. The total concentration of the hydrocarbons must be approximately 1 mass % (Warning—CS₂ is extremely volatile, flammable, and toxic.) The mixture shall cover the boiling range from n-C5 to n-C44, but it is not necessary to include every carbon number (see [Note 5](#), Procedure A, [7.8.1.1](#)).

17.1.1 The calibration mixture contains the normal paraffins with carbons numbers 5, 6, 7, 8, 9, 10, 11, 12, 14, 15, 16, 17, 18, 20, 24, 28, 32, 36, 40, and 44.

17.1.2 If samples contain hydrocarbons eluting prior to the elution of C5, it is necessary for the calibration mixture to contain paraffin with a boiling point below the IBP of the sample. Propane or butane can be added to the calibration mixture, non-quantitatively, by bubbling the gaseous compound into the calibration mixture contained in a septum sealed vial by using a gas syringe.

17.1.3 The calibration mixture has a limited concentration of the paraffins to a total of 1 %. This is necessary to maintain the skewness of the chromatographic peaks. CS₂ is usually used as a solvent. Cyclohexane has also been used as a solvent. These calibration mixtures are available from many chromatographic supply companies.

17.2 The gases used for the operation of the gas chromatograph are described in Procedure A [7.4-7.6](#).

17.2.1 Air cooling is necessary for inlets that use temperature programming. The air is provided by a separate line from that used in operating the FID detector. The purity requirement for this air source is oil and moisture free.

17.3 Reference Gas Oil #1-Batch 2—Used to check the overall system. This material is obtained from Chromatographic Suppliers. Users may also use Batch 1 if available.

17.4 Hydrocarbon Filters and Oxygen Traps—These are required to obtain good base signals and protect the column. It is desirable that the oxygen trap be provided with a visible indicator to determine the presence of oxygen in the system. Spent oxygen traps must be replaced.

17.5 CS₂—may be used to rinse the autosampler syringe between injections. (**Warning**—CS₂ is a toxic chemical. It is extremely flammable.)




FIG. 3 Determination of Column Resolution

18. Preparation of Apparatus

18.1 Install the capillary column according to the manufacturer of the Inlet used. Install the column at the detector end also.

18.2 Condition the column at intervals of 50°C both for inlet and GC oven. The signal will increase and then decrease. When the signal has decreased to half its value, increase both inlet and oven to the next 50°C interval. Repeat this process until you reach the final temperature of the column used (see Table 8). If during any of these intervals the baseline does not decrease to a lower value, stop the process immediately and return the oven to 40°C. Check for leaks in the system. Alternatively you can program the oven from 40°C to the final column temperature at rate of 5°C/min. Hold this temperature for 4 h until the baseline signal no longer drops. If the latter technique is chosen the system has to be leak-tight.

18.3 After the column has been conditioned, inject the retention time calibration sample containing the paraffins. Use the conditions of Table 8. Determine the column resolution and skewness as shown in Figs. 3 and 4. The column resolution was determined in the Precision Study to be between 4-11. The peak skewness of all peaks was determined in the interlaboratory study to be between 0.8-1.30 as shown in Table 9. A typical example of peak skewness calculation is shown in Fig. 4. Typical plot of the retention time vs. boiling point is shown in Fig. 6 which is obtained using any of the columns and conditions in Table 8.

18.3.1 When the instrument is commissioned verify that the relative response factors of paraffins in the calibration sample is unity as described in 9.3.2. One of the test that are performed when the Reference Gas Oil values do not meet the requirement is to obtain the relative response factor. The relative response factor determination can be obtained simultaneously with the retention time calibration if the concentrations of the hydrocarbons are known and if a blank is used to obtain the net area of the hydrocarbons. An example is shown in Table 10.

18.4 A blank is required for the analysis of the Reference Gas Oil and for the samples. Since no diluent is used in this method, a blank is a chromatogram with no injections. Acceptable blanks do not show appreciable ghost peaks and the signal at the end of the run is not higher in magnitude than the equivalent section of the sample signal. If the blank at the isothermal section displays a positive slope, correct the system for leaks. Column compensation can also be used, although it is not the preferred technique (see 12.2.2, Note 7).

19. Calibration

19.1 Inject the retention time calibration standard (see Fig. 5). Verify that the plot of the Boiling Point of the paraffins versus retention time has a shape as that shown in Fig. 6. The boiling points of the paraffins are given in Table 2. The calibration has a nonlinear portion for the first four paraffins of the calibration mixture. The second portion of the curve is essentially a straight line extending from C9 to C44. The chromatographic conditions to obtain the calibration curve are detailed in Table 8.

19.2 The retention time and peak area for the calibration curve data are acquired in the peak processing mode. Verify that the calibration standard yields the column resolution, peak skewness values for all components in the mixture as described in 18.3.

19.3 Inject the Reference Gas Oil with the conditions established in 18.3. Subsequently run a blank run without sample injection or use the blank in 18.4. Overlay the Reference Gas Oil chromatogram with the blank. Verify that the blank does not cross the Reference Gas Oil chromatogram especially at the end of the run. In addition verify that the blanks do not have ghost peaks. If either or both of these this occurs, inspect and or replace septum, glass liner and inspect the oxygen filter to ensure the absence of oxygen in to the chromatographic system. Further column conditioning may also be required. Typical chromatograms for the Reference Gas




FIG. 4 Determination of Peak Skewness

TABLE 9 Typical Skewness obtained from the Calibration Chromatogram (Procedure B)

Component	Skewness
n-C5	— ^A
n-C6	1.00
n-C7	1.00
n-C8	1.00
n-C9	1.04
n-C10	0.97
n-C11	1.00
n-C12	1.00
n-C13	1.00
n-C14	0.97
n-C15	1.03
n-C16	0.97
n-C17	1.03
n-C18	1.05
n-C19	1.08
n-C20	1.00
n-C24	0.94
n-C28	1.02
n-C32	1.02
n-C36	1.19
n-C40	0.89
n-C44	0.83

^A Skewness of n-C5 is excluded (see 18.3).

Oil Batch 2 are shown in Figs. 7 and 8 obtained with columns 1 to 3 described in Table 8.

19.4 If Biodiesel Mixtures (B5, B10 and B20) are analyzed, the chromatogram will show a prominent peak at the end of the analysis as shown in Fig. 9. The large peak is due to the elution of the major Fatty Acid Methyl Ester.

19.5 Obtain the Boiling Point Distribution (BP) for the Reference Gas Oil according to the calculations set forth in Section 12 Procedure A. Report the Boiling Point Distribution. Compare the values with those in Table 11. If the values do not agree with the Reproducibility reported in Table 11 refer to the possible problems reported in 19.3.

19.6 It is permissible to use a reference selected material that has similar boiling point characteristics as the samples analyzed. However, the principal reference is the ASTM Reference Gas Oil and the validity of a secondary standard as a reference is determined after compliance with the Reference Gas Oil analysis.

20. Sample Analysis

20.1 Sample Preparation:

20.1.1 Fill the autosampler vials to a volume which leaves a small headspace. For very viscous samples there are two options: 1) Adjust the autosampler withdrawal speed used to fill the syringe to a slow speed so as to not to create a vacuum which results in non-repeatable volumes or 2) Add 2 drops of CS₂ to the autosampler vial.

20.1.2 For the injection of the Reference Gas Oil, use if possible autosampler vials that have a reduce volume or use vial inserts that reduce the volume. The Reference Gas Oil is a valuable reference material where the supply available is limited and it should be conserved.

20.1.3 Adjust the autosampler to rinse the syringe with an adequate number of sample rinses if no CS₂ is used in the wash vials of the autosampler.




FIG. 5 Typical Retention Time Calibration (Columns 1-3, Table 8)




FIG. 6 Retention Times vs. Temperature Calibration obtained under conditions of Table 8

20.1.4 If the gas chromatograph has been sitting idle for more than 2 hours it is advisable to run a blank that will clear the accumulated trace amounts of septa, carrier impurities and other retain substances at the lower standby temperature of the inlet as well as those accumulated at the entrance of the column. This run is not used as a blank to verify instrument calibration.

20.2 Sequence Preparation:

20.2.1 If this is the first time that an analysis is carried out prepare the sequence to include the retention time calibration standard, the Reference Gas Oil and a blank which is necessary to calculate the Boiling Point Distribution of the Reference Gas Oil as well as for subsequent samples analysis. Calibration should be performed weekly when the instrument is in use, or

TABLE 10 Determination of Relative Response Factor^A

% Mass	Cn	Response Factor	% Deviation
0.186	C5	1.008	1
0.089	C6	1.003	0.3
0.0589	C7	1.087	8.7
0.0648	C8	1.049	4.9
0.0547	C9	1.016	1.6
0.0557	C10	1.000	0.0
0.0562	C11	0.997	-0.3
0.0937	C12	0.983	-1.7
0.0574	C13	0.984	-1.6
0.0579	C14	0.986	-1.4
0.0416	C15	0.978	-2.2
0.0352	C16	0.980	-2.0
0.0236	C17	0.982	-1.8
0.0247	C18	0.979	-2.1
0.0257	C19	0.979	-2.1
0.0393	C20	0.974	-2.6
0.0126	C24	0.983	-1.7
0.0129	C28	0.981	-1.9
0.013	C32	0.974	-2.6
0.0138	C36	1.006	0.6
0.0126	C40	1.050	5.0
0.0129	C44	1.021	2.1

^A Calculated by the use of Equation 2.

whenever maintenance is performed and as dictated by the lab on-site precision and or Quality Control protocol.

20.2.2 Adjust the autosampler to inject the amount stipulated in Table 8. Do not exceed the volume listed in the Table 8

20.2.3 Do not leave the retention time calibration vial exposed to ambient temperature since the evaporation will change the relative concentrations of the n-paraffins in the mixture which will potentially lead to a failure in the peak skewness and misidentification of the n-paraffins.

20.2.4 Periodically insert blanks in the sample analysis sequence. A determination of the state of the baseline can be made by examining the sample and baseline overlay in the data system.

20.3 Data Analysis:

20.3.1 *Reference Gas Oil*—After running the sequence described in 20.2.1, examine the linearity of the retention time calibration and examine the Boiling Point Distribution of the Reference Gas Oil (RGO). If the values are within the reproducibility values obtained from the ILS study than proceed to analyze samples. Periodically analyze the Reference Gas Oil in order to verify the system accuracy. The RGO should be conserved since the supply is limited. Alternatively, once the system has passed the test of the RGO a secondary standard can be used.

20.3.2 If the Reference Gas Oil (RGO) Boiling Point Distribution is not within the values of Table 11, examine all the recommendations in Sections 18 and 19. Subsequently a recalibration can be carried out. Further failure of the RGO will indicate an examination of the instrument components aided by recommendations of the instrument manufacturer.

21. Calculations

21.1 The calculations and algorithms for obtaining the Boiling Point Distributions are described in Section 12 Procedure A. Additional calculation (D86) is found in Appendix X4.

22. Report

22.1 Report the Boiling Point Distribution to the nearest 0.5°C (1.0°F) at 1 % intervals from the Initial Boiling Point (IBP) at 0.5 % up to the Final Boiling Point at 99.5 % (FBP). Select the % intervals as required by the sample nature.

23. Precision and Bias⁷

23.1 Precision was determined by an Interlaboratory Study (ILS 158). The study consisted in 10 samples (10 laboratories) which included 3 mixtures containing 5 %, 10 % and 20 % of FAME ester in Ultra Low Sulfur Diesel (B5, B10, B20). The Reference Gas Oil was included as an unknown in the samples. From this data the accepted values for the Reference Gas Oil were determined as shown in Table 11. It is to be noted that the values are almost numerically equal to the values listed in Table 3 of Procedure A.

23.2 *Precision*—The precision of this test method as determined by RR:D02-1761 is as follows. The precision values are to be used only in the Temperature Ranges of Table 12.

23.2.1 *Repeatability*—The difference between successive test results obtained by the same operator with the same apparatus under constant operating conditions on identical test material would, in the long run, in the normal and correct operation of the test method, exceed the following values by only one case in twenty. Values obtained from the precision study are shown in Table 13.

23.2.2 *Reproducibility*—The difference between two single and independent results obtained by different operators working in different laboratories on identical test material would, in the long run, exceed the following values only one case in twenty. Results of the precision study for Reproducibility are shown in Table 13.

23.2.3 Example calculations on the precision are shown in Table 14.

23.3 *Bias*—A study to comply with ASTM D6708 will be completed within 5 years in order to compare the results of Procedure A and Procedure B. In addition a statistical analysis was carried out by comparing the Boiling Point Distributions of the Reference Gas by using Procedure A and Procedure B, respectively. The results obtained with Procedure B are shown in Table 11. The results in Table 11 are well within the acceptable differences as those listed in Table 3.

23.3.1 The bias statement in 14.2.1 applies also to Procedure B.

24. Keywords

24.1 boiling range distribution; correlation; distillation; gas chromatography; petroleum; petroleum fractions; petroleum products; simulated distillation

⁷ Supporting data have been filed at ASTM International Headquarters and may be obtained by requesting Research Report RR:D02-D1761. Contact ASTM Customer Service at service@astm.org.




FIG. 7 Analysis of the Reference Gas Oil Under Accelerated Conditions (Column 2)




FIG. 8 Analysis of the Reference Gas Oil under Accelerated Conditions (Column 1)




FIG. 9 Chromatogram of a Biodiesel (B-10) mixture

TABLE 11 Precision Values Obtained for the ASTM Reference Gas Oil No. 1 Batch 2^A (Procedure B)

<i>M</i> , %	<i>T</i> _i , °C	<i>T</i> _i , °F	<i>r</i> _i , °C	<i>r</i> _i , °F	<i>R</i> _i , °C	<i>R</i> _i , °F
IBP	113.3	235.9	2.94	5.29	7.97	14.35
0.05	150.0	302.0	0.56	1.00	2.92	5.25
0.10	174.6	346.3	0.58	1.04	3.03	5.45
0.20	223.9	435.0	0.62	1.12	3.25	5.85
0.30	259.7	499.5	0.65	1.17	3.41	6.14
0.40	289.4	552.9	0.68	1.22	3.87	6.96
0.50	312.4	594.3	0.70	1.25	3.65	6.57
0.60	331.8	629.2	0.71	1.28	3.73	6.72
0.70	354.1	669.4	0.73	1.32	3.83	6.90
0.80	378.5	713.3	0.75	1.36	3.94	7.10
0.90	407.7	765.9	0.78	1.40	4.08	7.34
0.95	429.8	805.6	0.80	1.43	4.17	7.51
FBP	480.8	897.4	3.30	5.94	7.63	13.73

^A Values obtained from the ILS study (Research Report RR:D02-1760).

TABLE 13 Repeatability and Reproducibility, Procedure B (Accelerated D2887) Test^{A,B}

% Mass	Repeatability, <i>r</i> (°C)	Repeatability, <i>r</i> (°F)	Reproducibility, <i>R</i> (°C)	Reproducibility, <i>R</i> (°F)
IBP	2.94	5.29	7.97	9.52
5– 95 %	0.000857 (X+ 500)	0.000857(X+ 868)	0.00449(X+500)	0.00449(X+868)
FBP	3.32	6	7.63	10.8

^A Several Mass % did not meet the required DF (Degrees of Freedom) > 30 as stated by ASTM D6300-06, section 6.4.3 Note 1. Thus, the following sentence is added “Further Standardization is Recommended.”

^B The precision values are to be used only in the Temperature Ranges of Table 12.

TABLE 12 Temperature Ranges Covered in the ILS Study

% Off	Range, °C	Range, °F
IBP	110-131	230-268
5	138-201	280-394
10	144-282	291-540
20	159-322	318-612
30	170-340	338-644
40	184-350	363-662
50	196-360	385-680
60	208-370	406-698
70	221-384	430-723
80	236-396	457-745
90	259-423	498-793
95	268-439	514-822
FBP	288-534	550-993

TABLE 14 Calculation of Repeatability and Reproducibility at Selected Temperatures^A

% Mass	High, °C	High, °F	Low, °C	Low, °F	Median, °C	Median, °F	<i>r</i> , °C	<i>R</i> , °C	<i>r</i> , °F	<i>R</i> , °F
IBP	127	260.6	111	231.8	122	251.6	2.94	7.97	5.3	14.3
T5	193	379.4	139	282.2	174	345.2	0.58	3.03	1.0	5.5
T10	214	417.2	145	293	201	393.8	0.6	3.15	1.1	5.7
T20	239	462.2	160	320	221	429.8	0.62	3.24	1.1	5.8
T30	260	500	172	341.6	243	469.4	0.64	3.34	1.2	6.0
T40	289	552.2	184	363.2	254	489.2	0.65	3.39	1.2	6.1
T50	312	593.6	197	386.6	271	519.8	0.66	3.46	1.2	6.2
T60	369	696.2	209	408.2	290	554	0.68	3.55	1.2	6.4
T70	382	719.6	223	433.4	308	586.4	0.69	3.63	1.2	6.5
T80	397	746.6	236	456.8	332	629.6	0.71	3.74	1.3	6.7
T90	419	786.2	354	669.2	354	669.2	0.73	3.83	1.3	6.9
T95	438	820.4	266	510.8	364	687.2	0.74	3.88	1.3	7.0
FBP	497	926.6	289	552.2	406	762.8	2.32	7.63	4.2	13.7

^A The selected values were obtained from the precision study.

APPENDIXES

(Nonmandatory Information)

X1. BOILING POINTS OF NONPARAFFINIC HYDROCARBONS (PROCEDURES A AND B)

X1.1 There is an apparent discrepancy in the boiling point of multiple ring-type compounds. When the retention times of these compounds are compared with *n*-paraffins of equivalent atmospheric boiling point, these ring compounds appear to be eluted early from methyl silicone rubber columns. A plot showing 36 compounds other than *n*-paraffins plotted along the calibration curve for *n*-paraffins alone is shown in Fig. X1.1. The numbered dots are identified in Table X1.1. In this figure, the atmospheric boiling points are plotted against the observed retention times. If columns containing different percentages of stationary phase or different temperature programming rates were used, the slope and curvature of the *n*-paraffin curve (solid line) would change, but the relative relationships would remain essentially the same. Deviations of simulated distillation boiling points, as estimated from the curve, from actual boiling points for a few compounds are shown in Table X1.2. The deviations obtained by plotting boiling points at 10 mm rather than 760 mm are tabulated also. It is apparent that the deviation is much less at 10 mm pressure. This indicates that the distillation data produced by gas chromatography closely approximates those obtained in reduced pressure distillation.

Since the vapor-pressure-temperature curves for multiple-ring type compounds do not have the same slope or curvature as those of *n*-paraffins, an apparent discrepancy would exist when *n*-paraffin boiling points at atmospheric pressure are used.

X1.2 However, this discrepancy does not introduce any significant error when comparing with laboratory distillation because the pressure must be reduced in such procedures when overhead temperatures reach approximately 260°C (500°F) to prevent cracking of the sample. Thus, distillation data are subject to the same deviations experienced in simulated distillation by gas chromatography. A comparison of data obtained from TBP distillation with those obtained from simulated distillation of three high boiling petroleum fractions is shown in Table X1.3. The TBP distillations were made on 100 theoretical plate spinning band columns at 1 mm Hg pressure.

X1.3 The decanted oil is of particular interest because it contains a high presence of polycyclic aromatic compounds and the high sulfur coker gas oil should contain ring-type sulfur compounds and complex olefinic types.




FIG. X1.1 Boiling Point—Retention Time Relationships for Several High-Boiling Multiple-Ring Type Compounds (see Table X1.1)

TABLE X1.1 Compound Identification—Numbered Dots (see Fig. X1.1)

No.	Boiling Point, °C (°F)	Compound	Number	Boiling Point, °C (°F)	Compound
2	80 (176)	benzene	27	227 (441)	di- <i>n</i> -amylsulfide
3	84 (183)	thiophene	28	234 (453)	tri-isopropylbenzene
5	111 (231)	toluene	30	241 (466)	2-methylnaphthalene
6	116 (240)	pyridine	31	295 (473)	1-methylnaphthalene
8	136 (277)	2,5-dimethylthiophene			
9	139 (282)	<i>p</i> -xylene	34	254 (489)	indole
10	143 (289)	di- <i>n</i> -propylsulfide	35	279 (534)	acenaphthene
12	152 (306)	cumene			
13	159 (319)	1-hexahydroindan	38	298 (568)	<i>n</i> -decylbenzene
14	171 (339)	1-decene	39	314 (598)	1-octadecene
15	173 (344)	<i>sec</i> -butylbenzene			
17	178 (352)	2,3-dihydroindene	41	339 (642)	phenanthrene
18	183 (361)	<i>n</i> -butylbenzene	42	342 (647)	anthracene
19	186 (366)	<i>trans</i> -decalin			
20	194 (382)	<i>cis</i> -decalin	44	346 (655)	acridine
21	195 (383)	di- <i>n</i> -propyldisulfide	45	395 (743)	pyrene
23	213 (416)	1-dodecene	47	404 (796)	triphenylene
25	218 (424)	naphthalene	49	438 (820)	naphthacene
26	221 (430)	2,3-benzothiophene	50	447 (837)	chrysene

TABLE X1.2 Deviations of Simulated Distillation Boiling Points From Actual Boiling Points

Compound	Boiling Point, °C (°F) (760 mm)	Deviations from Actual Boiling Point, °C (°F)	
		(760 mm)	(10 mm)
Benzene	80 (176)	+ 3 (+ 6)	– 2 (–4)
Thiophene	84 (183)	+ 4 (+ 7)	+ 1 (+ 2)
Toluene	111 (231)	+ 2 (+ 3)	– 1 (–2)
<i>p</i> -Xylene	139 (282)	0 (0)	+ 2 (+ 4)
1-Dodecene	213 (416)	0 (0)	0 (0)
Naphthalene	218 (424)	– 11 (–20)	– 4 (–8)
2,3-Benzothiophene	221 (430)	– 13 (–23)	0 (0)
2-Methylnaphthalene	241 (466)	– 12 (–21)	– 2 (–3)
1-Methylnaphthalene	245 (473)	– 12 (–21)	– 1 (–1)
Dibenzothiophene	332 (630)	– 32 (–58)	– 6 (–10)
Phenanthrene	339 (642)	– 35 (–63)	– 9 (–16)
Anthracene	342 (647)	– 36 (–64)	– 8 (–15)
Pyrene	395 (743)	– 48 (–87)	– 16 (–29)
Chrysene	447 (837)	– 60 (–108)	^A

^A No data at 10 mm for chrysene.

TABLE X1.3 Distillation of Heavy Gas Oils

Weight Percent Off ^A	Virgin Gas Oil		High-Sulfur Coker Gas Oil		“Decanted” Oil	
	TBP, ^A °C (°F)	SD, ^B °C (°F)	TBP, °C (°F)	SD, °C (°F)	TBP, °C (°F)	SD, °C (°F)
IBP ^C	230 (446)	215 (419)	223 (433)	209 (409)	190 (374)	176 (348)
10	269 (517)	265 (506)	274 (526)	259 (498)	318 (605)	302 (575)
20	304 (580)	294 (562)	296 (565)	284 (544)	341 (645)	338 (640)
30	328 (622)	321 (610)	316 (600)	312 (593)	357 (675)	358 (676)
40	343 (650)	348 (659)	336 (636)	344 (651)	377 (710)	375 (707)
50	367 (693)	373 (704)	356 (672)	364 (688)	390 (734)	391 (736)
60	394 (742)	409 (749)	377 (710)	386 (727)	410 (770)	409 (768)
70	417 (783)	424 (795)	399 (751)	410 (770)	425 (797)	425 (797)
80	447 (836)	451 (844)	421 (800)	434 (814)	445 (833)	443 (830)
90	...	488 (910)	462 (863)	467 (872)	...	469 (876)
95	...	511 (951)	482 (900)	494 (922)	...	492 (918)
100	...	543 (1009)	...	542 (1007)	...	542 (1007)

^A TBP = True boiling point.

^B SD = Simulated distillation boiling point.

^C IBP = Initial boiling point.

X2. AGREEMENT WITH CONVENTIONAL DISTILLATION (PROCEDURES A AND B)

X2.1 Test Method **D2892** is the standard for conventional distillation of petroleum products.

X2.2 This test method has been compared with Test Method **D2892** on the same samples by a number of laboratories(1-3)⁸.

⁸ The boldface numbers in parentheses refer to the list of references at the end of this standard.

In all cases, agreement between the two test methods has been very good for petroleum products and fractions within the scope of this test method.

X2.3 The time required for analysis by this test method is approximately one tenth of that required for Test Method **D2892**, and Test Method **D2892** has difficulty establishing the IBP and FBP accurately.

X3. CALCULATION OF D86 CORRELATED DATA FROM D2887 DATA (PROCEDURE A ONLY)

X3.1 Correlations

X3.1.1 The resulting data obtained from carrying out an analysis by Test Method D2887 can be used to obtain Test Method D86 data via a correlation. The correlations used to convert selected Test Method D2887 distillation points (percent off) to Test Method D86 (percent off) are mathematical equations. There are two correlations presented in this appendix: the STP 577 correlation and the API correlation.

X3.2 STP 577 Correlation

X3.2.1 The correlation that has been used for a number of years is called the Atlantic Richfield correlation, which was published in an ASTM Special Technical Publication (STP 577) (4). See Table X3.1.

X3.2.2 The application of this correlation was also published by Kennard (5), which showed how the correlation can be optimized for a particular type of sample.

X3.2.3 This correlation has not been subjected to recent ASTM statistical treatment since its origin precedes the newer statistical methodologies. However, a limited number of comparisons of the use of the correlation is presented in Reference (4).

TABLE X3.1 STP 577 Correlation

D86-IBP	46.566 + 0.58289 (D2887 10 %) + 0.34795 (D2887 IBP)
D86-10 %	33.308 + 0.61562 (D2887 10 %) + 0.35110 (D2887 20 %)
D86-20 %	22.411 + 0.48903 (D2887 30 %) + 0.27528 (D2887 20 %) + 0.21713 (D2887 10 %)
D86-30 %	14.431 + 0.47035 (D2887 30 %) + 0.28369 (D2887 20 %) + 0.22784 (D2887 50 %)
D86-50 %	4.876 + 0.97597 (D2887 50 %)
D86-70 %	0.911 + 0.51975 (D2887 80 %) + 0.33260 (D2887 70 %) + 0.10159 (D2887 30 %)
D86-80 %	0.279 + 0.75936 (D2887 80 %) + 0.28333 (D2887 95 %) – 0.09975 (D2887 FBP)
D86-90 %	–1.973 + 0.61459 (D2887 90 %) + 0.31909 (D2887 95 %)
D86-FBP	34.179 + 1.14826 (D2887 95 %) – 0.59208 (D2887 90 %) + 0.31542 (D2887 FBP)

X3.3 API Correlation

X3.3.1 A second correlation that has been used is the API Procedure 3A3.2 (see Reference (6)).

X4. CORRELATION FOR JET AND DIESEL FUEL (PROCEDURES A AND B)

X4.1 The resulting data obtained from carrying out an analysis by Test Method D2887 can be used to obtain Test Method D86 data via a correlation. The correlations used to convert selected Test Method D2887 distillation points (percent off) to Test Method D86 (percent off) are mathematical equations.

X4.2 A correlation model is presented here for the calculation of Test Method D86 correlated data from boiling range distribution analysis by gas chromatography according to Test Method D2887. This correlation model is only valid for diesel and jet fuels, excluding biodiesels.

X4.3 This correlation model was validated by an analysis of variance procedure in accordance with Practice D6708.

X4.4 Significance and Use

X4.4.1 Valid data for conversion to Test Method D86 correlated data can be obtained by Test Method D2887. The model is only valid for diesel or jet fuel and samples must meet the specifications given in Test Method D2887.

X4.5 Summary of the Procedure

X4.5.1 Test Method D86 correlated data is calculated from Test Method D2887 data using Eq X4.1 and the coefficients specified in Table X4.1.

$$t_n = a_o + a_1 \cdot T_{n-1} + a_2 \cdot T_n + a_3 \cdot T_{n+1} \quad (\text{X4.1})$$

where:

- t_n = n th boiling point temperature of Test Method D86 correlated,
- a_i = i th coefficient from Table X4.1, and
- T_n = n th boiling point temperature of D2887.

X4.6 Basis

X4.6.1 This correlation model is based on data for 46 jet fuel samples and 39 diesel fuel samples analyzed in accordance with both Test Method D86 and D2887. From these results, a correlation model was determined using regression, specifying coefficients per recovery. A model of the remaining bias was determined by use of Practice D6708 on a dataset from the ASTM Interlaboratory Crosscheck Program of five jet fuels and six diesels analyzed by 38 laboratories by Test Method D2887 and 201 laboratories by Test Method D86.

X4.6.2 The bias correction model was used to correct the results from the correlation model, resulting in a new correlation matrix given in Table X4.1.

X4.6.3 Based on statistical significance tests, no sample specific biases were observed in the dataset used for the bias correction.

X4.6.4 Both methods were found sufficiently precise to distinguish among the samples.

X4.6.5 *Precision and Bias*⁹—Reproducibility after conversion of Test Method D2887 data into Test Method D86 data is equivalent to the reproducibility of Test Method D2887.

⁹ Supporting data have been filed at ASTM International Headquarters, and may be obtained by requesting Research Reports RR:D02-1553 and D02-1564.

X4.6.6 Cross-method reproducibility after conversion of Test Method D2887 data into Test Method D86 correlated data is given in Table X4.2.

TABLE X4.1 Correlation Coefficients

t_n °C	a_0	a_1	a_2	a_3	T_{n-1}	T_n °C	T_{n+1}
IBP	25.351	0.32216	0.71187	-0.04221	T_{IBP}	T_5	T_{10}
5 %	18.822	0.06602	0.15803	0.77898	T_{IBP}	T_5	T_{10}
10 %	15.173	0.20149	0.30606	0.48227	T_5	T_{10}	T_{20}
20 %	13.141	0.22677	0.29042	0.46023	T_{10}	T_{20}	T_{30}
30 %	5.7766	0.37218	0.30313	0.31118	T_{20}	T_{30}	T_{50}
50 %	6.3753	0.07763	0.68984	0.18302	T_{30}	T_{50}	T_{70}
70 %	-2.8437	0.16366	0.42102	0.38252	T_{50}	T_{70}	T_{80}
80 %	-0.21536	0.25614	0.40925	0.27995	T_{70}	T_{80}	T_{90}
90 %	0.09966	0.24335	0.32051	0.37357	T_{80}	T_{90}	T_{95}
95 %	0.89880	-0.09790	1.03816	-0.00894	T_{90}	T_{95}	T_{FBP}
FBP	19.444	-0.38161	1.08571	0.17729	T_{90}	T_{95}	T_{FBP}

X5. PREVIOUS CHANGES

TABLE X4.2 Cross-Method Reproducibility, °C

R °C	IBP	5 %	10 %	20 %	30 %	50 %	70 %	80 %	90 %	95 %	FBP
R °C	13.71	11.80	10.73	8.83	7.39	6.96	7.03	7.62	8.85	17.32	12.94

X5.1 Subcommittee D02.04 has identified the location of selected changes to this standard since the last issue (D2887–01a) that may impact the use of this standard.

X5.1.1 Changes:

(1) Corrected selected Fahrenheit boiling point values in Table 2 and added a note explaining the derivation of the corrected values.

(2) Corrected Footnote A in Table 2.

(3) Removed Column Resolution requirement.

(4) Added Appendix X3, Calculation Algorithm.

(5) Corrected rate of change of chromatographic signal in note.

(6) Modified paragraph 6.1.5 to address inlet pressure for open tubular columns.

X5.2 Subcommittee D02.04.0H has identified the location of selected changes to this standard since the last issue (D2887–08) that may impact the use of this standard.

X5.2.1 Changes:

(1) Revised Sections 12 and 13.

(2) Deleted original Appendix X3.

REFERENCES

- (1) Green, L. E., Schumauch, L. J., and Worman, J. C., *Analytical Chemistry*, Vol 32, 1960, p. 904.
- (2) Hickerson, J. F., *ASTM STP 577M*, ASTM International, 1973, p. 71.
- (3) Green, L. E., *Chromatograph Gives Boiling Point, Hydrocarbon Processing*, May, 1976.
- (4) Ford, D. C., Miller, W. H., Thren, R. C., and Wetzler, R., “Correlation of ASTM D2887-73 Boiling Range Distribution Data with ASTM Method D86-67D86 Distillation Data” In “Calculation of Physical Properties of Petroleum Products From Gas Chromatographic Analyses,” Ed. by L. E. Green and D. K. Albert, *ASTM STP 577*, ASTM International, 1975, pp. 20-30.
- (5) Kennard, C., “Correlated ASTM Distillation Distribution Based on Simulated Distillation (ASTM D2887) Data,” Hewlett Packard, Avondale, PA, Application Note AN230-5, April 1979.
- (6) API Technical Data Book for Petroleum Refining, Chapter 3, Petroleum Fraction Distillation Interconversions, 1999.

SUMMARY OF CHANGES

Subcommittee D02.04 has identified the location of selected changes to this standard since the last issue (D2887–12) that may impact the use of this standard.

- | | |
|--|--|
| <p>(1) Corrected selected Fahrenheit boiling point values in Table 2 and added a note explaining the derivation of the corrected values.</p> <p>(2) Corrected Footnote A in Table 2.</p> <p>(3) Removed Column Resolution requirement.</p> <p>(4) Added Appendix X3, Calculation Algorithm.</p> <p>(5) Corrected rate of change of chromatographic signal in Note 8.</p> | <p>(6) Modified 6.1.5 to address inlet pressure for open tubular columns.</p> <p>(7) Revised 7.9.</p> <p>(8) Revised 10.4.</p> <p>(9) Added Table 4 and Table 11.</p> <p>(10) Revised Table 3.</p> <p>(11) Revised Table 6.</p> |
|--|--|

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the ASTM website (www.astm.org/COPYRIGHT/).