



Standard Test Method for Heating Value of Gases in Natural Gas and Flare Gases Range by Stoichiometric Combustion¹

This standard is issued under the fixed designation D4891; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers the determination of the heating value of natural gases and similar gaseous mixtures within the range of composition shown in [Table 1](#), and [Table 2](#) that covers flare components but is not intended to limit the components to be measured in flare gases.

1.2 *This standard involves combustible gases. It is not the purpose of this standard to address the safety concerns, if any, associated with their use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:²

[D1826](#) Test Method for Calorific (Heating) Value of Gases in Natural Gas Range by Continuous Recording Calorimeter
[E691](#) Practice for Conducting an Interlaboratory Study to Determine the Precision of a Test Method

2.2 EPA Standard:³

[EPA-600 /2-85-106](#) Evaluation of the Efficiency of Industrial Flares: Flare Head Design and Gas Composition

3. Terminology

3.1 All of the terms defined in Test Method [D1826](#) are included by reference.

3.2 Definitions of Terms Specific to This Standard:

3.2.1 *combustion ratio, n* —the ratio of combustion air to gaseous fuel.

3.2.2 *burned gas parameter, n* —a property of the burned gas after combustion which is a function of the combustion ratio.

3.2.3 *critical combustion ratio, n* —for a specific burned gas parameter, the combustion ratio at which a plot of burned gas parameter versus combustion ratio has either maximum value or maximum slope.

3.2.4 *combustion air requirement index (CARI), n* —is the amount of air required for complete combustion of the gas being measured and can be used to index against other measured values such as the Wobbe Index or Heating Value.

3.2.5 *stoichiometric ratio, n* —the combustion ratio when the quantity of combustion air is just sufficient to convert all of the combustibles in the fuel to water and carbon dioxide.

4. Summary of Test Method

4.1 Air is mixed with the gaseous fuel to be tested. The mixture is burned and the air-fuel ratio is adjusted so that essentially a stoichiometric proportion of air is present. More exactly, the adjustment is made so that the air-fuel ratio is in a constant proportion to the stoichiometric ratio that is a relative measure of the heating value. To set this ratio, a characteristic property of the burned gas is measured, such as temperature or oxygen concentration.

5. Significance and Use

5.1 This test method provides an accurate and reliable procedure to measure the total heating value of a fuel gas, on a continuous basis, which is used for regulatory compliance, custody transfer, and process control.

5.2 Some instruments which conform to the requirements set forth in this test method can have response times on the order of 1 min or less and can be used for on-line measurement and control.

5.3 The method is sensitive to the presence of oxygen and nonparaffin fuels. For components not listed and composition ranges that fall outside those in [Table 1](#) and [Table 2](#), modifications in the method and changes to the calibration gas or gases being used may be required to obtain correct results.

6. Apparatus

6.1 A suitable apparatus for carrying out the stoichiometric combustion method will have at least the following four

¹ This test method is under the jurisdiction of ASTM Committee [D03](#) on Gaseous Fuels and is the direct responsibility of Subcommittee [D03.03](#) on Determination of Heating Value and Relative Density of Gaseous Fuels.

Current edition approved May 1, 2013. Published May 2013. Originally approved in 1989. Last previous edition approved in 2001 as D4891–89 (2001). DOI: 10.1520/D4891-89R06.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Available from United States Environmental Protection Agency (EPA), Ariel Rios Bldg., 1200 Pennsylvania Ave., NW, Washington, DC 20004, <http://www.epa.gov>.

TABLE 1 Natural Gas Components and Range of Composition Covered

Compound	Concentration Range, mole, %
Helium	0.01 to 5
Nitrogen	0.01 to 20
Carbon dioxide	0.01 to 10
Methane	50 to 100
Ethane	0.01 to 20
Propane	0.01 to 20
<i>n</i> -butane	0.01 to 10
isobutane	0.01 to 10
<i>n</i> -pentane	0.01 to 2
Isopentane	0.01 to 2
Hexanes and heavier	0.01 to 2

components: flow meter or regulator, or both; combustion chamber; burned gas sensor; and electronics. The requirement for each of these components is discussed below. The detailed design of each of these components can vary. Three different apparatus are shown in Fig. 1, Fig. 2 and Fig. 3. In each figure the equivalent of the four necessary components are enclosed in dashed lines.

6.2 Overview—Air and fuel enter the apparatus and the flow of each is measured. Alternatively, only one gas flow need be measured if the flow of the other is kept the same during measurement and calibration. This is illustrated in Fig. 2. Next there is a combustion chamber in which the air and fuel are mixed and burned. This can be as simple as a bunsen or meeker burner, but precautions should be taken that subsequent measurements of burned gas characteristics are not influenced by ambient conditions. Finally, there is a sensor in the burned gas which measures a property of this gas that is sensitive to the combustion ratio and has a unique feature at the stoichiometric ratio. Two such properties are temperature and oxygen concentrations, and either can be measured.

6.3 Flow Meter or Regulator, or both—The flow measurement part of the apparatus should have an accuracy and precision of the order of 0.1 %. Likewise, if the flow is to be kept constant, the flow regulator should maintain this constant value within 0.1 %. The meter or regulator for natural gas must maintain this precision and accuracy over the density and viscosity ranges consistent with the composition range in Table 1 or Table 2.

6.4 Combustion Chamber:

6.4.1 There are two different types of combustion chambers that may be used. In the first type the air and fuel are mixed and burned in a single burner. The apparatus shown in Fig. 1 has this type of combustion chamber.

6.4.2 In the second type of combustion chamber, the air and fuel are each divided into two streams, and combustion takes place simultaneously in two burners. The division of air flow must be such that the proportion of air going to each burner always remains the same. Likewise the division of fuel flow must always remain the same even through fuel composition changes. Another requirement is that the flow divisions be such that one burner has a mixture with a slightly higher combustion ratio than the other. The apparatus shown in Fig. 2 has this type of combustion chamber.

TABLE 2 Natural Gas Components and Range of Composition Covered^A

Compound	CAS Number
Volatile Analytes	
Acetone	67-64-1
Acetonitrile	75-05-8
Acrolein	107-05-8
Acrylonitrile	107-13-1
Benzene	71-43-2
1,3-Butadiene	106-99-0
Carbon disulfide	75-15-0
Chlorobenzene	108-90-7
Cumene	98-82-8
(isopropyl)benzene	
1,2-Dibromoethane	106-93-4
Ethylbenzene	100-41-4
Hexane	110-54-3
Methanol	67-56-1
Methyl isobutyl ketone	108-10-1
Methyl t-butyl ether	1634-04-4
Methylene chloride	75-09-2
Nitrobenzene	98-95-3
Nitropropane	79-46-9
Pentane2	109-66-0
Styrene	100-42-5
Tetrachloroethene	127-18-4
Toluene	108-88-3
Trichloroethene	79-01-6
Trimethylpentane	2 540-84-1
Xylenes (mixed isomers)	1330-20-7
Trimethylpentane	2 540-84-1
Xylenes (mixed isomers)	1330-20-7
Semi-volatile Analytes	
Acenaphthene	83-32-9
Acenaphthylene	208-96-8
Aniline	62-53-3
Anthracene	120-12-7
Benidine1	92-87-5
Benz[a]anthracene	56-55-3
Benzo[b]fluoranthene	205-99-2
Benzo[k]fluoranthene	207-08-9
Benzo[g,h,i]perylene	191-24-2
Benzo[a]pyrene	50-32-8
Benzo[e]pyrene2	192-97-2
Biphenyl2	92-52-4
Cresol (mixed isomers)	1319-77-3
Chrysene	218-01-9
Dibenz[a,h]anthracene	53-70-3
Dibenzofuran	132-64-9
Dibenzo(a,e)pyrene	192-65-4
3,3'- Dimethoxybenzidine	119-90-4
Dimethylaminobenzene	60-11-7
7,12-	57-97-6
Dimethylbenz(a)anthracene	
3,3'- Dimethylbenzidine	119-93-7
á,á-	122-09-8
Dimethylphenethylamine	
2,4-Dimethylphenol	105-67-9
Fluoranthene	206-44-0
Fluorene	86-73-7
Indeno(1,2,3-cd)pyrene	193-39-5
Isophorone	78-59-1
3-Methylcholanthrene	56-49-5
2-Methylnaphthalene	91-57-6
Naphthalene	91-20-3
Perylene2	198-55-0
Phenanthrene	85-01-8
Phenol	108-95-2
1,4-Phenylenediamine	106-50-3
Pyrene	129-00-0
o-Toluidine	95-53-4
Aldehydes	
Methanol	67-56-1
Formaldehyde	50-00-0
Acetaldehyde	75-07-0

TABLE 2 Continued

Compound	CAS Number	
Propanal	123-38-6	
C1 to C5 Hydrocarbons		
Description	Compound	CAS Number
C1 Alkanes	Methane	74-82-8
C2 Alkanes	Ethane	74-84-0
C3 Alkanes	Propane	74-98-6
C4 Alkanes	n-Butane	106-97-8
	Isobutane	75-28-5
	n-Pentane	109-66-0
C5 Alkanes	2-Methylbutane	78-78-4
	Cyclopentane	287-92-3
C2 Olefins	Ethylene	74-85-1
C2 Alkanes	Acetylene	74-86-2
C3 Olefins	Propylene	115-07-1
C4 Olefins	1-Butene	106-98-9
	2-Butene	107-01-7
	Isobutene	115-11-7
C5 Olefins	1-Pentene	109-67-1
	Cis-2-pentene	627-20-3
	Trans-2-pentene	646-04-8
	2-Methyl-1-butene	563-46-2
	3-Methyl-1-butene	563-45-1
	2-Methyl-2-butene	513-35-9
	Cyclopentene	142-29-0
	Propadiene	463-49-0
	1,2-Butadiene	590-19-2
	1,3-Butadiene	106-99-0
C5 Alkadienes	1,2-Pentadiene	591-95-7
	1-cis-3-Pentadiene	1574-41-0
	1-trans-3- Pentadiene	2004-70-8
	1,4-Pentadiene	591-93-5
	2,3-Pentadiene	591-96-8
	3-Methyl-1,2- butadiene	598-25-4
	2-Methyl-1,3- butadiene	78-79-5
	Cyclopentadiene	542-92-7
Heating Value Range		
Unit	Lower	Upper
Btu/ft³	83	2350

^aFlare Gas Heating Value range defined in Table 2 is derived from the Evaluation of the Efficiency of Industrial Flares: Flare Head Design and Gas Composition EPA-600 /2-85-106 September 1985 Table 1-1. Agency Information Collection Activities OMB Responses EPA ICR Number 2411.01; NSPS and NESHAP for Petroleum Refineries Sector Residual Risk and Technology; OMB Number 2060-0657.




FIG. 1 Gas Btu Transmitter (Functional Overview)

6.4.3 A third type utilizes a combustion oven operating in excess of 800°C (1472°F) to assure the combustion of gases within the natural or flare gas compositions being combusted as shown in Fig. 3.




FIG. 2 Stoichiometric Combustion Apparatus

6.5 Burned Gas Sensor:

6.5.1 The burned or combusted gas sensor must measure a characteristic of the burned gas which is a function of the combustion ratio and for which there is a critical combustion ratio related to the stoichiometric ratio. A combustion chamber of the first type (Fig. 1) would have one sensor in the burned gas and its output signal would constitute the desired measurement. In a combustion chamber of the second type (Fig. 2) there would be a sensor in the burned gas from each burner. The difference between the two output signals would constitute the desired measurement. In the third type (Fig. 3), the residual oxygen is measured and the resulting oxygen value is correlated to the CARI and Wobbe Index.

6.5.2 There are several properties of the burned gas which are related uniquely to the combustion ratio. A burned gas sensor may be selected which provides a measure of any one of these, for example, either temperature or oxygen partial pressure.

6.6 *Electronics*—Electronics are used to receive the signals from the components described above to control the flow of gases into the combustion chamber in response to the signal from the burned gas sensor and to provide a digital or analog output signal, or both, which is proportional to the heating value of the gaseous fuel.

6.7 *Temperature Stability and Operating Environment*—The method is capable of operating over a range of temperatures limited only by the specific apparatus used to realize the method. It is desirable to equilibrate the air and fuel temperatures before the gases are measured. The electronics should also be stabilized against temperature changes and the burned gas sensor should be insensitive to changes in the ambient conditions.

7. Reagents and Materials

7.1 *Physical Contamination*—The air and gas must be free of dust, liquid, water, liquid hydrocarbons, and other entrained solids. Foreign materials should be removed by a sample line filter. To avoid any problems in the line from any liquid accumulation, pitch the line to a low point and provide a drip leg.




FIG. 3 Residual Oxygen Stoichiometric Combustion Apparatus

7.2 Chemical Contamination—The air must be free of combustible compounds. The oxygen content and the absolute humidity of the air should be the same during measurement as during calibration.

8. Calibration and Standardization

8.1 The calibration factor, F , and the constant, B , in the equation, $C = F \cdot R + B$, are determined through an initial calibration, in which the critical combustion ratios of at least two standard gases of known but different heating values are measured using the procedure described in 9.1.

8.2 The calibration factor, F , is routinely redetermined at regular intervals under field conditions using a calibration gas of known heating value. The constant, B , is not adjusted in the routine calibrations. The interval between routine calibrations must be determined under the specific operating conditions, and is usually of the order of 24 h. Determination of F establishes the amount of net oxygen per standard volume of combustion air. Variations in net oxygen constant can be caused by several factors, such as changes in absolute humidity or the presence of contaminants in the air supply.

8.3 The calibration when utilizing multiple calibration gases in which these standards are used for calibrating the low, and in some calorimeter configurations the mid point of the expected measurement range for Heating Value, Specific Gravity (Relative Density; where Air = 1.0000). Refer to the calorimeters manufacturers manual for the proper calibration procedure. The interval between routine calibrations must be determined under the specific operating conditions, and can vary between 24 hours to 96 hours as determined by the operators to be appropriate for the users application.

8.4 Standardization of the calorimeters reported values such as Higher and Lower Heating Value, Specific Gravity, Density, Combustion Air Requirement Index (CARI) are based on the compositional makeup of the calibration gas standard(s) used to calibrate the calorimeter and the use of pressure base and temperature are use to calculate the values such as the higher or lower heating values.

9. Procedure

9.1 Measure the burned gas parameter at different combustion ratios and determine that combustion ratio for which the parameter has a specified characteristic such as a maximum, minimum, or maximum rate of change.

9.1.1 Use an apparatus such as in Fig. 1 or Fig. 2 where the components of the apparatus meet the requirements of Section 6. If the apparatus has two flow meters, the combustion ratio is the ratio of the output of the air flow meter divided by the output of the fuel flow meter. If the apparatus has only one flow meter, then the combustion ratio is set numerically equal to either the output of the air flow meter or the reciprocal of the output of the fuel flow meter. The burned gas parameter is a function of the combustion ratio and is measured at different combustion ratios. The critical combustion ratio, R , is taken as that point where this function has a maximum value, minimum value, or maximum rate of change. The heating value, C , is calculated from the equation

$$C = F \cdot R + B, \quad (1)$$

where the constants B and F are determined as described in 8.1 and 8.2.

9.1.2 This procedure may be automated, for example, by using a microprocessor in the electronics.

9.2 For making laboratory measurements of highest precision, use the following procedure:

9.2.1 First calibrate the instrument as described in 8.2. Then, before measuring the test gases, measure two other standard gases of known heating value. After the test gas measurements, measure the two standard gases again. The known heating values of these standard gases, CAL.VAL.LOW and CAL.VAL.HIGH, should bracket that of the unknown gas. Combine the measured values of the standard gases and the test gases to obtain a best estimate of the heating value of the test gas. Do this using the following calculation procedure.

9.2.1.1 *Step 1*—There are four measured values for the calibration gases, two for the high calorific gas and two for the low calorific gas. Average these four measurements together. The result is represented by the symbol, AV.STD.GASES.

9.2.1.2 *Step 2*—Average the two known heating values of the standard gases together. The result is represented by the symbol, AV.CAL.VAL. Thus, $AV.CAL.VAL = [(CAL.VAL.HIGH) + (CAL.VAL.LOW)]/2$.

9.2.1.3 *Step 3*—Calculate a correction to the test gas measurements. This correction is represented by the symbol, CORR. The calculation is as follows: $CORR = (AV.STD.GASES) - (AV.CAL.VAL)$.

9.2.1.4 *Step 4*—Subtract the quantity, CORR, that is calculated in Step 3 from each of the test gas measurements to give the corrected value.

9.2.2 *Example 1*—Standard gas low has CAL.VAL.LOW = 1000 Btu/standard cubic foot and measured values after calibration are 1002.0 and 1002.8. (All heating values in Example 1 and Example 2 have units of Btu per standard cubic foot.) Standard gas high has CAL.VAL.HIGH = 1200 and measured values of 1202.0 and 1203.2.

$$AV.STD.GASES = (1002.0 + 1002.8 + 1202.0 + 1203.2)/4 = 1102.5.$$

$$AV.CAL.VAL = (1000.0 + 1200.0)/2 = 1100.0$$

$$CORR = 1102.5 - 1100 = 2.5$$

$$TEST\ GAS\ MEASUREMENT = 1080.6$$

$$CORRECTED\ VALUE = (1080.6 - 2.5) = 1078.1$$

9.2.3 *Example 2*:

CAL.VAL.LOW = 1000 Btu/standard cubic foot.

Measured values are 998.0 and 998.2.

CAL.VAL.HIGH = 1200

Measured values are 1199.0 and 1199.2

$$AV.STD.VAL = (998.0 + 998.2 + 1199.0 + 1199.2)/4 = 1098.6$$

$$AV.CAL.VAL = 1100$$

$$CORR = (1098.6 - 1100) = -1.4$$

$$TEST\ GAS\ MEASUREMENT = 1076.7$$

$$CORRECTED\ VALUE = [1076.7 - (-1.4)] = 1078.1$$

9.3 For use of two or three calibration gas standards for the calibration process, see the manufactures manual for the correct calibration procedure and validation check of the calibration gases used.

10. Precision and Bias

10.1 To determine precision and bias, an interlaboratory study was carried out using two types of commercial instruments that implement the stoichiometric method. For each type of instrument six different laboratories each measured five different reference gases. Cylinders containing these reference gases were transported from laboratory to laboratory. Each laboratory used its own instrument and personnel to measure the heating values of the gases in these cylinders. The same calibration gas was used to calibrate each instrument.

10.2 The heating values of the reference gases were determined prior to the study by the Institute of Gas Technology. These values were established by averaging three recording calorimeter measurements. The values were unknown to the participants in the interlaboratory test program. At the end of the study, the heating values were remeasured at the Institute of Gas Technology to establish that the gas compositions did not change. The statistical analysis of the results was in accordance with the procedures in Practice E691.

10.2.1 *Repeatability*—The root mean square estimate of the within laboratory component of standard deviation was 0.76 Btu/standard cubic foot. The corresponding 95 % confidence repeatability interval was 2.1 Btu/standard cubic foot.

10.2.2 *Reproducibility*—The root mean square estimate of the between laboratory component of standard deviation was 1.67 Btu/standard cubic foot. The corresponding 95 % confidence reproducibility interval was 5.1 Btu/standard cubic foot.

10.2.3 *Bias*—The average of all measurements agreed with the average reference value within 0.1 %.

11. Keywords

11.1 calorimeter; flare gas; heating value; natural gas; stoichiometric conversion

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the ASTM website (www.astm.org/COPYRIGHT/).