



Standard Test Method for Determination of Aldicarb, Carbofuran, Oxamyl and Methomyl by Liquid Chromatography/Tandem Mass Spectrometry¹

This standard is issued under the fixed designation D7600; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

^{ε1} NOTE—This Test Method was changed editorially in 2012.

1. Scope

1.1 This procedure covers the determination of aldicarb, carbofuran, oxamyl and methomyl (referred to collectively as carbamates in this test method) in surface water by direct injection using liquid chromatography (LC) and detected with tandem mass spectrometry (MS/MS). These analytes are qualitatively and quantitatively determined by this method. This method adheres to multiple reaction monitoring (MRM) mass spectrometry.

1.2 This test method has been developed by US EPA Region 5 Chicago Regional Laboratory (CRL).

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 The Detection Verification Level (DVL) and Reporting Range for the carbamates are listed in [Table 1](#).

1.4.1 The DVL is required to be at a concentration at least 3 times below the Reporting Limit (RL) and have a signal/noise ratio greater than 3:1. [Fig. 1](#) displays the signal/noise ratios of the primary single reaction monitoring (SRM) transitions and [Fig. 2](#) displays the confirmatory SRM transitions at the DVLs for the carbamates.

1.4.2 The reporting limit is the concentration of the Level 1 calibration standard as shown in [Table 2](#) for the carbamates.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

¹ This test method is under the jurisdiction of ASTM Committee D19 on Water and is the direct responsibility of Subcommittee D19.06 on Methods for Analysis for Organic Substances in Water.

Current edition approved Dec. 1, 2009. Published February 2012. DOI: 10.1520/D7600-09E01.

2. Referenced Documents

2.1 ASTM Standards:²

- D1129 Terminology Relating to Water
- D1193 Specification for Reagent Water
- D2777 Practice for Determination of Precision and Bias of Applicable Test Methods of Committee D19 on Water
- D3856 Guide for Management Systems in Laboratories Engaged in Analysis of Water
- D3694 Practices for Preparation of Sample Containers and for Preservation of Organic Constituents
- D5847 Practice for Writing Quality Control Specifications for Standard Test Methods for Water Analysis
- E2554 Practice for Estimating and Monitoring the Uncertainty of Test Results of a Test Method in a Single Laboratory Using a Control Sample Program

2.2 Other Documents:

- EPA publication SW-846 Test Methods for Evaluating Solid Waste, Physical/Chemical Methods³

3. Terminology

3.1 Definitions:

3.1.1 *detection verification level (DVL), n*—a concentration that has a signal/noise ratio greater than 3:1 and is at least 3 times below the reporting limit (RL).

3.1.2 *reporting limit (RL), n*—the concentration of the lowest-level calibration standard used for quantification.

3.1.3 *carbamates, n*—in this test method, aldicarb, carbofuran, oxamyl and methomyl collectively.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Available from National Technical Information Service (NTIS), U.S. Department of Commerce, 5285 Port Royal Road, Springfield, VA, 22161 or at <http://www.epa.gov/epawaste/hazard/testmethods/index.htm>.

TABLE 1 Detection Verification Level and Reporting Range

Analyte	DVL (µg/L)	Reporting Range (µg/L)
Aldicarb	100	1–100
Carbofuran	100	1–100
Oxamyl	100	1–100
Methomyl	100	1–100

3.2 Abbreviations:

3.2.1 *ppt*—parts per trillion, ng/L

3.2.2 *ND*—non-detect

4. Summary of Test Methods

4.1 This is a performance-based method and modifications are allowed to improve performance.

4.2 For carbamate analysis, samples are shipped to the lab between 0°C and 6°C and analyzed within 7 days of collection. In the lab, the samples are spiked with surrogate, filtered using a syringe driven Millex HV PVDF filter unit and analyzed directly by LC/MS/MS.

4.3 Aldicarb, carbofuran, oxamyl, methomyl and 4-bromo-3,5-dimethylphenyl-*N*-methylcarbamate (BDMC, surrogate) are identified by retention time and two SRM transitions. The target analytes and surrogate are quantitated using the primary SRM transitions utilizing an external calibration. The final report issued for each sample lists the concentration of aldicarb, carbofuran, oxamyl, methomyl and the BDMC surrogate recovery.

5. Significance and Use

5.1 The *N*-methyl carbamate (NMC) pesticides: aldicarb, carbaryl, carbofuran, formetanate hydrochloride, methiocarb, methomyl, oxamyl, pirimicarb, propoxur, and thiodicarb have been identified by EPA as working through a common mechanism. They affect the nervous system by reducing the ability of the enzyme cholinesterase. Cholinesterase inhibition was the primary toxicological effect of regulatory concern to EPA in assessing the NMC's food, drinking water and residential risks. In most of the country, NMC residues in drinking water sources are at levels that are not likely to contribute substantially to the multi-pathway cumulative exposure. Shallow private wells extending through highly permeable soils into shallow, acidic ground water represent what the EPA believes to be the most vulnerable drinking water.⁴

5.2 This method has been investigated for use with reagent and surface water for the selected carbamates: aldicarb, carbofuran, oxamyl and methomyl.

6. Interferences

6.1 Method interferences may be caused by contaminants in solvents, reagents, glassware and other apparatus producing discrete artifacts or elevated baselines. All of these materials are demonstrated to be free from interferences by analyzing laboratory reagent blanks under the same conditions as samples.

⁴ Additional information about carbamate pesticides can be found on the Internet at <http://www.epa.gov> (2009).

6.2 All glassware is washed in hot water with a detergent, rinsed in hot water followed by distilled water. The glassware is then dried and heated in an oven at 250°C for 15 to 30 minutes. All glassware is subsequently cleaned with acetone, then methanol.

6.3 All reagents and solvents should be pesticide residue purity or higher to minimize interference problems.

6.4 Matrix interferences may be caused by contaminants that are co-extracted from the sample. The extent of matrix interferences can vary considerably from sample source depending on variations of the sample matrix.

7. Apparatus

7.1 LC/MS/MS System

7.1.1 *Liquid Chromatography (LC) System*—A complete LC system is needed in order to analyze samples.⁵ A system that is capable of performing at the flows, pressures, controlled temperatures, sample volumes and requirements of the standard may be used.

7.1.2 *Analytical Column-Waters*—XBridge C18, 150 mm × 2.1 mm, 3.5 µm particle size, or equivalent.

7.1.3 *Tandem Mass Spectrometer (MS/MS) System*—A MS/MS system capable of MRM analysis.⁶ A system that is capable of performing at the requirements in this standard may be used.

7.2 Filtration Device

7.2.1 *Hypodermic syringe*—A lock tip glass syringe capable of holding a Millex HV Syringe Driven Filter Unit PVDF 0.45 µm (Millipore Corporation, Catalog # SLHV033NS) or similar may be used.

7.2.1.1 A 25-mL lock tip glass syringe size is recommended since a 25-mL sample size is used in this test method.

7.2.2 *Filter*—Millex HV Syringe Driven Filter Unit PVDF 0.45 µm (Millipore Corporation, Catalog # SLHV033NS) or similar may be used.

8. Reagents and Materials

8.1 *Purity of Reagents*—High-performance liquid chromatography (HPLC) pesticide residue analysis and spectrophotometry grade chemicals shall be used in all tests. Unless indicated otherwise, it is intended that all reagents shall conform to the Committee on Analytical Reagents of the American Chemical Society.⁷ Other reagent grades may be used provided they are first determined to be of sufficiently high purity to permit their use without affecting the accuracy of the measurements.

8.2 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean reagent water conforming

⁵ A Waters Alliance High Performance Liquid Chromatography (HPLC) System was used to develop this test method. The multi-laboratory study included Agilent, Thermo Electron and Waters LC systems.

⁶ A Waters Quattro micro API mass spectrometer was used to develop this test method. The multi-laboratory study included Agilent, Applied Biosystems, Thermo Electron, Varian and Waters mass spectrometers.

⁷ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For Suggestions on the testing of reagents not listed by the American Chemical Society, see *Annual Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.




FIG. 1 Example Primary SRM Chromatograms Signal/Noise Ratios




FIG. 2 Example Confirmatory SRM Chromatograms Signal/Noise Ratios

TABLE 2 Concentrations of Calibration Standards (PPB)

Analyte/Surrogate	LV 1	LV 2	LV 3	LV 4	LV 5	LV 6
Aldicarb	1	5	10	25	50	100
Carbofuran	1	5	10	25	50	100
Oxamyl	1	5	10	25	50	100
Methomyl	1	5	10	25	50	100
BDMC (Surrogate)	2	10	20	50	100	200

to Type 1 of Specification **D1193**. It must be demonstrated that this water does not contain contaminants at concentrations sufficient to interfere with the analysis.

8.3 *Gases*—Ultrapure nitrogen and argon.

8.4 Acetonitrile (CAS # 75-05-8).

8.5 Methanol (CAS # 67-56-1).

8.6 Acetone (CAS # 67-64-1).

8.7 Ammonium acetate (CAS # 631-61-8).

8.8 Ammonium hydroxide (Concentrated, CAS # 1336-21-6).

8.9 Aldicarb (CAS # 116-06-3).

8.10 Carbofuran (CAS # 1563-66-2).

8.11 Oxamyl (CAS # 23135-22-0).

8.12 Methomyl (CAS # 16752-77-5).

8.13 4-Bromo-3,5-dimethylphenyl-N-methylcarbamate (BDMC, CAS # 672-99-1).

8.13.1 BDMC is used as a surrogate in this standard.

9. Hazards

9.1 Normal laboratory safety applies to this method. Analysts should wear safety glasses, gloves, and lab coats when working in the lab. Analysts should review the Material Safety Data Sheets (MSDS) for all reagents used in this method.

10. Sampling

10.1 *Sampling*—Grab samples must be collected in ≥25-mL pre-cleaned amber glass bottles with Teflon-lined caps demonstrated to be free of interferences. This test method requires a 25-mL sample size per analysis. Conventional sampling practices should be followed. Refer to Guide **D3856** and Practices **D3694**.

10.2 *Preservation*—Store samples between 0°C and 6°C from the time of collection until analysis. Analyze the sample within 7 days of collection.

11. Preparation of LC/MS/MS

11.1 *LC Chromatograph Operating Conditions*⁵:

11.1.1 Injection volumes of all calibration standards and samples are 100 µL. The first sample analyzed after the calibration curve is a blank to ensure there is no carry-over. The gradient conditions for the liquid chromatograph are shown in **Table 3**.

11.1.2 *Temperatures*—Column, 30°C; Sample compartment, 15°C.

11.1.3 *Seal Wash*—Solvent: 50 % Acetonitrile/50 % Water; Time: 5 minutes.

11.1.4 *Needle Wash*—Solvent: 50 % Acetonitrile/50 % Water; Normal wash, approximately 13 second wash time.

11.1.5 *Autosampler Purge*—Three loop volumes.

TABLE 3 Gradient Conditions for Liquid Chromatography

Time (min)	Flow (µL/min)	Percent CH ₃ CN	Percent 95 % Water/ 5 % CH ₃ CN	Percent 50 mmolar NH ₄ OAc/NH ₄ OH in 95 % Water/5 % CH ₃ CN
0	300	0	95	5
2	300	0	95	5
4	300	30	65	5
6	300	35	60	5
8	300	35	60	5
10	300	75	20	5
11.5	300	75	20	5
12	300	95	0	5
18	300	95	0	5
20	300	0	95	5
23	300	0	95	5

11.1.6 Specific instrument manufacturer wash/purge specifications should be followed in order to eliminate sample carry-over in the analysis of carbamates.

11.2 *Mass Spectrometer Parameters*⁶:

11.2.1 In order to acquire the maximum number of data points per SRM channel while maintaining adequate sensitivity, the tune parameters may be optimized according to your instrument. Each peak requires at least 10 scans per peak for adequate quantitation. This standard contains only one surrogate and four target compounds. The MRM experiment windows were set to acquire methomyl and oxamyl in one experiment window while aldicarb, carbofuran and BDMC are in their individual MRM experiment windows. This is required because the chromatographic resolution separating oxamyl and methomyl was not achieved. Variable parameters regarding retention times, SRM Transitions and cone and collision energies are shown in **Table 4**.

The instrument is set in the Electrospray (+) positive setting.

Capillary Voltage: 3.5 kV

Cone: Variable depending on analyte (**Table 4**)

Extractor: 2 Volts

RF Lens: 0.2 Volts

Source Temperature: 120°C

Desolvation Temperature: 300°C

Desolvation Gas Flow: 500 L/hr

Cone Gas Flow: 25 L/hr

Low Mass Resolution 1: 14.5

High Mass Resolution 1: 14.5

Ion Energy 1: 0.5

Entrance Energy: -1

Collision Energy: Variable depending on analyte (**Table 4**)

Exit Energy: 2

Low Mass Resolution 2: 15

High Mass resolution 2: 15

Ion Energy 2: 0.5

Multiplier: 650

Gas Cell Pirani Gauge: 3.3×10^{-3} Torr

Inter-Channel Delay: 0.02 seconds

Inter-Scan Delay: 0.1 seconds

Repeats: 1

Span: 0 Daltons

Dwell: 0.1 Seconds

12. Calibration and Standardization

12.1 The mass spectrometer must be calibrated per manufacturer specifications before analysis. In order that analytical values obtained using this test method are valid and accurate within the confidence limits of the test method, the following procedures must be followed when performing the test method.

TABLE 4 Retention Times, SRM Ions, and Analyte-Specific Mass Spectrometer Parameters

Analyte	Primary/ Confirmatory	Retention time (min)	Cone Voltage (Volts)	Collision Energy (eV)	SRM Mass Transition (Parent > Product)	Collision Energy (eV)
Aldicarb	Primary Confirmatory	11.00	10	7	208.2 > 115.9	2.12
			10	15	208.2 > 88.7	
Carbofuran	Primary Confirmatory	12.85	27	12	222.2 > 165.2	1.20
			27	20	222.2 > 123	
Oxamyl	Primary Confirmatory	8.25	15	8	237.2 > 71.6	2.38
			15	8	237.2 > 89.8	
Methomyl	Primary Confirmatory	8.45	17	8	163.1 > 87.7	1.58
			17	8	163.1 > 105.8	
BDMC (Surrogate)	Primary Confirmatory	14.50	25	24	258.1 > 122	1.31
			25	9	258.1 > 201.2	

12.2 Calibration and Standardization—To calibrate the instrument, analyze six calibration standards containing the six concentration levels of the carbamates and BDMC surrogate prior to analysis as shown in **Table 2**. A calibration stock standard solution is prepared from standard materials or purchased as certified solutions. Stock standard solution A (Level 6) containing aldicarb, carbofuran, oxamyl, methomyl and BDMC is prepared at Level 6 concentration and aliquots of that solution are diluted to prepare Levels 1 through 5. The following steps will produce standards with the concentration values shown in **Table 2**. The analyst is responsible for recording initial component weights carefully when working with pure materials and correctly carrying the weights through the dilution calculations.

12.2.1 Prepare stock standard solution A (Level 6) by adding to a 100-mL volumetric flask individual methanol solutions of the following: 50 μ L of aldicarb, carbofuran, oxamyl and methomyl each at 0.2 g/L and 50 μ L of BDMC at 0.4 g/L, dilute to 100 mL with 90% water/10% methanol. The preparation of the Level 6 standard can be accomplished using different volumes and concentrations of stock solutions as is accustomed in the individual laboratory. Depending on stock concentrations prepared, the solubility at that concentration will have to be ensured.

12.2.2 Aliquots of Solution A are then diluted with 90 % water/10 % methanol to prepare the desired calibration levels in 2-mL amber glass LC vials. The calibration vials must be used within 24 hours to ensure optimum results. Stock calibration standards are routinely replaced every 7 days if not previously discarded for quality control failure. Calibration standards are not filtered.

12.2.3 Inject each standard and obtain a chromatogram for each one. An external calibration technique is used monitoring the primary and confirmatory SRM transition of each analyte. Calibration software is utilized to conduct the quantitation of the target analytes and surrogate using the primary SRM transition. The ratios of the primary/confirmatory SRM transition area counts are given in **Table 4**. These are given as informative and will vary depending on the individual tuning conditions. The primary/confirmatory SRM transition area ratio must be within 30 % of the individual labs accepted primary/confirmatory SRM transition area ratio. The primary SRM transition of each analyte is used for quantitation and the

confirmatory SRM transition for confirmation. This gives added confirmation by isolating the parent ion, fragmenting it into two product ions, and relating it to the retention time in the calibration standard.

12.2.4 The calibration software manual should be consulted to use the software correctly. The quantitation method is set as an external calibration using the peak areas in ppt or ppb units as long as the analyst is consistent. Concentrations may be calculated using the data system software to generate linear regression or quadratic calibration curves. Forcing the calibration through the origin is not recommended.

12.2.5 Linear calibration may be used if the coefficient of determination, r^2 , is >0.98 for the analyte. The point of origin is excluded and a fit weighting of $1/X$ is used in order to give more emphasis to the lower concentrations. If one of the calibration standards other than the high or low point causes the r^2 of the curve to be <0.98 , this point must be re-injected or a new calibration curve must be regenerated. If the low or high point is excluded, minimally a five-point curve is acceptable but the reporting range must be modified to reflect this change.

12.2.6 Quadratic calibration may be used if the coefficient of determination, r^2 , is >0.99 for the analyte. The point of origin is excluded and a fit weighting of $1/X$ is used in order to give more emphasis to the lower concentrations. If one of the calibration standards causes the curve to be <0.99 this point must be re-injected or a new calibration curve must be regenerated. Minimally a six-point curve is acceptable using a quadratic fit. Each calibration point used to generate the curve must have a calculated percent deviation less than 25 % from the generated curve.

12.2.6.1 An initial seven-point curve over the calibration range is an option in the event that the low or high point must be excluded to obtain a coefficient of determination >0.99 . In this event, the reporting range must be modified to reflect this change.

12.2.7 The retention time window of the SRM transitions must be within 5 % of the retention time of the analyte in a midpoint calibration standard. If this is not the case, re-analyze the calibration curve to determine if there was a shift in retention time during the analysis and the sample needs to be re-injected. If the retention time is still incorrect in the sample, refer to the analyte as an unknown.

12.2.8 A midpoint calibration check standard must be analyzed at the end of each batch of 20 samples or within 24 hours after the initial calibration curve was generated. This end calibration check should be the same calibration standard that was used to generate the initial curve. The results from the end calibration check standard must have a percent deviation less than 30 % from the calculated concentration for the target analytes and surrogate. If the results are not within these criteria, the problem must be corrected and either: all samples in the batch must be re-analyzed against a new calibration curve, or the affected results must be qualified with an indication that they do not fall within the performance criteria of the test method. If the analyst inspects the vial containing the end calibration check standard and notices that the sample evaporated affecting the concentration, a new end calibration check standard may be made and analyzed. If this new end calibration check standard has a percent deviation less than 30 % from the calculated concentration for the target analytes and surrogate the results may be reported unqualified.

12.3 All samples are prepared using Class A glass volumetric glassware. The sample volume used throughout this test method is 25 mL. Every sample, the entire 25 mL volume, is filtered through the filtration device described in Section 7.2 only after all required spiking solutions are added and mixed throughout the sample.

12.3.1 A new filter unit is used for each sample. The syringe must be cleaned between each filtration. It is the analyst's responsibility to ensure that the syringe is clean. A possible way of cleaning the syringe between filtrations is first by rinsing with at least 5 syringe volumes of water, followed by at least 3 volumes of acetone, then 3 volumes of methanol and finally rinsed with water to remove any residual solvent.

12.4 If a laboratory has not performed the test before or if there has been a major change in the measurement system, for example, new analyst, new instrument, etc., perform a precision and bias study to demonstrate laboratory capability.

12.4.1 Analyze at least four replicates of a sample solution containing aldicarb, carbofuran, oxamyl, methomyl and BDMC at a concentration in the calibration range of Levels 3 to 5. The Level 4 concentration of the 6 point calibration curve was used to set the QC acceptance criteria in this method. The matrix and chemistry should be similar to the solution used in this test method. Each replicate must be taken through the complete analytical test method including any sample preservation and pretreatment steps.

12.4.2 Calculate the mean (average) percent recovery and relative standard deviation (RSD) of the four values and compare to the acceptable ranges of the quality control (QC) acceptance criteria for the Initial Demonstration of Performance in Table 5.

12.4.3 This study should be repeated until the single operator precision and mean recovery are within the limits in Table 5. If a concentration other than the recommended concentration is used, refer to Test Method D5847 for information on applying the F test and t test in evaluating the acceptability of the mean and standard deviation.

12.4.3.1 The QC acceptance criteria for the initial demonstration of performance in Table 5 were generated from a

TABLE 5 QC Acceptance Criteria

Analyte	Test Conc. (µg/L)	Initial Demonstration of Performance			Lab Control Sample	
		Recovery (%)		Precision	Recovery (%)	
		Lower Limit	Upper Limit	Maximum % RSD	Lower Limit	Upper Limit
Aldicarb	25	65	135	14	63	136
Carbofuran	25	66	132	15	64	134
Oxamyl	25	76	114	29	61	128
Methomyl	25	82	125	16	77	129
BDMC (Surrogate)	50	59	139	41	42	156

multi-laboratory method validation involving eight laboratories. The descriptive statistics from this validation are shown in the Precision and Bias Section. The analyst must be aware that the performance data generated from multiple-laboratory data tend to be significantly wider than those generated from single-laboratory data. It is recommended that the laboratory generate their own in-house QC acceptance criteria which meets or exceeds the criteria in this standard. References on how to generate QC acceptance criteria are ASTM Standards D2777, D5847, E2554 or Method 8000B in EPA publication SW-846, which may be helpful.

12.5 Surrogate Spiking Solution:

12.5.1 A surrogate standard solution containing BDMC is added to all samples. A stock surrogate spiking solution is prepared in methanol at 50 ppm. Spiking 25 µL of this spiking solution into a 25-mL water sample results in a concentration of 50 ppb of the surrogate in the sample. The result obtained for the surrogate recovery must fall within the limits of Table 5. If the limits are not met, the affected results must be qualified with an indication that they do not fall within the performance criteria of the test method.

12.6 Method Blank:

12.6.1 Analyze a reagent water blank with each batch of 20 or fewer samples. The concentration of the carbamates found in the blank must be below the DVL. If the concentrations of the carbamates are found above this level, analysis of samples is halted until the contamination is eliminated and a blank shows no contamination at or above this level, or the results must be qualified with an indication that they do not fall within the performance criteria of the test method.

12.7 Laboratory Control Sample (LCS):

12.7.1 To ensure that the test method is in control, analyze a LCS prepared with aldicarb, carbofuran, oxamyl and methomyl at a concentration in the calibration range of Levels 3 to 5. The LCS is prepared following the analytical method and analyzed with each batch of 20 samples or less. Prepare a stock matrix spiking solution in methanol containing aldicarb, carbofuran, oxamyl and methomyl each at 25 ppm. Spike 25 µL of this stock solution into 25 mL of water to yield a concentration of 25 ppb for the carbamates in the sample. The result obtained for the LCS must fall within the limits in Table 5.

12.7.2 If the result is not within these limits, analysis of samples is halted until the problem is corrected, and either all samples in the batch must be re-analyzed, or the results must be qualified with an indication that they do not fall within the performance criteria of the test method.

12.8 Matrix Spike (MS):

12.8.1 To check for interferences in the specific matrix being tested, perform a MS on at least one sample from each batch of 20 or fewer samples by spiking the sample with a known concentration of carbamates and following the analytical method. Prepare a stock matrix spiking solution in methanol containing aldicarb, carbofuran, oxamyl and methomyl at 25 ppm. Spike 25 µL of this stock solution into 25 mL of water to yield a concentration of 25 ppb of the carbamates in the sample.

12.8.2 If the spiked concentration plus the background concentration exceeds that of the Level 6 calibration standard, the sample must be diluted to a level near the midpoint of the calibration curve.

12.8.3 Calculate the percent recovery of the spike (P) using Eq 1:

$$P = 100 \frac{|A(V_s + V) - BV_s|}{CV} \quad (1)$$

where:

- A = concentration found in spiked sample,
- B = concentration found in unspiked sample,
- C = concentration of analyte in spiking solution,
- V_s = volume of sample used,
- V = volume of spiking solution added, and
- P = percent recovery.

12.8.4 The percent recovery of the spike shall fall within the limits in Table 6. If the percent recovery is not within these limits, a matrix interference may be present in the selected sample. Under these circumstances, one of the following remedies must be employed: the matrix interference must be removed, all samples in the batch must be analyzed by a test method not affected by the matrix interference, or the results must be qualified with an indication that they do not fall within the performance criteria of the test method.

12.8.5 The matrix spike/matrix spike duplicate (MS/MSD) limits in Table 6 were generated by eight laboratories across the country using surface waters collected near their facilities. The matrix variation between the six different surface waters may have a tendency to generate significantly wider control limits than those generated by a single laboratory in one surface water matrix. It is recommended that the laboratory generate their own in-house QC acceptance criteria which meets or exceeds the criteria in this standard.

12.8.5.1 The laboratory should generate their own in-house QC acceptance criteria after the analysis of 15–20 matrix spike samples of a particular surface water matrix. References on how to generate QC acceptance criteria are ASTM standards

D5847, D2777, E2554 or Method 8000B in EPA publication SW-846, which may be helpful.

12.9 Duplicate:

12.9.1 To check the precision of sample analyses, analyze a sample in duplicate with each batch of 20 or fewer samples. If the sample contains the analyte at a level greater than 5 times the detection limit of the method, the sample and duplicate may be analyzed unspiked; otherwise, an MSD should be used.

12.9.2 Calculate the relative percent difference (RPD) between the duplicate values (or MS/MSD values) as shown in Eq 2. Compare to the RPD limit in Table 6. Relative percent difference:

$$RPD = \frac{|MSR - MSDR|}{(MSR + MSDR)/2} \times 100 \quad (2)$$

where:

- RPD = relative percent difference,
- MSR = matrix spike recovery, and
- MSDR = matrix spike duplicate recovery.

The vertical bars in Eq 2 indicate the absolute value of the difference, hence, RPD is always expressed as a positive value.

12.9.3 If the result exceeds the precision limit, the batch must be re-analyzed or the results must be qualified with an indication that they do not fall within the performance criteria of the test method.

13. Procedure

13.1 The water samples are shipped chilled between 0°C and 6°C in ≥25 mL pre-cleaned amber glass bottles with Teflon-lined caps and stored in the laboratory between 0°C and 6°C. The samples must be analyzed within 7 days of collection. If the samples are above 6°C when received or during storage, or not analyzed within 7 days of collection, the data is qualified estimated and noted in the case narrative that accompanies the data.

13.2 In the laboratory, a 25-mL Class A glass volumetric flask is used to measure a 25-mL aliquot of the sample. Every sample is then spiked with the surrogate as described in Section 12.5. The laboratory control and matrix spike samples are then spiked with the target compound as described in Sections 12.7 and 12.8. The samples are then shaken in order to mix the spike solutions throughout the water sample. The sample, the entire 25-mL volume, is filtered through the filtration device described in Section 7.2. An aliquot of that filtered sample is placed into 2-mL amber glass LC vials for analysis.

13.3 Once a passing calibration curve is generated the analysis of samples may begin. An order of analysis may be: method blank(s), laboratory control sample(s), sample(s), duplicate(s), matrix spike sample(s) followed by an end calibration check standard.

14. Calculation or Interpretation of Results

14.1 For quantitative analysis of the carbamates and BDMC surrogate, the SRM transitions are identified by comparison of retention times in the sample to those of the standards. External calibration curves are used to calculate the amounts of aldicarb, carbofuran, oxamyl, methomyl and BDMC surrogate. Calculate the concentration in µg/L (ppb) for each analyte. The individual carbamates may be reported if present at or above

TABLE 6 MS/MSD QC Acceptance Criteria

Analyte	Test Conc. (µg/L)	MS/MSD		
		Recovery (%)		Precision
		Lower Limit	Upper Limit	Maximum RPD (%)
Aldicarb	25	48	143	26
Carbofuran	25	47	140	23
Oxamyl	25	42	140	36
Methomyl	25	61	132	30
BDMC (Surrogate)	50	29	164	34

the reporting limit. If the concentration of the analyte is determined to be above the calibration range, the sample is diluted with reagent water to obtain a concentration near the mid-point of the calibration range and re-analyzed.

15. Report

15.1 Determine the results in units of µg/L (ppb) in a water sample. Calculate the concentration in the sample using the linear or quadratic calibration curve generated. All data that does not meet the specifications in the test method must be appropriately qualified.

16. Precision and Bias

16.1 The determination of precision and bias was conducted through EPA and generated applicable data to determine the precision and bias as described in Practice D2777.

16.2 This test method was tested by US EPA Region 5 Chicago Regional Laboratory (CRL) on reagent water. The samples were spiked with target compounds to obtain a 25 ppb concentration of each and a 50 ppb concentration of surrogate as described in Section 12. Table 7 contains the recoveries and standard deviation (SD) for the surrogate and target compounds.

16.3 This test method was tested by US EPA Region 5 Chicago Regional Laboratory (CRL) on Chicago River water.

The samples were spiked with target compounds to obtain a 25 ppb concentration of each and a 50 ppb concentration of surrogate as described in Section 12. Table 8 contains the recoveries and standard deviation (SD) for the surrogate and target compound.

16.4 *Multi-Laboratory Validation*—This test method has been tested by eight laboratories using reagent water and their local surface waters. The incorporation of the testing laboratory's individual local surface water was chosen to validate the test method using various surface water matrices. The surface waters were from California, Colorado, Maryland, Mississippi, Massachusetts, Georgia, Ohio and Virginia. The reagent and local surface waters were spiked across the reporting range in quadruplicate for reagent water and duplicate for surface water. The multi-laboratory data for reagent water is shown in Table 9 and for surface waters in Table 10. Results of this collaborative study may not be typical of the results for matrices other than those studied. Grubbs' outliers were removed.

17. Keywords

17.1 carbamates; liquid chromatography; mass spectrometry; water

TABLE 7 Single-Laboratory Recovery Data in Reagent Water

Precision and Accuracy Samples	Measured ppb from 25 ppb spikes				Measured ppb from 50 ppb spikes
	Aldicarb	Carbofuran	Oxamyl	Methomyl	BDMC
1	26.6	24.9	25.5	24.7	51.9
2	26.8	26.2	25.5	24.8	51.9
3	27.7	26.3	24.6	24.4	52.2
4	27.7	26.6	25.5	25.3	50.8
5	26.7	26.7	25.2	24.3	50.1
6	27.1	26.2	25.4	25.4	52.9
Average Recovery	27.1	26.2	25.3	24.8	51.6
Average Percent Recovery	108.4 %	104.8 %	101.2 %	99.2 %	103.2 %
Standard Deviation (SD)	0.49	0.65	0.35	0.45	1.01
% Relative SD	1.8 %	2.5 %	1.4 %	1.8 %	2.0 %

TABLE 8 Single-Laboratory Recovery Data in Chicago River Water

Precision and Accuracy Samples	Measured ppb from 25 ppb spikes				Measured ppb from 50 ppb spikes
	Aldicarb	Carbofuran	Oxamyl	Methomyl	BDMC
Blank 1 Chicago River Water	ND	ND	ND	ND	47.6
Blank 2 Chicago River Water	ND	ND	ND	ND	45.4
1	26.9	25.9	14.6	17.5	46.8
2	27.0	25.7	13.3	16.9	45.3
3	27.0	25.8	14.0	17.4	47.0
4	26.7	25.8	14.3	17.6	48.4
5	27.0	24.8	13.2	16.8	46.3
6	26.0	25.6	13.6	16.9	46.4
Average Recovery	26.8	25.6	13.8	17.2	46.7
Average Percent Recovery	107.2 %	102.4 %	55.2 %	68.8 %	93.4 %
Standard Deviation (SD)	0.39	0.40	0.56	0.35	1.05
% Relative SD	1.5 %	1.6 %	4.1 %	2.0 %	2.2 %

TABLE 9 Multi-Laboratory Recovery Data in Reagent Water

Analyte	Spike Conc. (ppb)	# Results	# Labs	Bias			Precision			
				Mean Recovery (%)	Min Recovery (%)	Max Recovery (%)	Overall SD (%)	Pooled within-lab SD (%)	Overall RSD (%)	Pooled within-lab RSD (%)
(BDMC)	50	128	8	99.27	37.30	189.60	27.97	24.31	28.17	23.09
Aldicarb	1	32	8	85.64	40.00	166.10	31.11	14.69	36.32	16.97
Aldicarb	5	32	8	96.04	53.60	150.00	24.52	14.50	25.53	13.52
Aldicarb	25	32	8	99.79	67.00	135.70	15.79	7.49	15.82	7.64
Aldicarb	75	32	8	99.35	63.07	135.30	17.65	8.32	17.77	8.20
Carbofuran	1	32	8	85.40	46.70	126.80	17.66	12.28	20.68	14.33
Carbofuran	5	32	8	94.03	51.42	140.00	22.11	13.85	23.51	13.00
Carbofuran	25	32	8	98.87	67.20	130.90	15.48	8.31	15.66	8.35
Carbofuran	75	32	8	94.64	64.93	135.20	19.00	9.41	20.07	9.94
Methomyl	1	32	8	90.61	24.00	147.50	28.46	17.38	31.40	18.83
Methomyl	5	32	8	103.03	64.90	152.00	21.47	14.94	20.84	13.57
Methomyl	25	32	8	103.21	81.60	121.04	12.12	8.96	11.75	8.75
Methomyl	75	32	8	102.48	74.40	152.94	19.61	7.49	19.14	7.15
Oxamyl	1	32	8	90.84	10.00	202.00	43.79	34.00	48.20	43.78
Oxamyl	5	32	8	96.70	61.21	150.00	21.40	15.05	22.13	14.46
Oxamyl	25	32	8	94.98	51.05	130.00	16.03	15.39	16.88	16.93
Oxamyl	75	32	8	95.38	77.70	113.33	9.08	7.96	9.51	8.24

TABLE 10 Multi-Laboratory Recovery Data in Surface Water

Analyte	Spike Conc. (ppb)	# Results	# Labs	Bias			Precision			
				Mean Recovery (%)	Min Recovery (%)	Max Recovery (%)	Overall SD (%)	Pooled within-lab SD (%)	Overall RSD (%)	Pooled within-lab RSD (%)
(BDMC)	50	64	8	96.77	54.60	163.87	28.65	11.39	29.61	11.20
Aldicarb	1	16	8	104.51	0.00	225.90	52.04	13.40	49.79	50.52
Aldicarb	5	16	8	93.81	57.00	135.90	17.98	10.26	19.17	10.82
Aldicarb	25	16	8	95.49	52.40	130.00	20.45	7.39	21.42	9.71
Aldicarb	75	16	8	96.95	60.53	132.13	22.26	12.35	22.96	11.23
Carbofuran	1	16	8	103.88	14.00	200.70	44.53	14.77	42.86	33.00
Carbofuran	5	16	8	99.78	80.80	135.40	16.47	8.99	16.51	8.58
Carbofuran	25	16	8	93.87	54.40	121.00	19.74	6.56	21.03	6.83
Carbofuran	75	16	8	99.12	55.47	130.10	20.99	5.45	21.18	6.38
Methomyl	1	16	7	104.19	59.00	150.00	22.01	8.76	21.12	10.22
Methomyl	5	16	8	102.17	74.60	126.10	13.92	9.47	13.62	9.18
Methomyl	25	16	8	96.34	65.60	121.41	15.91	8.77	16.51	8.46
Methomyl	75	16	8	96.05	65.20	148.53	24.63	6.88	25.64	7.03
Oxamyl	1	16	8	80.54	0.00	130.00	42.53	30.84	52.80	66.69
Oxamyl	5	16	8	93.85	55.80	134.90	24.13	5.45	25.71	6.91
Oxamyl	25	16	8	91.38	67.20	140.00	21.58	10.02	23.62	11.02
Oxamyl	75	16	8	90.23	69.07	139.00	24.16	3.92	26.78	4.22

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the ASTM website (www.astm.org/COPYRIGHT/).