



Standard Test Method for Calibration of Primary Non-Concentrator Terrestrial Photovoltaic Reference Cells Using a Tabular Spectrum¹

This standard is issued under the fixed designation E1125; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method is intended to be used for calibration and characterization of primary terrestrial photovoltaic reference cells to a desired reference spectral irradiance distribution, such as Tables [G173](#). The recommended physical requirements for these reference cells are described in Specification [E1040](#). Reference cells are principally used in the determination of the electrical performance of photovoltaic devices.

1.2 Primary photovoltaic reference cells are calibrated in natural sunlight using the relative spectral response of the cell, the relative spectral distribution of the sunlight, and a tabulated reference spectral irradiance distribution.

1.3 This test method requires the use of a pyrheliometer that is calibrated according to Test Method [E816](#), which requires the use of a pyrheliometer that is traceable to the World Radiometric Reference (WRR). Therefore, reference cells calibrated according to this test method are traceable to the WRR.

1.4 This test method is a technique that may be used instead of the procedures found in Test Method [E1362](#). This test method offers convenience in its ability to characterize a reference cell under any spectrum for which tabulated data are available. The selection of the specific reference spectrum is left to the user.

1.5 This test method applies only to the calibration of a photovoltaic cell that shows a linear dependence of its short-circuit current on irradiance over its intended range of use, as defined in Test Method [E1143](#).

1.6 This test method applies only to the calibration of a reference cell fabricated with a single photovoltaic junction.

1.7 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.8 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 *ASTM Standards:*²

[E772](#) Terminology of Solar Energy Conversion

[E816](#) Test Method for Calibration of Pyrheliometers by Comparison to Reference Pyrheliometers

[E948](#) Test Method for Electrical Performance of Photovoltaic Cells Using Reference Cells Under Simulated Sunlight

[E973](#) Test Method for Determination of the Spectral Mismatch Parameter Between a Photovoltaic Device and a Photovoltaic Reference Cell

[E1021](#) Test Method for Spectral Responsivity Measurements of Photovoltaic Devices

[E1040](#) Specification for Physical Characteristics of Nonconcentrator Terrestrial Photovoltaic Reference Cells

[E1328](#) Terminology Relating to Photovoltaic Solar Energy Conversion (Withdrawn 2012)³

[E1362](#) Test Method for Calibration of Non-Concentrator Photovoltaic Secondary Reference Cells

[G173](#) Tables for Reference Solar Spectral Irradiances: Direct Normal and Hemispherical on 37° Tilted Surface

3. Terminology

3.1 *Definitions*—Definitions of terms used in this test method may be found in Terminology [E772](#) and Terminology [E1328](#).

3.2 *Symbols:*

3.2.1 The following symbols and units are used in this test method:

λ —Wavelength, nm or μm ,

¹ This test method is under the jurisdiction of ASTM Committee [E44](#) on Solar, Geothermal and Other Alternative Energy Sources and is the direct responsibility of Subcommittee [E44.09](#) on Photovoltaic Electric Power Conversion.

Current edition approved March 1, 2015. Published April 2015. Originally approved in 1986. Last previous edition approved in 2010 as E1125 – 10. DOI: 10.1520/E1125-10R15.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ The last approved version of this historical standard is referenced on www.astm.org.

I_{sc} —Short-circuit current, A,
 E —Irradiance, Wm^{-2} ,
 E_t —Total irradiance, Wm^{-2} ,
 $E(\lambda)$ —Spectral irradiance, $\text{Wm}^{-2}\mu\text{m}^{-1}$,
 $R(\lambda)$ —Spectral response, AW^{-1} ,
 $R_r(\lambda)$ —Reference cell spectral response, AW^{-1} ,
 T —Temperature, $^{\circ}\text{C}$,
 α —Temperature coefficient of reference cell I_{sc} , $^{\circ}\text{C}^{-1}$,
 n —Total number of data points,
 C —Calibration constant, Am^2W^{-1} ,
 M —Spectral mismatch parameter,
 F —Spectral correction factor, and
 S —Standard deviation.

4. Summary of Test Method

4.1 The calibration of a primary photovoltaic reference cell consists of measuring the short-circuit current of the cell when illuminated with natural sunlight, along with the total solar irradiance using a pyrheliometer. The ratio of the short-circuit current of the cell to the irradiance, divided by a correction factor similar to the spectral mismatch parameter defined in Test Method E973, is the calibration constant for the reference cell. Also, if the temperature of the cell is not $25 \pm 1^{\circ}\text{C}$, the short-circuit current must be corrected to 25°C .

4.1.1 The relative spectral irradiance of the sunlight is measured using a spectral irradiance measurement instrument as specified in Test Method E973.

4.2 The following is a list of measurements that are used to characterize reference cells and are reported with the calibration data:

4.2.1 The spectral response of the cell is determined in accordance with Test Methods E1021.

4.2.2 The cell's short-circuit current temperature coefficient is determined experimentally by measuring the short-circuit current at various temperatures and computing the temperature coefficient (see 7.2.2).

4.2.3 Linearity of short-circuit current versus irradiance is determined in accordance with Test Method E1143.

4.2.4 The fill factor of the reference cell is determined using Test Method E948. Providing the fill factor with the calibration data allows the reference cell to be checked in the future for electrical degradation or damage.

5. Significance and Use

5.1 The electrical output of a photovoltaic device is dependent on the spectral content of the illumination source, its intensity, and the device temperature. To make standardized, accurate measurements of the performance of photovoltaic devices under a variety of light sources, it is necessary to account for the error in the short-circuit current that occurs if the relative spectral response of the reference cell is not identical to the spectral response of the device to be tested. A similar error occurs if the spectral irradiance distribution of the test light source is not identical to the desired reference spectral irradiance distribution. These errors are accounted for by the spectral mismatch parameter (described in Test Method E973), a quantitative measure of the error in the short-circuit current measurement. It is the intent of this test method to provide a

recognized procedure for calibrating, characterizing, and reporting the calibration data for primary photovoltaic reference cells using a tabular reference spectrum.

5.2 The calibration of a reference cell is specific to a particular spectral irradiance distribution. It is the responsibility of the user to specify the applicable irradiance distribution, for example Tables G173. This test method allows calibration with respect to any tabular spectrum.

5.3 A reference cell should be recalibrated at yearly intervals, or every six months if the cell is in continuous use outdoors.

5.4 Recommended physical characteristics of reference cells can be found in Specification E1040.

5.5 Because silicon solar cells made on p-type substrates are susceptible to a loss of I_{sc} upon initial exposure to light, it is required that newly manufactured reference cells be light soaked at an irradiance level greater than 850 W/m^2 for 2 h prior to initial characterization in Section 7.

6. Apparatus

6.1 *Pyrheliometer*—A secondary reference pyrheliometer that is calibrated in accordance with Test Method E816. An absolute cavity radiometer may also be used. Because secondary reference pyrheliometers are calibrated against an absolute cavity radiometer, the total uncertainty in the primary reference cell calibration constant will be reduced if an absolute cavity radiometer is used.

6.2 *Collimator*—A collimator fitted to the reference cell during calibration that has the same field-of-view as the pyrheliometer. An acceptable collimator design is described in Annex A1.

6.3 *Spectral Irradiance Measurement Equipment*, as required by Test Method E973.

6.3.1 The spectral range of the spectral irradiance measurement shall be wide enough to include the spectral response of the cell to be calibrated.

6.3.2 The spectral range of the spectral irradiance measurement shall include 98 % of the total irradiance to which the pyrheliometer is sensitive.

6.3.3 If the spectral irradiance measurement is unable to measure the entire wavelength range required by 6.3.2, it is acceptable to use a reference spectrum, such as Tables G173, to supply the missing wavelengths. The reference spectrum is scaled to match the measured spectral irradiance data over a convenient wavelength interval within the wavelength range of the spectral irradiance measurement equipment. It is also acceptable to calculate the missing spectral irradiance data using a numerical model.

6.3.4 The spectral irradiance measurement equipment shall have the same field-of-view as the pyrheliometer and the reference cell collimator.

6.4 *Normal Incidence Tracking Platforms*—Tracking platforms used to follow the sun during the calibration and to hold the reference cell to be calibrated, the pyrheliometer, the collimator, and spectral irradiance measurement equipment. The pyrheliometer and the collimator must be parallel within

$\pm 0.25^\circ$. The platforms shall be able to track the sun within $\pm 0.5^\circ$ during the calibration procedure.

6.5 Temperature Measurement Equipment—An instrument or instruments used to measure the temperature of the reference cell to be calibrated, that has a resolution of at least 0.1°C , and a total error of less than $\pm 1^\circ\text{C}$ of reading.

6.5.1 Sensors such as thermocouples or thermistors used for the temperature measurements must be located in a position that minimizes any temperature gradients between the sensor and the photovoltaic device junction.

6.6 Electrical Measurement Equipment—Voltsmeters, ammeters, or other suitable electrical measurement instruments, used to measure the I_{sc} of the cell to be calibrated and the pyrheliometer output, that have a resolution of at least 0.02 % of the maximum current or voltage encountered, and a total error of less than 0.1 % of the maximum current or voltage encountered.

6.7 Spectral Response Measurement Equipment, as required by Test Method E1021.

6.7.1 The wavelength interval between spectral response data points shall be a maximum of 50 nm.

6.8 Temperature Control Block (Optional)—A device to maintain the temperature of the reference cell at $25 \pm 1^\circ\text{C}$ for the duration of the calibration.

7. Characterization

7.1 Prior to the characterization measurements, illuminate the reference cell to be calibrated at 1000 Wm^{-2} for 2 h. This is necessary to stabilize any light-induced degradation of the cell prior to calibration.

7.2 Characterize the reference cell being calibrated by the following methods:

7.2.1 Spectral Response—Determine the relative spectral response, $R(\lambda)$, (optionally the absolute spectral response) of the cell to be calibrated in accordance with Test Methods E1021.

7.2.2 Temperature Coefficient—Determine the temperature coefficient, α , of the cell to be calibrated as follows:

7.2.2.1 Using the electrical measurement equipment, measure I_{sc} at four or more temperatures over at least a 50°C temperature range centered around 35°C . The irradiance shall be at least 750 Wm^{-2} and less than 1100 Wm^{-2} , as measured with a second reference cell. Measure the temperature of the being calibrated at the same time.

7.2.2.2 Divide each value of I_{sc} by the normalized instantaneous irradiance level at the time of each measurement.

NOTE 1—The normalized instantaneous irradiance can be determined by dividing the second reference cell's I_{sc} by its calibration constant.

7.2.2.3 Determine the temperature coefficient by performing a least-squares fit of the I_{sc} versus T data to a straight line. The slope of the line divided by the value of the current from the least-squares fit at 25°C is the temperature coefficient, α .

7.2.3 Linearity—Determine the short-circuit current versus irradiance linearity of the cell being calibrated in accordance with Test Method E1143 for the irradiance range 750 to 1100 Wm^{-2} .

7.2.3.1 For reference cells that use single-crystal silicon solar cells, or for reference cells that have been previously characterized, the short-circuit current versus irradiance linearity determination is optional.

7.2.4 Fill Factor—Determine the fill factor of the cell to be calibrated from the I-V curve of the device, as measured in accordance with Test Methods E948.

8. Procedure

8.1 Mount the reference cell to be calibrated, the collimator, the pyrheliometer, and the spectral irradiance measurement equipment on the tracking platforms.

8.2 Measure the relative spectral irradiance of the sun, $E(\lambda)$, using the spectral irradiance measurement instrument (see 6.6) and the procedure of Test Method E973. During the spectral irradiance measurement, perform the following:

8.2.1 Measure the pyrheliometer output, E_p , and verify that the total irradiance is between 750 Wm^{-2} and 1100 Wm^{-2} .

8.2.2 Measure the short-circuit current of the reference cell, I_{sc} .

8.2.3 Measure the reference cell temperature, T .

8.2.4 Repeat 8.2.1 and 8.2.2 at least four times. These repetitions must be distributed in time during the spectral irradiance measurement. To assure temporal stability, the short-circuit current of the reference cell shall not vary by more than $\pm 0.2\%$ during the repetitions.

8.2.5 Average the short-circuit current and total irradiance values from 8.2.4 to obtain the I_{sc} and E_i that corresponds to the spectral irradiance measurement.

8.3 Perform a minimum of five replications of 8.2.

8.3.1 The five replications must be performed on at least three separate days. Therefore, five replications all performed on the same day would not be an acceptable data set for the calibration.

8.3.2 In order to reduce precision errors through averaging, it is recommended that at least 30 replications of 8.2 be performed.

9. Calculation of Results

9.1 Each spectral irradiance measurement obtained in 8.2 defines one data point. The total number of these data points is denoted as n .

9.1.1 For each data point, calculate the spectral correction factor, F , using the spectral mismatch parameter calculation, 8.1, of Test Method E973. To perform this calculation, replace the reference cell spectral response $R_{ref}(\lambda)$ with unity (this represents the spectral response of the pyrheliometer), and replace M with F .

9.2 Calculate the calibration constant for each data point, using the following:

$$C_i = \frac{I_{sc}}{E_i} \frac{1}{F} \frac{1}{1 - \alpha(25 - T)} \quad (1)$$

NOTE 2—The temperature correction term may be deleted if the measured reference cell temperature is $25 \pm 1^\circ\text{C}$.

9.3 Calculate the average calibration constant with the following:

$$C = \frac{1}{n} \sum_{i=1}^n C_i \quad (2)$$

9.4 Compute the standard deviation of the calibration constant using the following:

$$S = \left[\frac{\left[\sum_{i=1}^n (C_i)^2 \right] - nC^2}{n-1} \right]^{1/2} \quad (3)$$

9.4.1 The value of S shall be 1 % or less of the calibration constant C .

10. Report

10.1 Report, as a minimum, the following information:

- 10.1.1 Reference cell serial number.
- 10.1.2 Data of calibration.
- 10.1.3 *Calibration Constant*—From 9.4.
- 10.1.4 *Standard Deviation*—From 9.5.
- 10.1.5 *Fill Factor*—From 7.2.4.
- 10.1.6 *Linearity Verification*—From 7.2.3.
- 10.1.7 *Spectral Response*—From 7.2.1.
- 10.1.8 *Temperature Coefficient*—From 7.2.2.
- 10.1.9 Pyrheliometer (type, manufacturer, serial number, calibration constant, data last calibrated).
- 10.1.10 Complete description of measurement system.
- 10.1.11 Any deviations from the standard calibration procedure.
- 10.1.12 Any unusual occurrences during calibration.
- 10.1.13 Data for each point in calibration, that shall include the following:
 - 10.1.13.1 Cell temperature,
 - 10.1.13.2 Total irradiance,
 - 10.1.13.3 Short-circuit current, and
 - 10.1.13.4 Spectral correction factor.

11. Precision and Bias

11.1 *Precision*—It is not possible to specify the precision of the reference cell calibration test method using the results of an interlaboratory study because no laboratories were willing to participate in such study. The restrictions placed on the apparatus and the calibration conditions have been selected to minimize precision errors in the reference cell calibration constant. Factors that contribute to the total precision error include:

11.1.1 Temporal variations of the solar spectral and total irradiance during the measurement of the spectral irradiance (see 8.2) will introduce errors in the calibration period. To minimize these errors, calibration periods should be selected where variations in the short-circuit current of the reference cell and the pyrheliometer output are as small as possible.

11.1.2 The discussion of precision of spectral measurements in 9.1 of Test Method E973 is applicable to the reference cell calibration test method.

11.1.3 Temperature variations of the reference cell being calibrated within the $25 \pm 1^\circ\text{C}$ band will introduce small errors in the calibration constant if the temperature corrections are not employed (see Note 2). The temperature coefficient (see section 7.2.2) will determine the magnitude of these errors.

11.1.4 Electronic instrumentation used to measure the reference cell short-circuit current, the total irradiance, and the cell temperature will contribute precision errors to the calibration constant.

11.2 *Bias*—The contribution of bias to the total error will depend upon the bias of each individual factor used for the determination of the calibration constant. Possible individual contributions of bias include:

11.2.1 The slope of the cell's I-V curve near zero volts, and loading of the cell by the current measurement instrument due to nonzero input impedance can result in somewhat smaller values of the short-circuit current. This situation can be minimized by forcing the reference cell voltage as close to zero as possible during the short-circuit current measurement.

11.2.2 Measurement of the cell temperature at the back of the device will give a value that is lower than the junction temperature during exposure of the cell to sunlight. This may result in slightly too high a value for short-circuit current. Because the short-circuit current temperature coefficient is usually small, this source of bias tends to be small.

11.2.3 Each measurement instrument will introduce bias into the final calibration in varying amounts. It is assumed that all instruments are calibrated at regular intervals. However, bias will still affect any instrumentation even after careful calibration.

11.2.4 An absolute accuracy of 0.25 % for terrestrial solar radiometric measurements has been established for absolute cavity radiometers that have been compared with the World Radiometric Reference. If a secondary reference pyrheliometer is used, a 1 % transfer error from the cavity radiometer should be expected when utilizing the procedures of Test Method E816.

11.2.5 The discussion of bias in spectral measurements in 9.2 of Test Method E973 is applicable to the reference cell calibration test method.

12. Keywords

12.1 calibration; electrical performance; photovoltaic devices; primary terrestrial photovoltaic reference cells; spectral irradiance; spectral response; terrestrial photovoltaic reference cells

ANNEX

(Mandatory Information)

A1. COLLIMATOR DESIGN

A1.1 A collimator aligned with the solar beam is used to restrict the field of view of a solar cell to the direct beam of sunlight plus some amount of circumsolar radiation.

A1.1.1 **Fig. A1.1** schematically shows the design of a collimator tube. The amount of circumsolar radiation at the receiving aperture (radius r) is determined by the opening angle θ_o , which may also be specified in terms of the ratio $L/2R$ of length and diameter of the collimator tube. A value of $L/2R = 10$ to 12 may be used provided a collimator-to-collimator accuracy of 5 % is maintained. (An $L/2R = 10$ is the value used with standard pyrheliometer, corresponding to an opening angle of the cone of $2\theta_o = 5.72^\circ$).

A1.1.2 A slope angle θ_s is included in the collimator design to allow for a pointing and tracking error, and subtense angle of

the sun. The disk of the sun subtends a half angle of 0.26° from center to rim. The pointing and tracking errors may be as large as 0.5° . The design value for the slope angle, therefore, is a minimum of 1.26° . Larger slope angles permit greater tolerance in pointing and tracking, but result in a longer collimator tube. A practical upper limit for the slope angle is 2° . The receiving aperture radius r is determined by the size of the solar cell to be used with the collimator. For a rectangular cell, r is chosen as half the diagonal of the cell.

A1.2 *Design Summary*—With the slope angle, length-to-diameter ratio, and receiving aperture defined, the dimensions of the collimator are determined from geometrical considerations. The interior surface of the collimator shall be blackened and baffled to attenuate reflections from the interior walls.




FIG. A1.1 Collimator Design

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; <http://www.copyright.com/>