



Standard Test Method for Calibration of Narrow- and Broad-Band Ultraviolet Radiometers Using a Spectroradiometer¹

This standard is issued under the fixed designation G 130; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reappraisal. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reappraisal.

INTRODUCTION

Accurate and precise measurements of ultraviolet irradiance are required in the determination of the radiant exposure of both total and selected narrow bands of ultraviolet radiation for the determination of exposure levels in (1) outdoor weathering of materials, (2) indoor accelerated exposure testing of materials using manufactured light sources, and (3) UV-A and UV-B ultraviolet radiation in terms both of the assessment of climatic parameters and the changes that may be taking place in the solar ultraviolet radiation reaching earth.

Although meteorological measurements usually require calibration of pyranometers and radiometers oriented with axis vertical, applications associated with materials testing require an assessment of the calibration accuracy at orientations with the axis horizontal (usually associated with testing in indoor exposure cabinets) or with the axis at angles typically up to 45° or greater from the horizontal (for outdoor exposure testing). These calibrations also require that deviations from the cosine law, tilt effects, and temperature sensitivity be either known and documented for the instrument model or determined on individual instruments.

This test method requires calibrations traceable to primary reference standards maintained by a national metrological laboratory that has participated in intercomparisons of standards of spectral irradiance.

1. Scope

1.1 This test method covers the calibration of ultraviolet light-measuring radiometers possessing either narrow- or broad-band spectral response distributions using either a scanning or a linear-diode-array spectroradiometer as the primary reference instrument. For transfer of calibration from radiometers calibrated by this test method to other instruments, Test Method E 824 should be used.

NOTE 1—Special precautions must be taken when a diode-array spectroradiometer is employed in the calibration of filter radiometers having spectral response distributions below 320-nm wavelength. Such precautions are described in detail in subsequent sections of this test method.

1.2 This test method is limited to calibrations of radiometers against light sources that the radiometers will be used to measure during field use.

NOTE 2—For example, an ultraviolet radiometer calibrated against natural sunlight cannot be employed to measure the total ultraviolet irradiance of a fluorescent ultraviolet lamp.

1.3 Calibrations performed using this test method may be against natural sunlight, Xenon-arc burners, metal halide burners, tungsten and tungsten-halogen lamps, fluorescent lamps, etc.

1.4 Radiometers that may be calibrated by this test method include narrow-, broad-, and wide-band ultraviolet radiometers, and narrow-, broad, and wide-band visible-region-only radiometers, or radiometers having wavelength response distributions that fall into both the ultraviolet and visible regions.

NOTE 3—For purposes of this test method, narrow-band radiometers are those with $\Delta\lambda \leq 20$ nm, broad-band radiometers are those with $20 \text{ nm} \leq \Delta\lambda \leq 70$ nm, and wide-band radiometers are those with $\Delta\lambda \geq 70$ nm.

¹ This test method is under the jurisdiction of ASTM Committee G3 on Durability of Nonmetallic Materials and is the direct responsibility of Subcommittee G03.09 on Solar and Ultraviolet Radiation Measurement Standards.

Current edition approved June 1, 2006. Published July 2006. Originally approved in 1995. Last previous edition approved in 2002 as G 130–95(2002)

NOTE 4—For purposes of this test method, the ultraviolet region is defined as the region from 285 to 400-nm wavelength, and the visible region is defined as the region from 400 to 750-nm wavelength. The ultraviolet region is further defined as being either UV-A with radiation of wavelengths from 315 to 400 nm, or UV-B with radiation from 285 to 315-nm wavelength.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:²

E 772 Terminology Relating to Solar Energy Conversion

E 824 Test Method for Transfer of Calibration From Reference to Field Radiometers

3. Terminology

3.1 Definitions:

3.1.1 *broad-band radiometer*—a relative term generally applied to radiometers with interference filters or cut-on/cut-off filter pairs having a FWHM between 20 and 70 nm and with tolerances in center (peak) wavelength and FWHM no greater than ± 2 nm.

3.1.2 *diode array detector*—a detector with from 50 to 1000 silicon photodiodes affixed side-by-side in a linear array and mounted in the focal plane of the exit slit of a monochromator.

3.1.3 *full width at half maximum (FWHM)*—in a bandpass filter, FWHM is the interval between wavelengths at which transmittance is 50 % of the peak, frequently referred to as bandwidth.

3.1.4 *narrow-band radiometer*—a relative term generally applied to radiometers with interference filters with FWHM ≤ 20 nm and with tolerances in center (peak) wavelength and FWHM no greater than ± 2 nm.

3.1.5 *scanning monochromator*—a monochromator that uses either a single, or several interchangeable, detector(s) mounted at the exit slit, that is presented with dispersed light by sweeping the spectrum across the slit to illuminate the detector with a succession of different very narrow wavelength light distributions. The detector may be either a photomultiplier tube (PMT) or silicon photodiode (visible), or a PMT or an ultraviolet-enhanced silicon photodiode (ultraviolet and visible), or a lead sulfide cell or other solid state detector (near infrared), etc. The dispersed spectrum is swept across the monochromator's exit slit using a mechanical stage that rotates either a prism or a grating dispersive element, usually under the control of an external microprocessor or computer.

3.1.6 *spectroradiometer*—a radiometer consisting of a monochromator with special acceptance optics mounted to the entrance aperture and a detector mounted to the exit aperture, usually provided with electronic or computer encoding of wavelength and radiometric intensity. The monochromator of

such instruments is either of the linear diode (often called *diode array*) or the scanning type.

3.1.7 *wide-band radiometer*—a relative term generally applied to radiometers with combinations of cut-off and cut-on filters with FWHM greater than 70 nm.

3.2 For other terms relating to this test method, see Terminology **E 772**.

4. Significance and Use

4.1 This test method represents the preferable means for calibrating both narrow-band and broad-band ultraviolet radiometers. Calibration of narrow- and broad-band ultraviolet radiometers involving direct comparison to a standard source of spectral irradiance is an alternative method for calibrating ultraviolet radiometers. An ASTM test method describing this procedure is under development by Subcommittee G03.09 on Radiometry.

4.2 The accuracy of this calibration technique is dependent on the condition of the light source (for example, cloudy skies, polluted skies, aged lamps, defective luminaires, etc.), and on source alignment, source to receptor distance, and source power regulation.

NOTE 5—It is conceivable that a radiometer might be calibrated against a light source that represents an arbitrarily chosen degree of aging for its class in order to present to both the test and reference radiometers a spectrum that is most typical for the type.

4.3 Spectroradiometric measurements performed using either an integrating sphere or a cosine receptor (such as a shaped TFE³, or Al₂O₃ diffuser plate) provide a measurement of hemispherical spectral irradiance in the plane of the sphere's entrance port. As such, the aspect relative to the reference light source must be defined (azimuth and tilt from the horizontal for solar measurements, normal incidence with respect to the beam component of sunlight, or normal incidence and the geometrical aspect with respect to an artificial light source, or array). It is important that the geometrical aspect between the plane of the spectroradiometer's source optics and that of the radiometer being calibrated be as nearly identical as possible.

NOTE 6—When measuring the hemispherical spectral energy distribution of an array of light sources (for lamps), normal incidence is defined by the condition obtained when the plane of the sphere's aperture is parallel to the plane of the lamp, or burner, array.

4.4 Calibration measurements performed using a spectroradiometer equipped with a pyrheliometer-comparison tube (a sky-occluding tube), regardless of whether affixed directly to the monochromator's entrance slit, to the end of a fibre optic bundle, or to the aperture of an integrating sphere, shall not be performed unless the radiometer being calibrated is a true pyrheliometer (that is, unless it possesses a view-limiting device having the approximate optical constants of the spectroradiometer's pyrheliometer-comparison tube).

4.5 Spectroradiometric measurements performed using source optics other than the integrating sphere or the "standard" pyrheliometer comparison tube, shall be agreed upon in advance between all involved parties.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Tetrafluoroethylene such as a special grade of Teflon® or an equivalent material, has been found suitable.

4.6 Calibration measurements that meet the requirements of this test method are traceable to a national metrological laboratory that has participated in intercomparisons of standards of spectral irradiance, largely through the traceability of the standard lamps and associated power supplies employed to calibrate the spectroradiometer.

4.7 The accuracy of calibration measurements performed employing a spectroradiometer is dependent on, among other requirements, the degree to which the temperature of the mechanical components of the monochromator are maintained during field measurements in relation to those that prevailed during calibration of the spectroradiometer.

NOTE 7—This requirement is covered in detail in an ASTM standard under development in Subcommittee G03.09 on Radiometry.

5. Apparatus

5.1 Reference Spectroradiometers:

5.1.1 The spectroradiometer employed as the reference radiometer shall, regardless of whether it consists of a scanning or a linear-diode-array monochromator, be calibrated within the last month in accordance with the procedures specified by CIE Publication 63⁴ and the manufacturer.

5.1.1.1 It is recommended that the reference spectroradiometer, or one of its exact type, has been a participating spectroradiometer in an intercomparison of spectroradiometers either managed, sponsored, or sanctioned by a national metrological laboratory, or another appropriate body.

5.1.1.2 Alternatively, it is recommended that the reference spectroradiometer shall have participated in an intercomparison by measurement of a reference lamp source that is either managed, sponsored, or sanctioned by a national metrological laboratory, or another appropriate body.

5.1.2 If a linear diode-array spectroradiometer is used as the reference, it shall possess focusing optics internal to the monochromator and a linear diode array detector with a sufficient number of diodes that, together, result in a resolving power of 1 nm or less. The monochromator's dispersive element shall be a holographic grating, and the spectroradiometer's acceptance optics shall consist of either an integrating sphere with appropriately sized and oriented light entrance port, or a shaped translucent diffuser plate (such as a TFE³ or Al₂O₃ wafer) whose deviation from true cosine response is small and known. A further requirement is that the stray light rejection be determined for any diode-array spectroradiometers used to perform this test method and that it be 10⁵ or greater in the spectral region for which calibration is required.

5.1.2.1 A diode-array spectroradiometer shall not be used as the reference instrument below 300-nm wavelength. Further, when used in the wavelength region between 300 and 320-nm wavelength, evidence shall be presented with the calibration reports, or certificates, showing that the stray light has been eliminated by a combination of internal baffling, merging of two determinations in which the wavelength region below

320-nm is measured employing secondary filters to reject all wavelengths longer than 320 nm, other techniques, or combinations of these.

5.1.3 When an integrating sphere is used, the exit port (to the monochromator) and entrance port (that represents the receiver) should be oriented 90° to each other and the sphere should be equipped with a baffle to occlude all light that might reach the exit directly from the entrance port.

5.1.4 When a pyrheliometer-comparison tube, or other view-limiting device, is used for the purpose of calibrating, for example, ultraviolet pyrheliometers, the pyrheliometer-comparison tube should ideally be affixed to the entrance port of the integrating sphere such that the sphere's entrance port becomes the aperture stop of the view-limiting device. Under most circumstances, the pyrheliometer comparison tube should possess the optical geometry defined by the World Meteorological Organization, the principal one being a 5.6° field of view.

NOTE 8—When the sphere's entrance port is the occluder's aperture stop, no calibration of the spectroradiometer is required independent of the calibration with only the integrating sphere in place. If the occluder's aperture stop is integral with the occluder and of different smaller dimension than the sphere's entrance port, the spectroradiometer must be calibrated with the occluder attached to the integrating sphere ... resulting in greater uncertainties and the possibilities of significant errors.

5.2 Computational Facilities—The computer-based computational facilities used to import the raw data with respect to wavelength and intensity should be capable of providing analyzed spectral irradiance information integrated across any wavelength band chosen.

5.3 Instrument Mounts:

5.3.1 Equatorial Mount—An altazimuthal or equatorial, follow-the sun mount that is equipped with a platform on which the spectroradiometer is mounted is required for all hemispherical normal-incident and direct (beam) calibrations measurements.

5.3.2 Tilt Table—A stable, adjustable tilt table having tilt and azimuth adjustments is required for global solar radiation measurements (for example, at horizontal orientation) and hemispherical measurements at specified azimuthal and tilt positions.

NOTE 9—An altazimuthal mount so equipped also may be used as the tilt table.

5.3.3 Optical Platform—A stable, platform equipped with height adjustment is required for use in measuring the calibrating radiometers against light sources such as arrays, solar simulators, special lamps, and burners, etc.

NOTE 10—When using a fiber-optic/integrating sphere source configuration to calibrate radiometers, for example, against Xenon-arc lamps, carbon arcs, and other burners employed in indoor exposure cabinets, special fixtures may be required to rigidly mount and present the source optics to the source of irradiance. For UV-A and UV-B calibrations, the fiber-optic bundle must be constructed of quartz fibers.

6. Procedure

6.1 Calibrate the spectroradiometer in accordance with the manufacturer's instructions and CIE Publication 63⁴ unless the spectroradiometer's calibration is known to be stable within 30

⁴ The Spectroradiometric Measurement of Light Sources, Publication No. 63, The International Commission on Illumination (CIE).

days of the last intensity and wavelength calibration required. Verify calibration with a calibration check of at least one wavelength in the region of interest.

6.1.1 For weathering and exposure-testing applications requiring the measurement of UV-B radiation employing a single-filter radiometer, select a wavelength interval of 285 to 315 nm, regardless of the FWHM and CW of the filter radiometer being calibrated.

6.1.2 For weathering and exposure-testing applications requiring the measurement of UV-B and UV-A radiation using a multiple-filter radiometer, select a wavelength interval based on the FWHM of each of the specific filters of the radiometer calibrated.

6.1.3 For all other applications, such as UV-A, total ultraviolet, and specific narrow-band radiometry, select a wavelength interval that is either based on the FWHM of the instrument's spectral response function, or one that is agreed upon between the parties involved.

NOTE 11—When an application either requires, permits, or will likely result in, the use of filter radiometers from different manufacturers, calibration to the FWHM of the instrument's spectral response functions will result in significant instrument-to-instrument differences when measuring sources having the same spectral energy distributions. In this case, the users or specifications should state the exact wavelength interval that will be used for all calibrations.

6.2 Measurement of Light-Source Radiation for Calibration Against Sunlight:

6.2.1 Mount the radiometer to be calibrated in the geometrical configuration and aspect that will be employed in its end-use application.

6.2.2 Affix the spectroradiometer to the mount required for the measurements being performed (for example, an equatorial, follow-the-sun mount; a tilt table; or, a horizontal bench).

6.2.3 Ensure that both the radiometer being calibrated and the spectroradiometer are positioned at the same azimuth angle with respect to the sun, and at the same tilt from the horizontal.

6.2.4 Perform these calibration measurements only under clear sky conditions by ensuring that no cloud is within less than 30° of the sun during any one measurement sequence.

6.2.5 Determine the spectral irradiance distribution of the sun in conformance with the procedures specified in CIE Publication 63.⁴ Perform not less than five spectral irradiance measurements separated by at least 30 min. Ensure that at least one measurement is taken at, or not greater than 30 min from solar noon.

6.2.6 Record the instantaneous voltage signals of the radiometer, or radiometers, being calibrated at a frequency not less than every minute during the time period subtended by the spectral irradiance measurements.

6.3 Calibration Against Manufactured Light Sources:

6.3.1 Mount the radiometer to be calibrated in the geometrical configuration and aspect that will be employed in its end-use application. Ensure that the receptor aperture (for example, entrance port) of the spectroradiometer's sphere is at exactly the same distance from the light source (see section 5.2.2) as the radiometer being calibrated, and ensure that the reference spectroradiometer and radiometer being calibrated

have the same field of view of the source lamp or sun in terms of solid angle of the lamp's subtended.

6.3.2 Determine the geometrical aspect between the radiometer's aperture and the lamp by measuring the angle subtended between the aperture and the lamp. For non-circular lamp envelopes, measure the angle in two orthogonal planes, one of which is coincident with the long axis of the lamp.

6.3.3 Record the instantaneous voltage signals of the radiometer being calibrated over a length of time sufficient to establish that the reference light source is stable. For each instantaneous spectral measurement, record three output (voltage) readings of the radiometer being calibrated.

6.3.4 Carefully position the spectroradiometer and the source optics so that the aperture of the cosine receptor, or the integrating sphere (depending on the type of spectroradiometer being used), possesses the same geometrical aspect as the test radiometer being calibrated, and is exactly the same distance from the lamp's glass envelop. Ensure that the axis of the spectroradiometer's integrating sphere, or cosine receptor passes through both the entrance port and the center of the lamp. When measuring a single fluorescent tube lamp, or a Xenon-arc lamp, align the source optics with the exact center of the lamp and measure distance from the sphere aperture to the tube's glass envelop.

6.3.5 Determine the spectral irradiance distribution of the light source being employed in conformance with the procedures specified in CIE Publication 63.⁴ Take not less than three instantaneous spectral measurements spread over a 20-min period.

6.4 Computation of Instrument Sensitivity Constants When Calibrated to Sunlight:

6.4.1 Integrate the spectral irradiance data obtained by the spectroradiometer (see section 5.2.4) in the wavelength band corresponding to the wavelength band of, or assigned to, the radiometer being calibrated. For the most accurate calibration, the integral should be the FWHM of the test radiometer:

$$E_s(j) = \int_{\lambda_1}^{\lambda_2} E_{s,\lambda}(ij) d\lambda \quad (1)$$

where: $E_s(j)$ is the integrated irradiance for the measurement series j defined by the test radiometer being calibrated, λ_1 and λ_2 are the wavelength limits of integration defined above, and $E_{s,\lambda}(ij)$ is the spectral irradiance readings i in the wavelength interval $d\lambda$.

NOTE 12—The wavelength bands to which a radiometer is calibrated may be slightly larger, or slightly smaller than the "advertised" band-pass for the radiometer. The essential requirement is that the out-of-band spectrum of the reference light source, and, hence, the field source, must not represent a significantly greater irradiance than the average in-band irradiance, and the out-of-band irradiance must not exhibit poorer temporal stability than the average in-band irradiance.

6.4.2 For each value of integrated spectral irradiance $E_s(j)$, compute the average voltage $V_r(j)$ measured with the test radiometer in the interval j corresponding to the time interval of the reference spectral measurement:

$$V_r(j) = \frac{\sum_{i=1}^n V_r(ij)}{n(j)} \quad (2)$$

where: $V_r(ij)$ is the voltage reading i recorded by the radiometer being calibrated in the measurement series j summed from the first measurement $i = 1$ to the n th measurement, and where $n(j)$ is the number of readings taken during the measurement series.

6.4.3 Compute the radiometer's calibration factor $F(j)$... often referred to as either the instrument constant or sensitivity factor ... for each measurement of the spectral irradiance of the reference light source by:

$$F(j) = \frac{\sum_{i=1}^{j=n_v} V_r(ij)}{E_s(j)} \quad (3)$$

where: the n_v is the number of values of $V_r(j)$ obtained during the measurement of the integrated spectral irradiance $E_s(j)$. When the instrument employed to measure the spectral irradiance in the wavelength interval of interest is a linear diode array spectroradiometer, instantaneous measurements of spectral irradiance may, and should, be made within 15 s of the measurement made with the test radiometer. In this case, Eq 3 becomes:

$$F(j) = \frac{\sum_{i=1}^{j=n_v} V_r(ij)}{\sum_{j=1}^{n_e} E_s(j)} \quad (4)$$

where n_e should equal n_v . An alternative computation may be used in computing the instrument factor when employing a linear diode array instrument, one that would permit a determination of the standard deviation of the measurement series, is:

$$F(j) = \frac{\sum_{i=1}^{i=n} V_r(ij)}{\sum_{i=1}^{i=n} E_s(ij)} \quad (5)$$

6.4.4 The final calibration factor F is then computed from all $F(j)$'s using the following equation:

$$F = \frac{\sum_{j=1}^{j=n} F(j)}{n(j)} \quad (6)$$

7. Report

7.1 Report the following information:

7.1.1 *Title*—The title shall describe the radiometer calibrated and the reference light source that was used. For calibrations performed against the sun, only the most pertinent information should be included in the title (for example, normal incidence or tilt),

7.1.2 Manufacturer, model, serial number, and manufacturer's designated wavelength band-pass or radiometer(s) calibrated,

7.1.3 The wavelength interval, or intervals, for which the calibration was determined and for which the calibration is valid,

7.1.4 Manufacturer, model, serial number, and source optics of spectroradiometer used. Report most recent calibration history and traceability,

7.1.5 *Light Source Description*—If the sun, describe all pertinent information (solar time, aspect, component). If a lamp, include manufacturer, model number, serial number (if available), distance and aspect, and voltage (if other than standard line voltage). If a standard lamp is used as the reference source, report manufacturer, model number, serial number, calibration reference, traceability, and amperage used,

7.1.6 Radiometer(s) instrument constant derived in 6.4.2,

7.1.7 Date of calibration,

7.1.8 Date(s) of calibration of reference spectroradiometer, including any sanctioned intercomparisons in which the spectroradiometer participated,

7.1.9 Traceability chain (to a national metrological laboratory that has participated in intercomparisons of standards of spectral irradiance), and

7.1.10 Apply a calibration decal to the radiometer showing as a minimum the instrument constant and the date of calibration.

8. Precision and Bias

8.1 The precision in determining the instrument constant of an ultraviolet field radiometer used to measure the sun is influenced by sky conditions, and particularly by variations in cosine response when performing calibrations at low solar elevations and in the stability of the sun's ultraviolet spectrum during the calibration sequence.

8.2 The precision in determining the instrument constant of ultraviolet radiometers designed to measure the radiant exposure of manufactured ultraviolet sources is influenced in large part on the temporal stability of the source during the calibration sequence.

8.3 Repeatability of the average value of any calibration sequence the total of which is used to assign a calibration factor, or instrument sensitivity factor, should be such that the standard deviation is less than 1 % of the calibration value of the instrument.

8.4 Since there is no reference material source bias cannot be determined.

8.5 Reproducibility between instruments of the same manufacturer will depend on differences in the spectrum under which they were calibrated. Likewise, agreement between instruments of different manufacture will depend on differences in their spectral response distribution functions, as well as on the source spectrum against which they were calibrated.

8.6 Numerical differences and the standard deviation for data sets cannot be estimated. Hence, a need exists for conducting either field intercomparisons or interlaboratory measurements of reference sources (other than the standard lamps against which the reference spectroradiometers are calibrated).

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).