

Methods of test for Desiccants —

Part 3: Determination of pH

NOTE It is recommended that this Part be read in conjunction with BS 3482-1 “*Sampling, and preparation and storage of test samples*” which is issued separately.

Committees responsible for this British Standard

The preparation of this British Standard was entrusted by the Chemicals Standards Policy Committee (CIC/-) to Technical Committee CIC/2, upon which the following bodies were represented:

- Chemical Industries Association
- Ministry of Defence
- Royal Society of Chemistry
- Society of British Aerospace Companies Limited
- Society of Environmental Engineers

This British Standard, having been prepared under the direction of the Chemicals Standards Policy Committee, was published under the authority of the Standards Board and comes into effect on 31 July 1991

© BSI 10-1999

The following BSI references relate to the work on this standard:
Committee reference CIC/2
Draft for comment 90/55563 DC

ISBN 0 580 19699 2

Amendments issued since publication

Amd. No.	Date	Comments

Contents

	Page
Committees responsible	Inside front cover
Foreword	ii
1 Scope	1
2 Principle	1
3 Reagents	1
4 Apparatus	1
5 Procedure	1
6 Expression of results	1
7 Test report	1
Publication(s) referred to	Inside back cover

Foreword

This Part of BS 3482 has been prepared under the direction of the Chemicals Standards Policy Committee. This revision of BS 3482 comprises Part 1, which describes procedures for sampling and preparation and storage of test samples and a number of Parts describing methods of test for desiccants. This revision supersedes BS 3482:1962 which is withdrawn. It differs from BS 3482:1962 in that individual methods have been published in separate Parts and have been generally updated. Some additional methods not included in BS 3482:1962 have also been incorporated.

This Part of BS 3482 also supersedes the methods given in Appendix B of BS 2541:1960 and Appendix B of BS 3523:1962 which are withdrawn.

A British Standard does not purport to include all the necessary provisions of a contract. Users of British Standards are responsible for their correct application.

Compliance with a British Standard does not of itself confer immunity from legal obligations.

Summary of pages

This document comprises a front cover, an inside front cover, pages i and ii, pages 1 and 2, an inside back cover and a back cover.

This standard has been updated (see copyright date) and may have had amendments incorporated. This will be indicated in the amendment table on the inside front cover.

1 Scope

This Part of BS 3482 describes a method for the determination of the pH value of desiccants.

NOTE The titles of the publications referred to in this standard are listed on the inside back cover.

2 Principle

Boiling water is added to a test portion and stirred. The solution is allowed to cool to ambient temperature and the pH value is measured.

3 Reagents

3.1 General. Use only reagents of recognized analytical grade and water complying with grade 3 of BS 3978.

3.2 Three buffer solutions, of pH values 4, 7 and 9, complying with BS 1647-2.

4 Apparatus

4.1 Ordinary laboratory apparatus and the following are required.

4.2 pH meter, with glass and calomel electrode assembly, complying with BS 2586 and BS 3145, capable of discriminating to within 0.05 unit of pH.

Calibrate the pH meter using the buffer solutions (3.2) which cover the pH range between 4 and 9. Clean the probe by thoroughly rinsing it in water before and after each test until the indicated pH value changes by no more than 0.05 unit in 5 min. Carry out the calibration prior to the first measurement and check periodically during the course of each test (e.g. after every 1 h to 2 h).

4.3 Beaker, tall-form, 250 mL capacity.

4.4 Thermometer, general purpose type, complying with BS 1704, capable of indicating 20 °C to an accuracy of ± 2 °C.

4.5 Water bath, capable of being maintained at 20 ± 2 °C (see 4.4).

5 Procedure

5.1 Test portion

Weigh, to the nearest 1 g, 10 g of the laboratory sample into the beaker (4.3).

5.2 Preparation of the test solution

Add 100 mL of boiling water to the test portion (5.1) and stir. Place the beaker containing the test solution in the water bath (4.5) maintained at 20 ± 2 °C until the temperature of the test solution has reached 20 ± 2 °C as determined using the thermometer (4.4).

5.3 Determination

Determine the pH using the pH meter (4.2) operated in accordance with the manufacturer's instructions whilst gently stirring the test solution (5.2).

6 Expression of results

Record the pH value determined to the nearest 0.1 pH unit.

7 Test report

The test report shall include the following information.

- a) A complete identification of the sample.
- b) A reference to this British Standard method, i.e. BS 3482-3:1991.
- c) The results expressed in accordance with clause 6.
- d) Any unusual features noted during the determination.

Publication(s) referred to

BS 1647, *pH measurement*.

BS 1647-2, *Specification for reference value standard solutions and operational reference standard solutions*.

BS 1704, *Specification for solid-stem general purpose thermometers*.

BS 2586, *Specification for glass and reference electrodes for the measurement of pH*.

BS 3145, *Specification for laboratory pH meters*.

BS 3978, *Specification for water for laboratory use*.

BSI — British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover. Tel: 020 8996 9000. Fax: 020 8996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services. Tel: 020 8996 9001. Fax: 020 8996 7001.

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

Information on standards

BSI provides a wide range of information on national, European and international standards through its Library and its Technical Help to Exporters Service. Various BSI electronic information services are also available which give details on all its products and services. Contact the Information Centre. Tel: 020 8996 7111. Fax: 020 8996 7048.

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration. Tel: 020 8996 7002. Fax: 020 8996 7001.

Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

If permission is granted, the terms may include royalty payments or a licensing agreement. Details and advice can be obtained from the Copyright Manager. Tel: 020 8996 7070.